

Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)				
NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018				
Schedule No. 10- Flexible Hystero- Fiberscope Set(Rfx/Event number 3000002786)				
Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms
1	Page 21 Para 2a	Separate/combined module from video processor	Karl Storz	Video Processor to be specified or to be deleted.
2	Page 21 Para 2c	Powerful 300 Watt Xenon Lamp with backup lamp of Halogen/LED	Karl Storz	Powerful 300 Watt Xenon Lamp
3	Page 21 Para 2d	Lamp life should not be less than 500 hrs (approx), one spare 300W xenon lamp should be supplied in case of xenon light source		Amended as: Powerful 300 Watt Xenon or LED light source. Amended as: Lamp life should not be less than 500 hrs (approx), 2 nos. spare xenon bulbs to be supplied along with light source in case of xenon light source

for

Swiller

for

Swiller

for

Swiller