

Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018

Schedule No. 12 - Digital Video Colposcopy System (Rfx no. 3000002788)

Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
1	Page 75 Para 2.f	Technical Specifications: f. Focussing range : 130 mm	Olympus	Kindly amendment it as focusing adjustment range as 220-350 mm	Amended as: Focussing range : 130 mm or (220-350mm)
2	Page 75 Para 2.g	Technical Specifications: g. Inclination angle of Binocular tube: 45 degree	Olympus	Kindly amend it as optional as same is not mentioned in our technical literature	Deleted
3	Page 75 Para 2.i	Separate/combined module from video processor			Deleted
4	Page 75 Para 2.i	Powerful 300 Watt Xenon Lamp with backup lamp of Halogen/LED			Amended as: Powerful 300 Watt Xenon Lamp/LED light source
5	Page 75 Para 2.i	Lamp life should not be less than 500 hrs (approx), one spare 300 W xenon lamp should be supplied in case of xenon light source			Amended as: Lamp life should not be less than 500 hrs (approx). 2 nos. spare 300 W xenon bulbs should be supplied in case of xenon light source
6		Added Para:			Added Para: Light source should be supplied alongwith 2nos. Light guide cables of 3mtr or more length
7	Page 76 Para 4	4. Gynaecology Examination couch with doctors stool and Light	Olympus	Kindly delete this section as we don't manufacture Gynaecology Examination couch with doctors stool and Light	No Change

Swikar

Dr. [Signature]

[Signature]

[Signature]