

Amendment No. 12**Date: 21/01/2016****Sub: Amendment to the Tender Enquiry Document****Ref: NIT No.: HLL/PCD/PMSSY-II/07/15-16 dated 28/10/2015 and subsequent amendments**

The following changes have been incorporated in the referred NIT.

Section I
Notice Inviting Tender

Sch no	Event Number	Name of Item	Qty.	EMD	Due date & time for submission of online bids	Due date & time for submission of Tender fee and EMD in physical form	Date and time of opening of online bids
7	3000000609	OT Light LED	6	₹ 180,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
14	3000000617	ENT Treatment Unit (IMPORTED)	2	₹ 60,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
16	3000000619	Video Endoscope System under GIP Pediatrics Surgery	1	₹ 60,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
18	3000000621	Endoscope system of Neurosurgery	1	₹ 60,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
26	3000000629	Cardio tocography machine (6 Nos) with One Central Station	1	₹ 70,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
27	3000000630	Video Endoscope unit	1	₹ 120,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
31	3000000766	Ventilators High End	12	₹ 360,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs
32	3000000767	Patient Bed - I.C.U Beds (Advance)	108	₹ 540,000.00	01-02-2016 1800 Hrs	02-02-2016 1130 Hrs	02-02-2016 1230 Hrs

Section VII
Technical Specification**Item SI No. 07**
OT Light – LED

1. **Existing Para:** Operating Room Surgical Lighting System should provide an ideal combination of brightness, maneuverability, and shadow resolution without sacrificing color accuracy through a consistent LED technology with a unique faceted reflector design technology.

Read as: Operating Room Surgical Lighting System should provide an ideal combination of brightness, maneuverability, and shadow resolution without sacrificing color accuracy through a consistent LED technology with a unique faceted reflector design technology **or lenses based technology.**

2. **Existing Para:** Light head area: 5000 square cm (+/- 10%).

Read as: Deleted.

3. **Added Para:** Light head diameter: 70cm \pm 10% (Major Dome) and 50cm \pm 10% (Minor Dome)
4. **Added Para:** OT light should be homogenous and shadow free

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids