

Amendment No. 10

Date: 08/12/2015

Technical Amendment

TENDER REF: HLL/PCD/PMSSY-II/NAGPUR/10/15-16 dated 07.10.2015

Schedule No. 2

Fully Automatic Coagulation Analyzer

- 1. Existing Para:** Tests available: PT, APTT, Fibrinogen, TT, LA, All Factors, ATIII, Heparin, PC, PS, PLG, AP, APCR, DD FDP, FM, vWF, etc. Also provide reagents for 500 cases of each test

Read as: Should have provision to perform PT, APTT, Fibrinogen, TT, LA, All Factors, ATIII, Heparin, PC, PS, PLG, AP, APCR, DD FDP, FM, vWF, etc. **Reagents for 1000 cases of PT and 500 cases of APTT test**

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids