



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2024/B/5733312  
Dated/दिनांक : 20-12-2024

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	30-12-2024 15:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	30-12-2024 15:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Health And Family Welfare
Department Name/विभाग का नाम	Department Of Health And Family Welfare
Organisation Name/संगठन का नाम	Hll Lifecare Limited
Office Name/कार्यालय का नाम	Hll Bhavan, Registered Office, Poojapura
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	1.5 T MRI Machine (Q2)
Minimum Average Annual Turnover of the bidder (For 3 Years)/बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का)	600 Lakh (s)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	3 Year (s)
MSE Exemption for Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Past Performance/विगत प्रदर्शन	30 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	Yes

<b>Bid Details/बिड विवरण</b>	
<b>RA Qualification Rule</b>	50% Lowest Priced Technically Qualified Bidders
<b>Type of Bid/बिड का प्रकार</b>	Two Packet Bid
<b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b>	2 Days
<b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>	No
<b>Evaluation Method/मूल्यांकन पद्धति</b>	Total value wise evaluation
<b>Financial Document Required/वित्तीय दस्तावेज की आवश्यकता है।</b>	Yes
<b>Arbitration Clause</b>	No
<b>Mediation Clause</b>	No

#### **EMD Detail/ईएमडी विवरण**

Advisory Bank/एडवाइजरी बैंक	HDFC Bank
EMD Amount/ईएमडी राशि	2400000

#### **ePBG Detail/ईपीबीजी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	14

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

#### **Beneficiary/लाभार्थी :**

General Manager (HR)  
Power Grid Corporation of India Ltd Northern Region-III Plot No. 2A/INS2 Awadh Vihar Yojana Amar Shaheed Path Lucknow, UP-226002  
(General Manager Hr)

**MII Purchase Preference/एमआईआई खरीद वरीयता**

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

**MSE Purchase Preference/एमएसई खरीद वरीयता**

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

2. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

4. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

5. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is

validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

7. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

**Excel Upload Required/एक्सेल में अपलोड किए जाने की आवश्यकता :**

BOQ MVA MCH Ghazipur - [1734686118.xlsx](#)

**1.5 T MRI Machine ( 1 pieces )**

**(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)**

**Technical Specifications/तकनीकी विशिष्टियाँ**

\* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
MAGNET	<b>Magnet Configuration</b>	Normal bore, Wide Bore
GANTRY	<b>Bore diameter at isocenter (cm)</b>	60, 70
	<b>Maximum bore length with cover (cm)</b>	140-149, 150-159, 160-169, 170-179, 180-189, 190 and above
SHIM SYSTEM	<b>Type of shimming</b>	Active, Passive, Active and Passive, Passive, linear and higher order active
PATIENT TABLE	<b>Safe patient transfer cart provided</b>	Yes, No
	<b>Table movements</b>	Vertical, Horizontal, Vertical and Horizontal
	<b>Lateral movement in Table</b>	Yes, No
GRADIENT SYSTEM	<b>Peak Amplitude with each axis independently ( mT/m)</b>	? 30, ? 40
	<b>Standard slew rate, with each axis independently (T/m/sec)</b>	? 100, ? 200

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
RF SYSTEM	<b>Minimum RF transmitter and receiver power output (KW)</b>	10, 15, 20, 25, 30
	<b>Minimum Number of independent / quadrature RF receiver channels (Acquisition in single channel and single FOV)</b>	8, 16, 32, 16 and 32 (scalable), 64, 32 with 64 (Scalable), 48, 128, 24
	<b>RF system's minimum number of elements that can be connected simultaneously in single FOV and single scan</b>	8, 16, 24, 32, 48, 64, Coil and FOV dependent
COIL SYSTEM	<b>Parallel imaging -uses redundant information acquired by multiple coil elements to significantly accelerate image acquisition The overall image time can be reduced by a factor of two or more</b>	Yes, No
	<b>Parallel imaging - Maximum parallel acquisition acceleration factor</b>	? 2, ? 4, ? 6, ? 8, ? 10, ? 16, Varies by coil
	<b>Element selection type</b>	Automatic, Manual
	<b>Head and Neck Coil - Number of elements (minimum)</b>	8, 10, 14, 16, 20, 32 Or higher
	<b>Head and Neck coil, Detachable face - a detachable front section is Intended to help reduce patient anxiety</b>	Yes, No
	<b>Head and neck Coil, Patient mirror(s) - include a small mirror that enables the patient to see outside the bore, which can reduce patient anxiety</b>	Yes, No
	<b>Spine Coil - Number of elements (minimum)</b>	12, 16, 18, 24, 32, 40 Or higher
	<b>Anterior Body coil</b>	Yes, No
	<b>Dedicated Anterior body coil with Spine Coil (body imaging at Max FOV) (Single or combination of coils) - Maximum elements used simultaneously</b>	1, 2, 3, 9, 32, 6, 12, 18, 8 Or higher
<b>Breast Coil</b>	Yes, No	

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	<b>Breast Coil - Number of elements (minimum)</b>	4, 7, 8, 16, 18, 0 Or higher
	<b>Dedicated Breast Biopsy Coil</b>	Yes, No
	<b>Spine (CTL)</b>	Yes, No
	<b>Head + Neck, Neuro-Vasculature</b>	Yes, No
	<b>Whole CNS (Head + Whole Spine)</b>	Yes, No
	<b>Chest, Heart</b>	Yes, No
	<b>Abdomen, Pelvis</b>	Yes, No
	<b>Whole Abdomen</b>	Yes, No
	<b>Peripheral Vasculature</b>	Yes, No
	<b>Prostate, Colon, Rectum, Cervix</b>	Yes, No
	<b>Dedicated Knee coil</b>	Yes, No
	<b>Knee coil - number of elements (minimum)</b>	8, 15, 0 Or higher
	<b>Dedicated Ankle coil</b>	Yes, No
	<b>Ankle coil - number of elements (minimum)</b>	0, 8, 16 Or higher
	<b>Dedicated Shoulder coil</b>	Yes, No
	<b>Shoulder coil - number of elements (minimum)</b>	0, 3, 6, 16 Or higher
	<b>Dedicated Wrist coil</b>	Yes, No
	<b>Wrist coil - number of elements (minimum)</b>	0, 8, 16 Or higher
	<b>Multi-Purpose Coils</b>	Yes, No
	<b>Temporo-Mandibular Joints</b>	Yes, No
	<b>Whole Body</b>	Yes, No
	<b>Flex coils small and large at least 4 channel</b>	Yes, No
SCANNING TECHNIQUES	<b>Standard pulse sequences (Series of RF and gradient pulses used to generate MR signals for image acquisition)</b>	Yes, No
	<b>Image acquisition - Patient movement compensated, head and body</b>	Yes, No

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	<b>Isotropic 3-D (T1) and Isotropic 3-D fast spin echo (T2)</b>	Yes, No
	<b>Fat suppressed Single breath-hold 3-D body imaging using parallel acquisition of gradient echo pulse sequences</b>	Yes, No
	<b>Single-point Dixon acquisition - this technique generates different image contrasts such as water-only, fat-only, in-phase, and out-of-phase</b>	Yes, No
	<b>Gradient Echo - An imaging sequence in which echo is created by switching a pair of dephasing and rephasing gradients such as spoiled, rewound and Steady state free precession techniques</b>	Yes, No
	<b>Spectroscopy for Breast</b>	Yes, No
	<b>Spectroscopy for Prostate</b>	Yes, No
	<b>Multi parametric study for prostate</b>	Yes, No
	<b>Bilateral Breast imaging with fat suppression with time intensity graph</b>	Yes, No
	<b>Non-contrast angiographic imaging technique (time of flight, phase contrast and SSFP) is used to visualize flow within vessels</b>	Yes, No
WORKSTATION QUANTITATIVE IMAGE ANALYSIS TOOLS	<b>fMRI tools</b>	Yes, No
	<b>Diffusion Imaging tools</b>	Yes, No
	<b>Diffusion tensor imaging tools</b>	Yes, No
	<b>Cardiac imaging tools</b>	Yes, No
WORKSTATION SPECIFICATIONS	<b>Image post processing workstation model</b>	Standalone, Client Server
	<b>Number of nodes to be provided</b>	1, 2, 3, 4, Not Applicable in case of Standalone Workstation
	<b>Number of Concurrent Licenses to be provided</b>	1, 2, 3, Not Applicable in case of Standalone Workstation

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	<b>All necessary compatible and up gradable software including post processing software for all offered application including DTI and other associated post processing like MIP, MPR, and surface reconstruction to be provided</b>	Yes, No
	<b>Software's for analyzing DWI, Spectroscopy, perfusion, CSF flow data, T1 relaxation, T2 relaxation, T2*relaxation to be provided</b>	Yes, No
	<b>System should be capable of having number of DTI directions (minimum)</b>	12, 32, 128, 256 Or higher
	<b>Cardiac package for morphology and cardiac angiography</b>	Yes, No
	<b>Output in the form of jpeg, avi /equivalent formats</b>	Yes, No
	<b>Display of cardiac cine images in movie mode with rapid avi creation in workstation</b>	Yes, No
SERVER HARDWARE	<b>Minimum Processing Power of Server</b>	20,000 concurrent slices, Not Applicable in case of Standalone Workstation, 16,000 concurrent slices
	<b>Ram Size</b>	? 32 GB, ? 64 GB, Not Applicable in case of Standalone Workstation
	<b>Monitor Size of Server</b>	? 19 inch, Not Applicable in case of Standalone Workstation
	<b>Monitor Type of Server</b>	LCD/LED/TFT, Not Applicable in case of Standalone Workstation
WORKSTATION HARDWARE	<b>Minimum Ram Size of Workstation</b>	4, 8, 16 Or higher ( <b>gigabyte</b> )
	<b>Minimum Storage capacity of Workstation</b>	500, 1000 Or higher ( <b>gigabyte</b> )
ACCESSORIES	<b>MRI Compatible auto pressure injector</b>	Yes, No
	<b>MR Compatible basic ventillator</b>	Yes, No
	<b>MR Compatible Monitor (ECG, SpO2, NIBP, RR and HR/PR) with slave display in control room</b>	Yes, No



Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	<b>MR Compatible medical O2 gas pipeline with cylinder and standard accessories</b>	Yes, No
	<b>MRI compatible trolley and wheel chair to transfer patient in and out of gantry room</b>	Yes, No
	<b>Dry laser camera with minimum resolution of 500 DPI and with three trays to accommodate three different size of films</b>	Yes, No
	<b>Online UPS for backup of at least 30 minutes for entire system</b>	Yes, No
PRE-INSTALLATION SITE REQUIREMENTS	<b>Power Requirements(KVA)</b>	30, 40, 60, 90, 120, 160 Or lower
	<b>Preparation of MRI Room, console Room , Technical room, patient preparation room, one wash room etc to be done by the seller in the area of approx 900 sq feet</b>	Yes, No
	<b>Vinyl flooring, wooden panels and good quality false ceiling in MRI room to be done by the seller</b>	Yes, No
	<b>Tile based false ceiling and floor tiling and wall tiling till false ceiling in rooms other than MRI room to be done by the seller</b>	Yes, No
	<b>Necessary modification civil work, electrical , plumbing, furnishing, air conditioning in the given area to be done by the Seller</b>	Yes, No
	<b>Doors other than main MRI room to be of good quality Aluminium doors and to be done by the seller</b>	Yes, No
	<b>RF shielding for doors , wall, glass viewer etc to be done by the seller</b>	Yes, No
WARRANTY	<b>Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b>	5, 2, 3 Or higher

**Additional Specification Parameters - 1.5 T MRI Machine ( 1 pieces )**

Specification Parameter Name	Bid Requirement (Allowed Values)
TECHNICAL SPECIFICATION	The specification provided in GEM tender document is only for reference and for actual specifications Bidders are requested to refer the Tender Document uploaded by HLL as ATC under the heading 'Buyer Added Bid Specific Terms and Conditions'.
PRODUCT COMPLIANCE SHEET	Bidder has to provide item by item compliance sheet as per attached ATC technical specification.

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Prasad K R	695012,HLL LIFECARE LTD, HLL BHAVAN, POOJAPURA, THIRUVANANTHAPURAM	1	30

**Special terms and conditions-Version:3 effective from 14-07-2023 for category 1.5 T MRI Machine**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.  
OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

## **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

### **1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 50% of the contracted Å quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

### **2. Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---