

Amendment No.6

Date: 14/10/2016

Sub: Amendment No.6 to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/11/16-17 dated: 22.08.2016.

The following changes are made to the Tender enquiry document.

Section – VII Technical Specifications

Item Sl. No. 1

Photo – Slit lamp with Applanation tonometer

1. Existing Para 12: 6-40 mm field of view.

Read as: field of view: **6-35mm or more**

2. Existing Para 16: Beam splitter.

Read as: Deleted.

3. Existing Para 17.1: Integrated camera at least 3 Megapixel high resolution

Read as: Integrated camera **3 -8 Megapixel or more resolution**

4. Existing Para 19.1: Bulbs- 06 nos.

Read as: LED light source should be covered under warranty. **2 Nos. LED bulbs should be provided.**

All other contents of the tender enquiry including terms & conditions remain unaltered.