

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/MCD-01/14-15**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	Malathion Technical	Coromandel International Ltd	Responsive		NA		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/MCD-04/14-15/TE**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	Temephos EC 50% or Equivalent	Coromandel International Ltd	Responsive		NA		Responsive
2		Paramount Pesticides Ltd	Non-Responsive	Performance is not complied as per qualification criteria of the TED	NA		Non-Responsive
3		Chemets Wets & Flows Pvt Ltd	Responsive		NA		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/MCD-06/14-15/TG**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	Temephos Granules 1%	Kilpest India Ltd	Responsive		NA		Responsive
2		Chemets Chemicals Pvt. Ltd	Responsive		NA		Responsive
3		Heranba Industries Ltd	Non-Responsive	Performance is not complied as per qualification criteria of the TED	NA		Non-Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/RIMS-03/13-14**

<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
1	C-Arm Image Intensifier	Allengers Medical System Ltd	Responsive		Responsive		Responsive
2	Monocular Microscope Lab	Quality Scientific & Mechanical Works	Responsive		Responsive		Responsive
		Radical Instruments	Responsive		Responsive		Responsive
		Dewinter Optical Inc.	Non-Responsive	Qualification criteria not fulfilled as per tender requirement.	Responsive		Non-Responsive
3	Mobile X-Ray Machine 100 mA	Skarray Healthcare Pvt. Ltd	Responsive		Responsive		Responsive
		Allengers Medical System Ltd	Responsive		Non-Responsive	Weight of the unit is 250 Kg in place of not more than 130 Kg	Non-Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/ICCONS-03/13-14**

<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
1	ICU Ventilator	Drager Medical India Pvt.Ltd	Responsive		Responsive		Responsive
		MAQUET Getinge Group	Responsive		Responsive		Responsive
2	Electrical Suction Apparatus	<b>No Bid Received</b>					
3	Multipara Monitor	Mediserve	Responsive		Responsive		Responsive
4	Pulse Oximeter	<b>No Bid Received</b>					
5	Tilt Table Electric- Pediatric	<b>No Bid Received</b>					
6	Tilt Table Electric- Adult	<b>No Bid Received</b>					
7	Wobble Cushion	<b>No Bid Received</b>					
8	Balance Pad	<b>No Bid Received</b>					
9	Adjustable Wobble board	<b>No Bid Received</b>					
10	Physiotherapy Couch Electric	<b>No Bid Received</b>					
11	Physiotherapy Couch Hydraulic	<b>No Bid Received</b>					
12	Traction Table	<b>No Bid Received</b>					
13	Muscle Stimulator	<b>No Bid Received</b>					
14	Tread Mill Medium Duty	<b>No Bid Received</b>					
15	Tread Mill - High End	<b>No Bid Received</b>					
16	Puretone audiometer - 2: Channel	Alliance Biomedical Pvt.Ltd	Non-Responsive	Tender fee is not submitted	Not evaluated due to Commercial Non-Responsiveness		Non-Responsive
		Starkey Laboratories India Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
		Alps international Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/ICCONS-03/13-14**

<b>Item No</b>	<b>Item Description</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Reason for Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Reason For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
17	Hearing aid analyser	Alps international Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
18	BERA/ASSR/OAE/VEMP	Alliance Biomedical Pvt.Ltd	Non-Responsive	Tender fee is not submitted	Not evaluated due to Commercial Non-Responsiveness		Non-Responsive
19	OAE Screener	Alliance Biomedical Pvt.Ltd.	Non-Responsive	Tender fee is not submitted	Not evaluated due to Commercial Non-Responsiveness		Non-Responsive
		Starkey Laboratories India Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
		Alps international Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
20	Impedance audiometer	Alliance Biomedical Pvt.Ltd	Non-Responsive	Tender fee is not submitted	Not evaluated due to Commercial Non-Responsiveness		Non-Responsive
		Starkey Laboratories India Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
		Alps international Pvt Ltd	Non-Responsive	BG validity is short	Responsive		Non-Responsive
21	EEG - 32 Channel	Mediserve	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No- HLL/PCD/ICCONS-03/13-14**

Item No	Item Description	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Axxonet System Technologies Pvt.Ltd	Non-Responsive	Tender validity is mentioned till 02.07.14 which is lesser than 120 days from tender opening	Non-Responsive	Manufacturer does not have ISO certification. Quoted model does not have European CE/US FDA approval	Non-Responsive
		Axis Healthcare	Non-Responsive	BG in the form of Fixed deposit not acceptable as per the TED Clause 19.3 of GIT	Responsive		Non-Responsive
		Medicaid systems	Responsive		Non-Responsive	Quoted model does not have European CE/US FDA Approval	Non-Responsive
22	Water bath	S.M.Scientific Instruments Pvt.Ltd	Responsive		Responsive		Responsive
23	Microscope binocular	<b>No Bid Received</b>					
24	Semi auto analyser	Transasia Bio-Medicals Ltd	Responsive		Responsive		Responsive
25	ISE Analyser	Transasia Bio-Medicals Ltd	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	CT Scan - 64 Slice	Wipro GE Healthcare Pvt Ltd	Responsive		Responsive		Responsive
		Erbis Engineering Co. Ltd	Responsive		Non Responsive	AERB Type approval not submitted.	Non Responsive
		Siemens Ltd	Responsive		Responsive		Responsive
		Philips Electronics India Ltd	Responsive		Responsive		Responsive
2	MRI- 1.5 Tesla	Siemens Ltd	Responsive		Responsive		Responsive
		Philips Electronics India Ltd	Responsive		Responsive		Responsive
3	ICU Beds	Surgicojn Medequip Pvt. Ltd	Non Responsive	Tender fee submitted in the form of demand draft, expired on the date of TC opening	Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		MDD Medical Systems(India) Pvt Ltd	Responsive		Responsive		Responsive
		Poct Services	Responsive		Non Responsive	Not demonstrated.	Non Responsive
		Godrej & Boyce Mfg. Co. Ltd	Responsive		Responsive for Model Chrysalis ICCU XP with single wheel central locking	Other two models not demonstrated and not having central brake as per data available in the bid.	Responsive
		Med Freshe	Responsive		Non Responsive	Demo observation: CPR not available.	Non Responsive
		Medi Works	Responsive		Non Responsive	1. Valid European CE or USFDA not submitted. 2. Back section of the bed is not completely radio translucent as per demo.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Hansraj Nayyar Medical India	Responsive		Responsive		Responsive
4	Pulse oximeter with NIBP and Central Monitor	Schiller Healthcare India Pvt. Ltd.	Non Responsive	Jurisdiction quoted only in Mumbai high court instead of Delhi as per the TED terms	Non Responsive	Battery back up is 2-4 Hrs instead of 5 Hrs.	Non Responsive
		Instromedix India Pvt. Ltd.	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		HOSPIMAX Healthcare Private Limited	Responsive		Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		DIGITEX Medical Systems(P) Ltd	Non Responsive	1.End user certificates not submitted as per the tendered quantity 2. Warranty terms not accepted	Non Responsive	1. Valid European CE or USFDA not submitted. 2. Heart rate 25 - 250 BPM in place of 18 - 300BPM	Non Responsive
5	Syringe infusion pump	B.Braun Medical (India) Pvt. Ltd	Responsive		Responsive for Model Perfusor Space	Model: Perfusor Compact Non responsive due to- 1. Flow rate: 0.1 to 99.9ml/hr instead of 0.1 to 1000ml/hr 2. Bolus rate: 800ml/hr/beep instead of 100-1500ml/hr or more	Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Allied Medical Limited	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo observation: KVO alarm can not be disabled automatically (i.e even after removal of occlusion) and have to reset the pump for infusion. - Major deviation	Non Responsive
		Instromedix India Pvt. Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo Observation: Accuracy +/-3% instead of +/-2% or better.	Non Responsive
		Medical Systems & Services	Responsive		Non Responsive	Demo observation: Bolus rate can not be programmed	Non Responsive
		My Source Innoventures Pvt. Ltd.	Non Responsive	1.End user certificates not submitted as per the tendered quantity 2. Delivery period not accepted	Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		Rohanika Electronics & Medical Systems	Responsive		Responsive		Responsive
		SS Medical Systems(India) Pvt. Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo observation: 1. KVO function can not be activated or deactivated by the user. 2. Syringe accuracy is not as per requirement.	Non Responsive



**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Kopran Laboratories Ltd	Non Responsive	1.End user certificates not submitted as per the tendered quantity 2. Validity terms not accepted	Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
7	Fibrosopic Bronchoscope - Adult	Karl Storz Endoscopy India Pvt. Ltd	Responsive		Responsive		Responsive
8	Ventilator-portable	Allied Medical Limited	Non Responsive	Not meeting the qualification criteria requirement.	Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		Aspen Diagnostic Pvt Ltd	Responsive		Responsive		Responsive
		Instromedix India Pvt Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo obseravtion: 1. High pressure oxygen port not available. 2. Turbine based airmixing technology used, it may not provide 100% Oxygen.	Non Responsive
		Covidien healthcare India Pvt Ltd	Responsive		Responsive		Responsive
		Premier Medical Systems & Devices Pvt Ltd	Responsive		Non Responsive	1. Weight 12kg instead of less than 10kg 2. Battery not available. 3. FiO2 40 -100% 4. Valid European CE or USFDA not submitted.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Rohanika Electronics & Medical Systems	Responsive		Non Responsive	PEEP: 0 -25 cmH2O instead of 0 -40cmH2O	Non Responsive
		Innovative Medicare Technologies Pvt. Ltd.	Responsive		Non Responsive	1. PEEP: 0 -25 cmH2O instead of 0 -40cmH2O 2. PS: 0 -30cmH2O instead of 0-60cmH2O	Non Responsive
		Wipro GE Healthcare Pvt Ltd	Responsive		Non Responsive	Demo obseravtion: Weight of the machine 12.5kg instead of less than 10 kg	Non Responsive
9	Ventilator high end (ICU)	Maquet Medical India Pvt Ltd	Responsive		Non Responsive	Et tube leak compensation not available.	Non Responsive
		Schiller Healthcare India Pvt Ltd	Non Responsive	Jurisdiction quoted only in Mumbai high court instead of Delhi as per the TED terms	Non Responsive	Not demonstrated due to commercial non responsiveness.	Non Responsive
		Instromedix India Pvt Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo observation: 1. 10 minutes resolution of trends instead of 5 minutes. 2. Compressor of different make (locally supplied)	Non Responsive
		Covidien healthcare India Pvt Ltd	Responsive		Non Responsive	Demo observation: Integrated EtCO2 not available.	Non Responsive
		IMT Medical AG	Responsive		Non Responsive	Demo observation: Integrated EtCO2 not available.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Rohanika Electronics & Medical Systems	Responsive		Non Responsive	Permanent flow sensors not offered.	Non Responsive
		HOSPIMAX Healthcare Private Limited	Responsive		Responsive		Responsive
		Wipro GE Healthcare Pvt Ltd	Responsive		Responsive		Responsive
11	Complete Monitoring System for ICU	Instromedix India Pvt Ltd	Responsive		Responsive		Responsive
		Spacelabs healthcare	Responsive		Non Responsive	Demo Observation: Inbuilt NMT not available.	Non Responsive
		SS Medical Systems(India) Pvt Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	1. Not demonstrated. 2. Valid European CE or USFDA not submitted.	Non Responsive
		Philips Electronics India Ltd	Responsive		Responsive		Responsive
		Nihon Kohden India Private Limited	Responsive		Responsive		Responsive
12	Portable ultrasound with colour Doppler system	Instromedix India Pvt Ltd	Responsive		Non Responsive	Tissue harmonic imaging offered without contrast media.	Non Responsive
		FUJIFILM Sonosite India Pvt Ltd	Responsive		Responsive		Responsive
		Wipro GE Healthcare Pvt Ltd	Responsive		Responsive		Responsive
13	Colour Doppler echocardiography system	Philips Electronics India Ltd.	Responsive		Responsive		Responsive
		Wipro GE Healthcare Pvt Ltd.	Responsive		Responsive		Responsive
14	Upper GI endoscope	Karl Storz Endoscopy India Pvt. Ltd	Responsive		Responsive		Responsive
		Rohit Surgical Pvt. Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Responsive		Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
15	Open care system for neonates	Phoenix Medical Systems Private Limited	Responsive		Responsive		Responsive
		AVI Healthcare Pvt Ltd	Responsive		Non Responsive	Valid European CE or USFDA Certificates not submitted.	Non Responsive
		Medi Works	Responsive		Responsive		Responsive
16	Hemodialysis machine	B. Braun Medical (India) Pvt. Ltd.	Responsive		Responsive		Responsive
		Vishal Surgical Equipment Co.	Non Responsive	EMD submitted in form of Bank guarantee, BG validity short by 1 day	Responsive		Non Responsive
17	Operation table with accessories	Rohit Surgical Pvt. Ltd.	Responsive		Non Responsive	Minimum height 740mm instead of 600 - 650mm	Non Responsive
		Bet Medical (P) Ltd.	Responsive		Non Responsive	1. Leg section adjustment: (+15 deg to -90 deg) instead of (+50 deg to -110deg)	Non Responsive
		Indian Surgical Equipment Co. Pvt. Ltd.	Responsive		Non Responsive	1. Valid European CE or USFDA not submitted. 2. Several technical points not clarified.	Non Responsive
		STERIS(India) Private Limited	Responsive		Responsive for Model STERIS Amsco 3085SP	Model: Steris EASYMAX is Non Responsive due to; Leg section adjustment: (+10 to 90 deg) instead of (+50 to -110 deg)	Responsive
		Rohanika Electronics & Medical Systems	Responsive		Non Responsive	Valid European CE or USFDA not submitted.	Non Responsive
		Hardik Medi - Tech	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Med Freshe	Responsive		Responsive		Responsive
		HOSPIMEDICA International Ltd.	Responsive		Non Responsive	1. Maximum height: 895mm instead of 1050mm. 2. Exhaustion protection and low battery warning not available.	Non Responsive
		Kopran Laboratories Ltd.	Non Responsive	1. End user certificates not submitted as per the tendered quantity 2. Validity terms not accepted	Non Responsive	1. Valid European CE or USFDA not submitted. 2. Leg section adjustment: +/- 90 deg offered instead of + 50 deg and -110deg	Non Responsive
18	Surgical Diathermy with accessories	Rohit Surgical Pvt. Ltd.	Responsive		Responsive		Responsive
		Alan Electronic Systems Pvt. Ltd.	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	1. Valid European CE or USFDA not submitted. 2. So many points not clarified.	Non Responsive
		Divlabs (India)	Responsive		Non Responsive	1. High leakage current. 2. Availability of Automatic stoppage of output in case of malfunction with audio visual alarm not specified.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Vishal Surgical Equipment Co.	Non Responsive	EMD submitted in form of Bank guarantee, BG validity short by 1 day	Non Responsive	1. Valid European CE or USFDA not submitted. 2. Not clarified availability of programmable memory. 3. Quoted model does not have facility of vessel sealing.	Non Responsive
		Imaging services	Responsive		Responsive		Responsive
19	Laparoscopic Surgery Set	Rohit Surgical Pvt. Ltd.	Responsive		Responsive		Responsive
		Karl Storz Endoscopy India Pvt. Ltd.	Responsive		Responsive		Responsive
		Imaging services	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	1. Valid European CE or USFDA not submitted. 2. Several clarifications were sought for compliance on safety features but bidder not replied.	Non Responsive
		Global Life Care Pvt Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	1. Instruments are nearly 5cm shorter than the required length of 36cm. 2. Facility of digital zoom not clarified despite asking. 3. Cusher Liver retractor not offered.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
20	Anesthesia work station	Allied Medical Limited	Responsive		Non Responsive	Demo observation: 1. Electronic gas mixing not available. 2. Built in NMT module not available.	Non Responsive
		Instromedix India Pvt Ltd	Non Responsive	End user certificates not submitted as per the tendered quantity	Non Responsive	Demo observation: 1. Graphic display for gas regulators not available. 2. Full leak compensation not available.	Non Responsive
		Draeger Medcial India Pvt. Ltd	Responsive		Responsive		Responsive
		Wipro GE Healthcare Pvt Ltd	Responsive		Responsive		Responsive
		Spacelabs healthcare	Responsive		Non Responsive	Demo observation: 1. Built in NMT module not available.	Non Responsive
		Philips Electronics India Ltd.	Responsive		Non Responsive	Demo observation: Offered model not demonstrated.	Non Responsive
21	Plasma sterilizer	CL Micromed Pvt. Ltd	Responsive		Responsive		Responsive
		STERIS(India) Private Limited	Responsive		Non Responsive	Not Plasma technology	Non Responsive
		Johnson & Johnson Limited	Responsive		Responsive		Responsive
		Medical Point (I) Ltd	Responsive		Non Responsive	Valid IEC certificate not submitted.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		STERICOL - GOA TEKNOLOJI	Non Responsive	1. Conditional warranty terms, 2. Price quoted are in FOB and it is submitted that air freight, handling and insurance will be charged at actual, 3. End user certificates not submitted as per the tendered quantity	Non Responsive	Not Plasma technology	Non Responsive
		Chengdu Laoken Technology Co. Ltd	Responsive		Responsive		Responsive
22	Ultrasonic energy cleaner(high load)	Roop Telsonic Ultrasonix Ltd	Responsive		Responsive		Responsive
23	Paediatric Cystoscope and resectoscope	Karl Storz Endoscopy India Pvt. Ltd	Responsive		Responsive		Responsive
		Eagle Medical Systems Pvt. Ltd	Non Responsive	1. Warranty terms not accepted, quoted 2 years instead of 5 years, 2. Delivery period not accepted, offered 90 days instead of 60 days	Responsive		Non Responsive
24	Ultrasonic Cutting and Coagulating Device	Aprameya Engineering (I) Pvt. Ltd.	Responsive		Responsive		Responsive



**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

<b>Item No</b>	<b>Name of Equipment</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Resaon For Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Resaon For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
25	Digital X-ray Unit (1000 mA)	Cura Healthcare Pvt. Ltd	Non Responsive	NSIC certificate submitted but as per the tender terms bidder has not quoted the equipment manufactured by the company, infact he has quoted the equipment from the different manufacturer, hence exemption is not allowed.	Non Responsive	Not complied to: Any two components out of three (X-Ray tube, X-Ray Generator and Flat panel detectors) should be from same manufacturer of the main complete system.	Non Responsive
		Carestream Health India Pvt. Ltd	Responsive		Non Responsive	Not complied to: Any two components out of three (X-Ray tube, X-Ray Generator and Flat panel detectors) should be from same manufacturer of the main complete system.	Non Responsive
		FUJIFILM India Private Limited	Responsive		Non Responsive	Not complied to: Any two components out of three (X-Ray tube, X-Ray Generator and Flat panel detectors) should be from same manufacturer of the main complete system.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/01/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Resaon For Commercial Non Responsiveness	Technical Responsiveness	Resaon For Technical Non Responsiveness	Overall Responsiveness
		Allengers Medical Systems Limited	Responsive		Non Responsive	Not complied to: Any two components out of three (X-Ray tube, X-Ray Generator and Flat panel detectors) should be from same manufacturer of the main complete system.	Non Responsive
		Wipro GE Healthcare Pvt. Ltd	Responsive		Responsive		Responsive
		Apelem-DMS Group	Non Responsive	1. Payment terms are not accepted, 2. Delivery terms not accepted	Non Responsive	Not complied to: Any two components out of three (X-Ray tube, X-Ray Generator and Flat panel detectors) should be from same manufacturer of the main complete system.	Non Responsive
		Konica Minolta Healthcare India Pvt. Ltd	Responsive		Non Responsive	AERB Type approval not submitted.	Non Responsive
		Siemens Ltd	Responsive		Non Responsive	AERB Type approval not submitted.	Non Responsive
		Philips Electronics India Ltd	Responsive		Responsive		Responsive
26	Clinical CE/IVD Flow Cytometer	Beckman Coulter India Pvt. Ltd	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	Anesthesia work station with monitor and ventilator	Wipro GE Healthcare	Responsive		Responsive		Responsive
		Philips India Limited	Responsive		Responsive		Responsive
		Drager Medical India Pvt Ltd	Responsive		Responsive		Responsive
2	Drug Cart	POCT Services	Responsive		Non Responsive	Valid CE or FDA certificate not submitted	Non Responsive
		Surgicoim Mediquip Pvt Ltd.	Responsive		Responsive		Responsive
3	Defibrillator with CPR capability	Innovative Medicare Technologies Pvt.Ltd	Responsive		Responsive		Responsive
		Biomedicon Services(I) Pvt.Ltd	Responsive		Responsive		Responsive
4	Patient warming System	Covidien healthcare India Pvt.Ltd	Responsive		Responsive		Responsive
		Indian Surgical Equipment Co. Pvt. Ltd	Responsive		Responsive		Responsive
5	Blood & Fluid Warming Device	Taurus Health Care Pvt Ltd	Responsive		Responsive		Responsive
		B L Lifesciences Pvt Ltd.	Non Responsive	Not fulfilling the qualification criteria as per the requirement of TED	Responsive		Non Responsive
		Hospimax Healthcare Private Limited	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
6	Difficult Airway Cart includes (C- track, All supraglottic Device –reusable, Tracheal Bougie, Laryngoscope of all types and sizes, McCoy Laryngoscope)	<b>No Bid Received</b>					
7	Non –Invasive cardiac support pump with defibrillator	Biomedicon Services(I) Pvt.Ltd	Responsive		Responsive		Responsive
8	Fibreoptic bronchoscope with monitor and recording facility for adult and paediatric	Karl Storz Endoscopy India Pvt. Ltd.	Responsive		Responsive		Responsive
9		Allied Medical Ltd.	Non Responsive	Not fulfilling the qualification criteria as per the requirement of TED	Non Responsive	1. Valid European CE or US FDA certificate not submitted for the quoted model. 2. Disengaged pluger not available.	Non Responsive
		Rohanika Electronics &Medical Systems	Responsive		Responsive		Responsive
		Longfian Scitech Co. Ltd	Non Responsive	Not fulfilling the qualification criteria as per the requirement of TED .	Non Responsive	Valid European CE or US FDA certificate not submitted for the quoted model.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		B.Braun Medical India Pvt Ltd	Responsive		Non Responsive	1. Valid European CE or US FDA certificate not submitted for the quoted model. 2. Flow rate 0.1 to 99.9ml/hr instead of 0.1 to 1000ml/hr 3. Bolus rate 800 ml/hr instead of programmable from 40 -1000ml.hr 4. Display of drug directory not available. 5. Key board locking system not available. 6. KVO set not available.	Non Responsive
		Instromedix (India) Pvt. Ltd.	Responsive		Non Responsive (Both the models)	<u>Model 1: Array SP 5</u> Valid European CE or US FDA certificate not submitted for the quoted model. <u>Model 2: SK-500II</u> Display of 40 drugs instead of more than 50 drugs.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Hospimax Healthcare Private Limited	Responsive		Non Responsive	1. Valid European CE or US FDA certificate not submitted for the quoted model. 2. Can not display drug names.	Non Responsive
10	Recovery trolley	Medical Point(I) Ltd	Responsive		Responsive (both models)		Non Responsive
		POCT Services	Responsive		Responsive		Responsive
		Surgicoim Mediquip Pvt Ltd.	Responsive		Responsive		Responsive
11	Recovery ward monitor	Nihon Kohden India Pvt Ltd.	Responsive		Non Responsive (Both models)	Both the offered models are non modular.	Responsive
		Wipro GE Healthcare	Responsive		Responsive		Responsive
		Philips India Limited,	Responsive		Responsive		Responsive
		Instromedix (India) Pvt. Ltd.	Responsive		Responsive		Responsive
12	Nerve Stimulator	Medical Point(I) Ltd	Responsive		Responsive		Responsive
13	ICU ventilator	VBM India Co	Responsive		Non Responsive	Non paramagnetic oxygen sensor offered instead of paramagnetic sensors	Non Responsive
		Covidien healthcare India Pvt.Ltd.	Responsive		Non Responsive	Non paramagnetic oxygen sensor offered instead of paramagnetic sensors	Non Responsive
		Air Liquide Medical Systems Pvt Ltd.	Responsive		Non Responsive	Non paramagnetic oxygen sensor offered instead of paramagnetic sensors	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Schiller Healthcare India Pvt. Ltd.	Responsive		Non Responsive	Non paramagnetic oxygen sensor offered instead of paramagnetic sensors	Non Responsive
		Wipro GE Healthcare	Responsive		Non Responsive	Compressor not from the same manufacturer.	Non Responsive
14	ICU monitor	Nihon Kohden India Pvt Ltd	Responsive		Non Responsive	Offered model is non modular type.	Non Responsive
		Wipro GE Healthcare	Responsive		Responsive		Responsive
		Philips India Limited	Responsive		Responsive		Responsive
		Instromedix (India) Pvt. Ltd.	Responsive		Responsive		Responsive
		Drager Medical India Pvt Ltd	Responsive		Responsive		Responsive
15	DVT Pump	POCT Services	Responsive		Responsive (Model 1: Flowtron Universal AC 600)	Model: 2- Flowtron ACS 800 Valid European CE or USFDA not submitted	Responsive
		Adison Equipment Co	Responsive		Responsive		Responsive
16	View Box	MEX India	Responsive		Responsive		Responsive
17	Blood Gas Analyser	Roche Diagnostics India Pvt Ltd.	Responsive		Responsive		Responsive
		Siemens Ltd	Responsive		Non Responsive	Cartridge based system offered instead of reagent based.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Resaon For Commercial Non Responsiveness	Technical Responsiveness	Resaon For Technical Non Responsiveness	Overall Responsiveness
		Vitan Medical System Pvt. Ltd	Responsive		Responsive		Responsive
18	ECG machine	Innovative Medicare Technologies Pvt.Ltd	Responsive		Responsive		Responsive
		Recorders & Medicare Systems (P) Ltd.	Responsive		Non Responsive	Lead acid battery offered instead of Li battery	Non Responsive
		Schiller Healthcare India Pvt. Ltd.	Responsive		Responsive		Responsive
		Taurus Health Care Pvt Ltd.	Responsive		Responsive		Responsive
19	Transport ventilator	Innovative Medicare Technologies Pvt.Ltd.	Responsive		Responsive		Responsive
		Covidien healthcare India Pvt.Ltd	Responsive		Responsive		Responsive
		Wipro GE Healthcare	Responsive		Non Responsive	Net weight of equipment is 12 kg instead of < 10kg	Non Responsive
		Drager Medical India Pvt Ltd	Responsive		Responsive		Responsive
20	Transport Monitor	Innovative Medicare Technologies Pvt.Ltd	Responsive		Non Responsive	1. 8.5" display instead of 10" or more 2. 2 channel recorder instead of 3 channel.	Non Responsive
		Schiller Healthcare India Pvt. Ltd.	Responsive		Non Responsive	1. Usage of monitor out side hospital not started properly. 2. 2 channel recorder instead of 3 channel.	Non Responsive
		Wipro GE Healthcare	Responsive		Responsive		Responsive



**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/03/13-14**

<b>Item No</b>	<b>Name of Equipment</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Resaon For Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Resaon For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
		Instromedix (India) Pvt. Ltd.	Responsive		Responsive		Responsive
21	Non invasive ventillator	Innovative Medicare Technologies Pvt.Ltd	Responsive		Responsive		Responsive
		Resmed India Pvt Ltd.	Responsive		Responsive (both models)		Responsive
		Drager Medical India Pvt Ltd	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
1	Analgesiometer - Tail Flick (Digital)	Marsap Services Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED, Conditional warranty .	Non Responsive	CE/ BIS not submitted ISO certificate not submitted	Non Responsive
2	Polygraph (Sixteen channel research)	MEDI WORKS	Responsive		Responsive		Responsive
		Gentech Marketing & Distribution (P) Ltd	Responsive		Responsive		Responsive
		Medicaid Systems	Responsive		Responsive		Responsive
3	Electroconvulsimeter (with ear and corneal electrodes)	Marsap Services Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	1. Stimulus current range: 0.00 to 99mA instead of 0.25 to 500mA 2. Data export to PC not available 3. ISO certificate for manufacturer not submitted	Non Responsive
		Gentech Marketing & Distribution (P) Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED .	Non Responsive	1. CE/BIS not submitted. 2. ISO certificate for manufacturer not submitted 3. Current range: 0-300mA instead of 0.25 to 500mA	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness	
		Medicaid Systems	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive	
4	Cook's Pole Climbing Apparatus	<b>No Bid Received</b>						
5	Rotarod (6 compartments)- Computerized	Marsap Services Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	ISO certificate not submitted. CE/BIS not submitted	Non Responsive	
6	Photoactometer	Marsap Services Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	ISO certificate not submitted. CE/BIS not submitted	Non Responsive	
7	Elevated Plus Maze	Marsap Services Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	CE/BIS not submitted	Non Responsive	
		Gentech Marketing & Distribution (P) Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	CE/BIS not submitted	Non Responsive	

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
9	Portable Autoclave (25L)	<b>No Bid Received</b>					
10	Incubator (20 - 100 degC)	Narang Scientific Works Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	Temperature alarm not available	Non Responsive
		S.M. Scientific Instruments Pvt. Ltd.	Responsive		Responsive		Responsive
		Surya Scientific Industries	Responsive		Responsive		Responsive
11	Digital Spirometer	Lifecare Medica	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
12	Bicycle ergometer with digital display	<b>No Bid Received</b>					
13	Double beam UV spectrophotometer	Spectrum Medical Industries Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	Clause by clause compliance not submitted. ISO certificate not submitted.	Non Responsive
14	Refrigerator -20 deg C (With boxes to store samples)	Thermo Fisher Scientific India Pvt. Ltd.	Responsive		Responsive		Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Narang Scientific Works Pvt. Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	Availability of audio visual alarm neither mentioned in catalogue nor clarified by the bidder.	Non Responsive
		Rattan Sales Corporation	Responsive		Responsive		Responsive
15	Treadmill (motorized) for humans with cardiac monitor	Lifecare Medica	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
16	Digital Reaction Time apparatus	<b>No Bid Received</b>					
17	Multiple Choice Apparatus (with digital display)	Medicaid Systems	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
18	Critical flicker fusion apparatus	Medicaid Systems	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
19	Quartenary FHPLC	Agilent Technoloies India Pvt. Ltd.	Responsive		Responsive		Responsive
		Waters (India) Private Limited	Non Responsive	Conditional warranty	Responsive		Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Spinco Biotech Pvt Ltd.	Responsive		Responsive		Responsive
20	Water purification system for HPLC	Rions Inida Lab Water Systems Pvt.	Responsive		Responsive		Responsive
21	ELISA reader (automated)	Transasia Bio-Medicals Ltd.	Non Responsive	EMD not submitted,Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
		Techcomp India Pvt. Ltd.	Responsive		Responsive		Responsive
22	Electrolyte Analyzer	CL Micromed Pvt. Ltd.	Responsive		Responsive		Responsive
		Transasia Bio-Medicals Ltd.	Non Responsive	EMD not submitted,Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
23	ECG Machine 12 channel	Taurus Healthcare Pvt. Ltd.	Responsive		Responsive		Responsive
25	Isolated Organ bath	MEDI WORKS	Responsive		Model 1: Non Responsive Model 2: Responsive	Model 1: 1. External water tank 2. Thermostat not available.	Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

Item No	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason For Commercial Non Responsiveness	Technical Responsiveness	Reason For Technical Non Responsiveness	Overall Responsiveness
		Medicaid Systems	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
26	Lab. Centrifuge Machine (Digital)	Thermo Fisher Scientific India Pvt. Ltd.	Responsive		Responsive		Responsive
27	Electronic Balance	Techcomp India Pvt. Ltd.	Responsive		Responsive		Responsive
28	Cardiac Monitor	Surbhi Mediteck	Responsive		Responsive		Responsive
		Digitex Medical Systems (P) Limited	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Responsive		Non Responsive
29	Multi Channel Pipette (Manual)	<b>No Bid Received</b>					
30	Bioelectrical Impedance Analyzer for bodycomposition	<b>No Bid Received</b>					
31	Vortex Mixer	<b>No Bid Received</b>					
32	Autoimmunostainer	Biogenex Life Sciences Pvt Ltd.	Non Responsive	Not meeting the qualification criteria as per the requirement of TED	Non Responsive	Valid European CE / USFDA not submitted. Many points not clarified.	Non Responsive

**Techno Commercial Responsiveness & Non Responsiveness Bidders for Tender Reference No-HLL/PCD/PMSSY/AIIMS-II/09/13-14**

<b>Item No</b>	<b>Name of Equipment</b>	<b>Name of Bidder</b>	<b>Commercial Responsiveness</b>	<b>Resaon For Commercial Non Responsiveness</b>	<b>Technical Responsiveness</b>	<b>Resaon For Technical Non Responsiveness</b>	<b>Overall Responsiveness</b>
33	Bench top Flow cytometer			<b>No Bid Received</b>			
34	FAST WESTERN BLOT APPARATUS			<b>No Bid Received</b>			
35	Immuno Assay Fully Automated Singe Dose Entry			<b>No Bid Received</b>			