

13th December 2018**Amendment No. 1****Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/08/18-19 Dated 28.11.2018.**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII
Technical Specification****Sch. No. I/01
SPHYGMOMANOMETER ANEROID (Rfx no. 3000003577)**

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 45 Para 10	Approved Make: Welch Allyn, Diamond, Omron, Richter	Approved Make: Welch Allyn, Diamond, Omron, Riester

**Sch. No. I/03
Stethoscope (Rfx no. 3000003579)**

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 45 Para 4	Chest piece with Diaphragm & bell- Ø45mm(adult) , - Ø35mm (paediatric)	Chest piece with Diaphragm & bell- Ø45mm (adult) , - Ø35mm (paediatric) with +/- 10% tolerance is acceptable.
2	Pg 45 Para 8	Approved Make: Welch Allyn, Heine, Littman	Approved Make: Welch Allyn, Heine, Littman, Riester

**Sch. No. I/ 06
X ray Viewer (LED) 2 Panel (Rfx no. 3000003582)**

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 46 Para 9	It should be CE/ FDA/BIS certified product	It should be CE/ FDA/BIS/ ISO 13485 certified product

**Sch. No. I/07
Automatic Blood Pressure Monitor on Roller Stand (Rfx no. 3000003583)**

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 47 Para 9	Approved Make: Welch Allyn, Heine, Littman	Approved Make: Welch Allyn, Heine, Littman, Riester

**Sch. No. I/ 08
Glucometer (Rfx no. 3000003584)**

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
--------	--------------------	----------------------	---------

1	Pg 47 Para 1.1	Should have precise digital temperature control with selectable range of 37degree celsius to 42 degree Celsius	Deleted
2	Pg 47 Para 1.2	Measuring range should be 20 to 400 mg/Dl	Measuring range should be 20 to 600 mg/Dl
3			Added Para: Unit Price of the strip should be quoted separately and it will be valid for 2 years.

Sch. No. I/10
Exam Lamp - Mobile (Rfx no. 3000003586)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 48 Para 2.1	Manufacturer should have ISO certification	Manufacturer should have ISO 13485 certification

Sch. No. I/16
Digital Haemoglobinometer Portable (For Medicine & OBG) (Rfx no. 3000003592)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 49 Para 3	The measurement device should use cuvettes to collect blood sample for analysis.	The measurement device should use cuvettes / strips to collect blood sample for analysis.
2	Pg 49 Para 7	1000 cuvettes should be supplied with the initial purchase of equipment	1000 cuvettes / strips should be supplied with the initial purchase of equipment. Unit Price of the strip / cuvettes should be quoted separately and it will be valid for 2 years.

Sch. No. II/02
Otoscope (Rfx no. 3000003595)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Para 1.6	Should have in built rechargeable battery with charger	Should have in built rechargeable battery with desktop charger

Sch. No. II/03
Head Light (Rfx no. 3000003596)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 45 Para 7	Available with rechargeable battery option	Available with rechargeable battery option, with minimum 10 hours battery backup (by single or multiple battery. Charging may be separately through charger)

Sch. No. III/03
STREAK RETINOSCOPE (Rfx no. 3000003600)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
--------	--------------------	----------------------	---------

1	Pg 53 Para 1.2	Should have crossed-linear polarizing filter, optional orange filter	Should preferably have crossed-linear polarizing filter, optional orange filter
2	Pg 53 Para 1.5	Should use LED streak lamp.	Should use LED/ Xenon streak lamp. (If Xenon offered, one spare xenon lamp should be provided)

Sch. No. III/08
Portable Autoclave (Rfx no. 300003605)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 57 Para 23	The control system should be microprocessor based PLC system specially designed for sterilization applications. The control system should have CPU processor. Digital input/output controls, analog measuring inputs & COM ports for printer.	The control system should be microprocessor based / PLC based system specially designed for sterilization applications. The control system should have CPU processor. Digital input/output controls, analog measuring inputs & COM ports for printer.

Sch. No. III/10
Ophthalmoscope for EYE OPD (Rfx no. 300003607)

Sl. No	Tender Page & Para	TENDER SPECIFICATION	Read As
1	Pg 58	Approved Make: Welch Allyn, Heine, Keeler	Approved Make: Welch Allyn, Heine, Keeler, Riester

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids