

Amendment No. 1

Date: 08/07/2014

Subject: Amendment to the Tender Enquiry Document

Ref: NIT No: HLL/PCD/GNCTD/04/GBPH/14-15 dated 29.05.2014

The following changes have been incorporated in the referred tender .

Section – VII Technical Specifications

Item No.1

Multi headed (5 headed) microscope

1. ***Existing Para 13:*** Trinocular tube 22mm/25mm for camera attachment with appropriate C-mount relay for attachment of projection system with provision of light sending to eyepiece /camera as full/50%-50%

Read as: Trinocular tube 22mm/25mm for camera attachment with appropriate C-mount relay for attachment of projection system with provision of light sending to eyepiece /camera as full **"50%/50% or 80%/20%"**

2. ***Added Para:***

- i) **The microscope, camera and software should be from the same manufacturer.**
- ii) **The model offered should be European CE and/or USFDA approved.**

Item No.3

Fully Automated Endoscope Reprocessor (Endowasher)

1. ***Added Para:***

The model offered should be European CE and/or US-FDA approved.

All other contents of the tender enquiry including terms & conditions remain unaltered.