

# **TENDER ENQUIRY DOCUMENT**

## **FOR PURCHASE OF MEDICAL EQUIPMENT**

On behalf of

ALL INDIA INSTITUTE OF MEDICAL SCIENCES Bhubaneswar,  
(Odisha) – 751 019 & Institutes under PMSSY Upgradation

**HITES/PCD/AIIMS-BBSR/03/18-19**

*Through*



## **HLL INFRA TECH SERVICES LIMITED**

(Subsidiary of HLL Lifecare Ltd., a Govt. of India Enterprise)

B-14 A, Sector-62, Noida - 201 307

Phone: 0120-4071500; Fax: 0120-4071513

URL: [www.hllhites.com](http://www.hllhites.com)

Email: [pcd@hllhites.com](mailto:pcd@hllhites.com)

**INDEX**

<b>Section</b>	<b>Topic</b>	<b>Page No.</b>
Section I	– Notice inviting e-Tender (e-NIT) -----	03
Section II	– General Instructions to Tenderers (GIT) -----	06
Section III	– Special Instructions to Tenderers (SIT) -----	27
Section IV	– General Conditions of Contract (GCC) -----	28
Section V	– Special Conditions of Contract (SCC) -----	44
Section VI	– List of Requirements -----	45
Section VII	– Technical Specifications -----	48
Section VIII	– Quality Control Requirements -----	79
Section IX	– Qualification Criteria -----	80
Section X	– Tender Form -----	82
Section XI	– Price Schedules -----	83
Section XII	– Questionnaire -----	84
Section XIII	– Bank Guarantee Form for EMD -----	85
Section XIV	– Manufacturer’s Authorisation Form -----	86
Section XV	– Bank Guarantee Form for Performance Security /CMC Security -----	87
Section XVI	– Contract Form (A & B) -----	88
Section XVII	– Proforma of Consignee Receipt Certificate -----	92
Section XVIII	– Proforma of Final Acceptance Certificate by the Consignee -----	93
Section XIX	– Consignee List -----	95
Appendix A	– Public Procurement (Preference to Make in India), Order, 2017-----	96
Appendix B	– Integrity Pact -----	102

**SECTION I**  
**NOTICE INVITING TENDER (NIT)**

**Tender Enquiry No.: HITES/PCD/AIIMS-BBSR/03/18-19****Dated: 29.09.2018**

Procurement & Consultancy Services Division of **HLL Infra Tech Services Limited (HITES)**, a fully owned subsidiary of HLL Lifecare Ltd. (HLL), for and on behalf of ALL INDIA INSTITUTE OF MEDICAL SCIENCES Bhubaneswar, (Odisha), invites sealed tenders, from eligible and qualified tenderers for supply of Medical Equipments of ALL INDIA INSTITUTE OF MEDICAL SCIENCES Bhubaneswar, (Odisha) & Institutes under PMSSY Upgradation.

SI No	RFx No	Equipment	Department	Quantity	Tender Processing Fee (INR)*	EMD (INR)
1	3000003442	Grossing Station-Advance Model	Pathology & Lab Medicine	1	2360	1,00,000
2	3000003443	Automated Media Pouring Station Including Stacker and Perparator	Microbiology	1	1180	82,000
3	3000003444	Fully Automatic Chemiluminescence Analyser	Microbiology	1	2360	1,00,000
4	3000003445	Advance Robotic Rehabilitation System	Physical Medicine & Rehabilitation (PMR)	1	3540	9,80,000
5	3000003446	Automated Platelet Aggregometer	Transfusion Medicine & Blood Bank (TM& BB)	1	1180	70,000
6	3000003447	Refrigerated Centrifuge	Transfusion Medicine & Blood Bank (TM& BB)	1	1180	80,000
7	3000003448	532nm Green laser with slit lamp (Photoocoagulative)	Ophthalmology	1	1180	80,000
8	3000003449	Immunoassay Analyser	Biochemistry	1	2360	1,00,000
9	3000003450	Echocardiography machine	Anesthesiology	1	2360	1,00,000
10	3000003451	Surgical Microscope	Burn & Plastic Surgery	1	3540	3,40,000
11	3000003452	Transcranial Magnetic Stimulator (TMS) with paired Pulse	Neurology	1	2360	1,80,000
12	3000003453	Automatic Function Test (AFT) with Tilt Table	Physiology	1	2360	1,46,000
13	3000003454	128 Channel EEG ERP System	Physiology	1	2360	1,96,000
14	3000003455	Dual Energy X-Ray Absorptiometry	Endocrinology	1	3540	2,40,000
15	3000003456	Laparoscopic Surgery Set	General Surgery	1	2360	1,50,000

**Note:\* Tender processing Fee is inclusive of GST @18% (Our GSTIN: 09AADCH4882R1ZP)**

## (2) Tender timeline:

Sl. No.	Description	Schedule
a.	Last date for receipt of Pre-bid queries	10.10.2018, 06:00 PM
b.	<b>Pre-bid meeting date, time</b>	12.10.2018, 11:00 AM
c.	Closing date & time for submission of online bids	12.11.2018, 01:00 PM
d.	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	12.11.2018, 02:00 PM
e.	Time and date of opening of online bids	12.11.2018, 02:30 PM
f.	Venue for :- <ul style="list-style-type: none"> <li>• Pre-bid meeting</li> <li>• Submission of tender processing fee, EMD in physical form.</li> <li>• Tender Opening-Tech Bid</li> </ul>	HLL Infra Tech Services Limited, Procurement & Consultancy Services Division, B-14 A, Sector-62, Noida-201307

\* Bidders have to submit Original Bank Instruments for tender processing fee and EMD/ documentary proof for EMD exemption within the above mentioned date and time

**SPECIFIC Instructions for e-Tender Participation:-**

3. Bidders should have valid Class 3-B Digital Signature Certificate with encryption.
4. Bidders are requested to read the bidders help document on e-tender web site link before proceeding for bidding.
5. The prospective bidders have to register with the E-procurement system of HLL at <https://etender.lifecarehll.com/irj/portal>. On completion of the registration process, the bidders will be provided user ID and password within 48 hours (excluding non-working days). In order to submit the bids electronically, bidders are required to have a valid Class 3-B Digital Signature Certificate (signing and encryption/ decryption certificates).
6. Post receipt of User ID & Password, Bidders can log on for downloading & uploading tender document.
7. The tenderers shall submit Tender Processing Fee and EMD in physical form at the scheduled time and venue.
8. Tenderer may download the tender enquiry documents from the web site [www.hllhites.com](http://www.hllhites.com) or [www.lifecarehll.com](http://www.lifecarehll.com) or [www.eprocure.gov.in/cppp](http://www.eprocure.gov.in/cppp) or <https://etender.lifecarehll.com/irj/portal>.
9. The bidders shall submit the required Tender Processing Fee (in form of Demand Draft or Banker's Cheque) and EMD (as per GIT clause no. 19.3) in physical form in favour of 'HLL Infra Tech Services Limited' at the scheduled time and venue. **Tender processing Fee is required from all the bidders irrespective of their registration with NSIC or any other Govt. Organisation.**
10. All the tender related documents to be scanned in .pdf format with lower resolution and 100% readability and submitted online. The bidders shall not submit any other documents in physical form other than the documents mentioned at point no 9 above.
11. Prospective bidders may send their queries 02 (two) days before the pre-bid meeting so that they can be studied and addressed during pre-bid meeting. Query can also be raised during pre-bid meeting. No queries/ representations will be entertained after pre-bid meeting
12. All prospective tenderers may attend the Pre Tender meeting. The venue, date and time indicated above.
13. Bidders shall ensure that their bids complete in all respects, are submitted **online through HLL's e-portal (as described above) ONLY. No DEVIATION is acceptable.**
14. Bidders may simulate bid submission (technical & financial) at least one week in advance of the bid submission deadline. No clarifications/troubleshooting regarding any problems being faced during online bid submission shall be entertained in the last week of bid submission

**IMPORTANT NOTE:-**

**Tender Processing Fee and EMD** (as applicable) should be deposited within the scheduled date & time in the Tender Box located at:

**HLL Infra Tech Services Limited,  
Procurement and Consultancy Division,  
B-14 A, Sector-62, Noida-201307, Uttar Pradesh**

**CEO  
HLL Infra Tech Services Limited**

**SECTION - II****GENERAL INSTRUCTIONS TO TENDERERS (GIT)****CONTENTS**

<b>Sl. No.</b>	<b>Topic</b>	<b>Page No.</b>
<b>A</b>	<b>PREAMBLE</b>	
1	Definitions and Abbreviations	8
2	Introduction	9
3	Availability of Funds	9
4	Language of Tender	10
5	Eligible Tenderers	10
6	Eligible Goods and Services	10
7	Tendering Expense	10
<b>B</b>	<b>TENDER ENQUIRY DOCUMENTS</b>	
8	Contents of Tender Enquiry Documents	10
9	Amendments to Tender Enquiry Documents	11
10	Clarification of e-Tender Enquiry Documents	11
<b>C</b>	<b>PREPARATION OF TENDER</b>	
11	Documents Comprising the Tender	11
12	Tender Currencies	13
13	Tender Prices	13
14	Indian Agent	15
15	Firm Price	16
16	Alternative Tenders	16
17	Documents Establishing Tenderer's Eligibility and Qualifications	16
18	Documents Establishing Good's Conformity to Tender Enquiry Document	16
19	Earnest Money Deposit (EMD)	17
20	Tender Validity	17
21	Digital Signing of Tender	18
<b>D</b>	<b>SUBMISSION OF TENDERS</b>	
22	Submission of e-Tenders	18
23	Late Tender	19
24	Alteration and Withdrawal of e-Tender	19
<b>E</b>	<b>TENDER OPENING</b>	
25	Opening of e-Tenders	19
<b>F</b>	<b>SCRUTINY AND EVALUATION OF TENDERS</b>	
26	Basic Principle	20
27	Scrutiny of Tenders	20

<b>Sl. No.</b>	<b>Topic</b>	<b>Page No.</b>
28	Minor Infirmary/Irregularity/Non-Conformity	21
29	Discrepancy in Prices	21
30	Discrepancy between original and copies of Tender	21
31	Qualification Criteria	22
32	Conversion of Tender Currencies to Indian Rupees	22
33	Schedule-wise Evaluation	22
34	Comparison of Tenders	22
35	Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders	22
36	Tenderer's capability to perform the contract	25
37	Contacting the Purchaser	25
<b>G</b>	<b>AWARD OF CONTRACT</b>	
38	Purchaser's Right to Accept any Tender and to Reject any or All Tenders	25
39	Award Criteria	25
40	Variation of Quantities at the Time of Award	25
41	Notification of Award	25
42	Issue of Contract	26
43	Non-receipt of Performance Security and Contract by the Purchaser/Consignee	26
44	Return of EMD	26
45	Publication of Tender Result	26
46	Corrupt or Fraudulent Practices	26

## A. PREAMBLE

### 1. Definitions and Abbreviations

1.1 The following definitions and abbreviations, which have been used in these documents shall have the meanings as indicated below:

#### 1.2. Definitions:

- (i) **“Purchaser”** means ALL INDIA INSTITUTE OF MEDICAL SCIENCES Bhubaneswar, (Odisha) – 751 019 & Institutes under PMSSY Upgradation.
- (ii) **“e-Tender”** means Bids / Quotation / Tender received from a Firm / Tenderer / Bidder online.
- (iii) **“Tenderer”** means Bidder/the Individual or Firm submitting Bids/Quotation/e-Tenders.
- (iv) **“Supplier”** means the individual or the firm supplying the goods and services as incorporated in the contract.
- (v) **“Goods”** means the articles, material, commodities, livestock, furniture, fixtures, raw material, spares, instruments, machinery, equipment, medical equipment, industrial plant etc. which the supplier is required to supply to the purchaser under the contract.
- (vi) **“Services”** means services allied and incidental to the supply of goods, such as transportation, installation, commissioning, provision of technical assistance, training, after sales service, maintenance service and other such obligations of the supplier covered under the contract.
- (vii) **“Earnest Money Deposit” (EMD)** means Bid Security/ monetary or financial guarantee to be furnished by a tenderer along with its tender.
- (viii) **“Contract”** means the written agreement entered into between the purchaser and/or consignee and the supplier, together with all the documents mentioned therein and including all attachments, annexure etc. therein.
- (ix) **“Performance Security”** means monetary or financial guarantee to be furnished by the successful tenderer for due performance of the contract placed on it. Performance Security is also known as Security Deposit.
- (x) **“Consignee”** means the Hospital/Institute/Medical College/ person to whom the goods are required to be delivered as specified in the Contract. If the goods are required to be delivered to a person as an interim consignee for the purpose of despatch to another person as provided in the Contract then that “another” person is the consignee, also known as ultimate consignee.
- (xi) **“Specification”** means the document/standard that prescribes the requirement with which goods or service has to conform.
- (xii) **“Inspection”** means activities such as measuring, examining, testing, gauging one or more characteristics of the product or service and comparing the same with the specified requirement to determine conformity.
- (xiii) **“Day”** means calendar day.
- (xiv) **“Local supplier”** means a supplier or service provider whose product or service offered for procurement meets the minimum local content as prescribed under this Order or by the competent Ministries/ Departments in pursuance of this order.
- (xv) **“Local content”** means the amount of value added in India which shall, unless otherwise prescribed by the Nodal Ministry, be the total value of the item procured excluding net domestic indirect taxes) minus the value of imported content in the item (including all customs duties) as a proportion of the total value in percent.
- (xvi) **Margin of purchase preference’** means the maximum extent to which the price quoted by a local supplier may be above the L1 for the purpose of purchase preference.

#### 1.3 Abbreviations:

- (i) **“TE Document”** means Tender Enquiry Document
- (ii) **“NIT”** means Notice Inviting Tenders.



- (iii) “GIT” means General Instructions to Tenderers
- (iv) “SIT” means Special Instructions to Tenderers
- (v) “GCC” means General Conditions of Contract
- (vi) “SCC” means Special Conditions of Contract
- (vii) “DGS&D” means Directorate General of Supplies and Disposals
- (viii) “NSIC” means National Small Industries Corporation
- (ix) “PSU” means Public Sector Undertaking
- (x) “CPSU” means Central Public Sector Undertaking
- (xi) “LSI” means Large Scale Industry
- (xii) “SSI” means Small Scale Industry
- (xiii) “LC” means Letter of Credit
- (xiv) “DP” means Delivery Period
- (xv) “BG” means Bank Guarantee
- (xvi) “CD” means Custom Duty
- (xvii) “RR” means Railway Receipt
- (xviii) “BL” means Bill of Lading
- (xix) “FOB” means Free on Board
- (xx) “FCA” means Free Carrier
- (xxi) “FOR” means Free On Rail
- (xxii) “CIF” means Cost, Insurance and Freight
- (xxiii) “CIP (Destinations)” means Carriage and Insurance Paid up to named port of destination. Additionally the Insurance (local transportation and storage) would be extended and borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery.
- (xxiv) “DDP” means Delivery Duty Paid named place of destination (consignee site)
- (xxv) “INCOTERMS” means International Commercial Terms as on the date of Tender Opening
- (xxvi) ”MOH&FW” means Ministry of Health & Family Welfare, Government of India
- (xxvii) “Dte. GHS” means Directorate General and Health Services, MOH&FW.
- (xxviii) “CMC” means Comprehensive maintenance Contract (labour, spare and preventive maintenance)
- (xxix) “RT” means Re-Tender.
- (xxx) “GST” means Goods and Services Tax

## 2. Introduction

- 2.1 The Purchaser has issued these TE documents for purchase of goods and related services on behalf of MoHFW, Govt of India as mentioned in Section – VI – “List of Requirements”, which also indicates, *inter alia*, the required delivery schedule, terms and place of delivery.
- 2.2 This section (Section II - “General Instruction Tenderers”) provides the relevant information as well as instructions to assist the prospective bidders in preparation and submission of bids. It also includes the mode and procedure to be adopted by the purchaser for receipt and opening as well as scrutiny and evaluation of tenders and subsequent placement of contract.
- 2.3 The tenderers shall also read the Special Instructions to Tenderers (SIT) related to this purchase, as contained in Section III of these documents and follow the same accordingly. Whenever there is a conflict between the GIT and the SIT, the provisions contained in the SIT shall prevail over those in the GIT.
- 2.4 Before formulating the tender and submitting the same to the purchaser, the tenderer should read and examine all the terms, conditions, instructions, etc. contained in the TE documents. Failure to provide and/or comply with the required information, instructions etc. incorporated in these TE documents may result in rejection of its tender.

## 3. Availability of Funds

- 3.1 Expenditure to be incurred for the proposed purchase will be met from the funds available with the purchaser/ consignee.

#### 4. Language of Tender

- 4.1 The tender submitted by the tenderer and all subsequent correspondence and documents relating to the tender exchanged between the tenderer and the purchaser, shall be written in the English language, unless otherwise specified in the Tender Enquiry. However, the language of any printed literature furnished by the tenderer in connection with its tender may be written in any other language provided the same is accompanied by an English translation and, for purposes of interpretation of the tender, the English translation shall prevail.
- 4.2 The tender submitted by the tenderer and all subsequent correspondence and documents relating to the tender exchanged between the tenderer and the purchaser, may also be written in the Hindi language, provided that the same are accompanied by English translation, in which case, for purpose of interpretation of the tender etc., the English translations shall prevail.

#### 5. Eligible Tenderers

- 5.1 This invitation for tenders is open to all suppliers who fulfil the eligibility criteria specified in these documents.

#### 6. Eligible Goods and Services

- 6.1 All goods and related services to be supplied under the contract shall have their origin in India or any other country with which India has not banned trade relations. The term “origin” used in this clause means the place where the goods are mined, grown, produced, or manufactured or from where the related services are arranged and supplied.

#### 7. Tendering Expense

- 7.1 The tenderer shall bear all costs and expenditure incurred and/or to be incurred by it in connection with its tender including preparation, mailing and submission of its tender and for subsequent processing the same. The purchaser will, in no case be responsible or liable for any such cost, expenditure etc. regardless of the conduct or outcome of the tendering process.

### B. e-TENDER ENQUIRY DOCUMENTS

#### 8. Content of Tender Enquiry Documents

- 8.1 In addition to Section I – “Notice inviting e-Tender” (NIT), the TE documents include:

- Section II – General Instructions to Tenderers (GIT)
- Section III – Special Instructions to Tenderers (SIT)
- Section IV – General Conditions of Contract (GCC)
- Section V – Special Conditions of Contract (SCC)
- Section VI – List of Requirements
- Section VII – Technical Specifications
- Section VIII – Quality Control Requirements
- Section IX – Qualification Criteria
- Section X – Tender Form
- Section XI – Price Schedules
- Section XII – Questionnaire
- Section XIII – Bank Guarantee Form for EMD
- Section XIV – Manufacturer’s Authorisation Form
- Section XV – Bank Guarantee Form for Performance Security/CMC Security
- Section XVI – Contract Forms A & B
- Section XVII – Proforma of Consignee Receipt Certificate
- Section XVIII – Proforma of Final Acceptance Certificate by the consignee
- Section XIX – Consignee List
- Appendix A – DIPP – Public Procurement (Preference to Make in India), Order 2017**
- Appendix B – Integrity pact**

8.2 The relevant details of the required goods and services, the terms, conditions and procedure for tendering, tender evaluation, placement of contract, the applicable contract terms and, also, the standard formats to be used for this purpose are incorporated in the above-mentioned documents. The interested tenderers are expected to examine all such details etc. to proceed further.

## **9. Amendments to TE documents**

9.1 At any time prior to the deadline for submission of tenders, the purchaser may, for any reason deemed fit by it, modify the TE documents by issuing suitable amendment(s) to it.

9.2 Such an amendment will be notified in writing by registered/speed post or by fax/telex/e-mail, to all prospective tenderers, who have received the TE documents and will be binding on them.

9.3 In order to provide reasonable time to the prospective tenderers to take necessary action in preparing their tenders as per the amendment, the purchaser may, at its discretion extend the deadline for the submission of tenders and other allied time frames, which are linked with that deadline.

## **10. Clarification of TE documents**

10.1 A tenderer requiring any clarification or elucidation on any issue of the TE documents may take up the same with the purchaser in writing on their letter head duly signed and scanned through email to [pcd@hllhites.com](mailto:pcd@hllhites.com) and [bmenoida@hllhites.com](mailto:bmenoida@hllhites.com). The purchaser will respond to such request provided the same is received by the purchaser **within the due date mentioned in the NIT. Any queries/ representations received later shall not be taken into cognizance.**

## **C. PREPARATION OF e-TENDERS**

### **11. Documents comprising the e-Tender**

11.1 The tender(s) shall only be submitted online as mentioned below:

- (i) Technical Bid (Consisting of Techno-Commercial bids in excel format provided with the tender enquiry along with the supporting documents i.e. scanned copies of Tender Processing Fee, EMD, Eligibility Criteria & Technical Specifications viz. Product Specification Sheets/ Brochures, OEM Certificate, etc.) has to be attached in the C-folder of e-tendering module. Bidders have to ensure that the documents uploaded in pdf format are legible.
- (ii) Price Bid has to be submitted in the prescribed excel format provided with the tender enquiry.

#### Note:

- (i) The Tender Processing Fee and EMD, in favor of HLL Infra Tech Services Ltd, are to be submitted in physical form as per Section - I, Notice Inviting Tender, of this tender enquiry.
- (ii) The bidders have to follow the steps listed in *Bidding Manual – Attachment Mode* available in the *Bidder Help Documents* of e-tender portal login screen for uploading the Techno-Commercial Bid.

### **A) Details of Technical Tender (Un priced Tender)**

**Bidders shall furnish the following information along with technical tender:.**

- i) Techno-Commercial Bid in excel format provided with the tender enquiry

- ii) Earnest money Deposit (EMD) furnished in accordance with GIT clause 19.1 alternatively, documentary evidence as per GIT clause 19.2 for claiming exemption from payment of earnest money.
- iii) Tender Form as per Section X (without indicating any prices).
- iv) Documentary evidence, as necessary in terms of clauses 5 and 17 establishing that the tenderer is eligible to submit the tender and, also, qualified to perform the contract if its tender is accepted.
- v) Tenderer/Agent who quotes for goods manufactured by other manufacturer shall furnish Manufacturer's Authorization **strictly as per the prescribed format (Section - XIV)**.
- vi) Power of Attorney issued by Competent Authority in favour of the person **who is digitally signing/ uploading the tender(s)**.
- vii) Documents and relevant details to establish in accordance with GIT clause 18 that the goods and the allied services to be supplied by the tenderer conform to the requirement of the TE documents.
- viii) Performance Statement as per section IX along with relevant copies of orders and end users' satisfaction certificate.
- ix) Price Schedule(s) as per Section XI filled up with all the details including Make, Model etc. of the goods offered with prices blank (without indicating any prices).
- x) Certificate of Incorporation.
- xi) Self-Attested copies of GST registration certificate and PAN Card.
- xii) Non conviction /no pending conviction certification issued by Notary on judicial stamp paper for preceding three years.
- xiii) Self-Attested copies of quality certificates i.e. US FDA /CE/ BIS Certificate issued by competent authority, if applicable.
- xiv) Documentary evidence stating the status of bidder.
- xv) List of procurement agencies of repute to which the tendered product have been supplied during last 12 months.
- xvi) Self-attested copies of annual report, audited balance sheet and profit & loss account for preceding three years from the date of tender opening.
- xvii) Notarized affidavit that tenderer does not have any relation with the person authorized to evaluate technically or involve in finalizing the tender or will decide the use of tendered items.
- xviii) A self-declaration on Rs. 10/- non-judicial Stamp Paper that the rates quoted in the tender are the lowest and not quoted less than this to any Government Institution (State/Central/ other Institute in India).
- xix) Copies of original product catalogues/ data sheet must be enclosed of all quoted items.
- xx) A tenderer quoting imported goods located within India shall produce documentary evidence of the goods having been imported and already located within India (i.e. Bills of Entry for the quoted items and a self-declaration confirming that the quoted items were imported for the purpose of storage in bidder warehouse and for further sale), along with their techno-commercial bid.
- xxi) ***The Integrity pact (At Appendix-B) shall be a part and parcel of this document and has to be signed by bidder(s) at the pre-tendering stage itself, as a pre-bid obligation and should be submitted along with the Techno-Commercial Bids. All bidders are bound to comply with the integrity pact clauses. Bids submitted without signing the integrity pact will be ab initio rejected without assigning any reason.***

**B) Price Bid:**

Prices are to be quoted in the prescribed Price Bid format in excel provided along with the tender enquiry in the e-tender portal. The price should be quoted for the accounting unit indicated in the e-tender document.

**Note:**

- (i) **The bidder has to be diligent while filling up the Techno-Commercial Bid and Price Bid provided in excel formats and must not tamper with the contents of the sheets.**
  - (ii) It is the responsibility of bidder to go through the TE document to ensure furnishing all required documents in addition to above, if any.
  - (iii) The bidders have to follow the steps listed in *Bidding Manual – Attachment Mode* available in the *Bidder Help Documents* of e-tender portal login screen for uploading the Price Bid.
- 11.2 A person signing (manually or digitally) the tender form or any documents forming part of the contract on behalf of another shall be deemed to warrant that he has authority to bind such other persons and if, on enquiry, it appears that the persons so signing had no authority to do so, the purchaser may, without prejudice to other civil and criminal remedies, cancel the contract and hold the signatory liable for all cost and damages.
- 11.3 A bid, which does not fulfill any of the above requirements and/ or give evasive information/reply against any such requirement, shall be liable to be ignored.
- 11.4 Tender sent by fax/telex/cable shall be ignored.

**12. Tender currencies**

- 12.1 The tenderer supplying indigenous goods or already imported goods shall quote only in Indian Rupees (INR).
- 12.2 For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currency say US Dollar, Euro, GBP or Japanese Yen. As regards price(s) for allied services, if any required with the goods, the same shall be quoted in Indian Rupees only (INR), if such services are to be performed /undertaken in India. Commission for Indian Agent, if any and if payable shall be indicated in the space provided for in the price schedule and will be payable in Indian Rupees only.
- 12.3 **Tenders, where prices are quoted in any other currency may not be accepted and are liable to be ignored.**
- 12.4 A tenderer quoting imported goods located within India shall produce documentary evidence of the goods having been imported and already located within India (i.e. Bills of Entry for the quoted items and a self-declaration confirming that the quoted items were imported for the purpose of storage in bidder warehouse and for further sale), along with their techno-commercial bid.

**13 Tender Prices**

- 13.1 The Tenderer shall indicate on the Price Schedule provided under Section XI all the specified components of prices shown therein including the unit prices and total tender prices of the goods and services it proposes to supply against the requirement. All the columns shown in the price schedule should be filled up as required.
- 13.2 If there is more than one schedule in the List of Requirements, the tenderer has the option to submit its quotation for any one or more schedules. However, while quoting for a schedule, the tenderer shall quote for the complete requirement of goods and services as specified in that particular schedule.
- 13.3 The quoted prices for goods offered from within India and that for goods offered from abroad are to be indicated separately in the applicable Price Schedules attached under Section XI.
- 13.4 While filling up the columns of the Price Schedule, the following aspects should be noted for compliance:
- 13.4.1 For domestic goods or goods of foreign origin located within India, the prices in the corresponding price schedule shall be entered separately in the following manner:

- a) The price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like, Custom Duty and/or GST already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc;
- b) Any taxes and duties including Custom duty and/or GST, which will be payable on the goods in India if the contract is awarded;
- c) Charges towards Packing & Forwarding, Inland Transportation, Insurance (local transportation and storage), Loading & Unloading etc. would be borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Price Schedule;
- d) The price of Incidental Services, as mentioned in List of Requirements and Price Schedule;
- e) The prices of Site Modification Work (if any), as mentioned in List of Requirements, Technical Specification and Price Schedule.
- f) The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Price Schedule.

13.4.2 For goods offered from abroad, the prices in the corresponding price schedule shall be entered separately in the following manner:

- a) The price of goods quoted FOB/FCA port of shipment, as indicated in the List of Requirements and Price Schedule;
- b) Price of goods quoted CIP (name port of destination) in India as indicated in the List of Requirements, Price Schedule and Consignee List
- c) The charges for Insurance (local transportation and storage), would be extended and borne by the Supplier from ware house to the consignee site for a period including 3 months beyond date of delivery. Other local costs and Incidental costs, as specified in the List of Requirements and Price Schedule;
- d) The charges for Incidental Services, as in the List of Requirements and Price Schedule;
- e) The prices of Site Modification Work ( if any), as mentioned in List of Requirements, Technical Specification and Price Schedule; and
- f) The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Price Schedule.

### 13.5 **Additional information and instruction on Duties and Taxes:**

13.5.1 If the Tenderer desires to ask for GST or any other taxes to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.

### 13.5.2 **Local Duties & Taxes, if any applicable:**

Normally, goods to be supplied to government departments against government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned government department. Keeping this in view, the supplier shall ensure that the stores to be supplied by the supplier against the contract placed by the purchaser are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the purchaser. The purchaser should issue the certificate to the supplier within 21 days from the date of receipt of request from the supplier.

However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the purchaser to enable the purchaser reimburse the supplier and take other necessary action in the matter.

### 13.5.3 **Customs Duty:**

The Purchaser will pay the Customs duty wherever applicable upon actual production of documentary evidence.

### 13.5.4 **Goods and Services Tax (GST):**

- a. If a tenderer asks for Goods and Services Tax to be paid extra, the rate and nature of Goods and Services Tax applicable should be shown separately. The Goods and Services Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction is legally liable to Goods and Services Tax and is payable as per the terms of the contract. If any refund of Tax is received at a later date, the Supplier must return the amount forth-with to the purchaser.
  - b. In case within the delivery period stipulated in the contract, there is an increase in the statutory taxes like GST, Custom Duty, or fresh imposition of taxes which may be levied in respect of the goods and services specified in the contract, reimbursement of these statutory variation shall be allowed to the extent of actual quantum of taxes paid by the supplier. This benefit, however, cannot be availed by the supplier in case the period of delivery is extended due to unexcused delay by the supplier.
  - c. But nevertheless, the Purchaser/Consignee shall be entitled to the benefit of any decrease in price on account of reduction in or remission of customs duty and/or GST or any other duty or tax or levy or on account of any other grounds. In case of downward revision in taxes/duties, the actual quantum of reduction of excise duty must be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.
- 13.6 For transportation of imported goods offered from abroad, relevant instructions as incorporated under GCC Clause 10 shall be followed.
- 13.7 For insurance of goods to be supplied, relevant instructions as provided under GCC Clause 11 shall be followed.
- 13.8 Unless otherwise specifically indicated in this TE document, the terms FCA, FOB, FAS, CIF, CIP, DDP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris
- 13.9 The need for indication of all such price components by the tenderers, as required in this clause (viz., GIT clause 13) is for the purpose of comparison of the tenders by the purchaser and will no way restrict the purchaser's right to award the contract on the selected tenderer on any of the terms offered.

## 14. **Indian Agent**

- 14.1 If a foreign tenderer has engaged an agent in India in connection with its tender, the foreign tenderer, in addition to indicating Indian agent's commission, if any, in a manner described under GIT sub clause 12.2 above, shall also furnish the following information:
- a) The complete name and address of the Indian Agent and its permanent income tax account number as allotted by the Indian Income Tax authority.
  - b) The details of the services to be rendered by the agent for the subject requirement.
  - c) Details of Service outlets in India, nearest to the consignee(s), to render services during Warranty and CMC period.

- d) A copy of agreement between the Agent & their principal detailing the terms & conditions as well as services and after sales services as above to be rendered by the agent and the precise relationship between them and their mutual interest in the business as laid out in section VII (Technical specifications).
- e) Principal's/Manufacturer's original Proforma Invoice with the price bid

## 15. Firm Price

- 15.1 Unless otherwise specified in the SIT, prices quoted by the tenderer shall remain firm and fixed during the currency of the contract and not subject to variation on any account. Bidders are requested to quote BOQ wise unit price (**uniform unit prices must be quoted for same BOQ items across India**) and total price. If a firm quotes NIL Charges/ consideration, the bid shall be treated as unresponsive and will not be considered

## 16. Alternative Tenders

- 16.1 Alternative Tenders are not permitted.
- 16.2 However the Tenderers can quote alternate models meeting the tender specifications of same manufacturer with single EMD.
- 16.3 If an agent submits bid on behalf of the Principal/OEM, the same agent shall not submit a bid on behalf of another Principal/OEM in the same tender for the same item/product. In a tender, either the Indian Agent on behalf of the Principal/OEM or Principal/OEM itself can bid but both cannot bid simultaneously for the same item/product in the same tender.

## 17 Documents Establishing Tenderer's Eligibility and Qualifications

- 17.1 Pursuant to GIT clause 11, the tenderer shall furnish, as part of its tender, relevant details and documents establishing its eligibility to quote and its qualifications to perform the contract if its tender is accepted.
- 17.2 The documentary evidence needed to establish the tenderer's qualifications shall fulfil the following requirements:
- a) in case the tenderer offers to supply goods, which are manufactured by some other firm, the tenderer has been duly authorised by the goods manufacturer to quote for and supply the goods to the purchaser. The tenderer shall submit the manufacturer's authorization letter to this effect as per the standard form provided under Section XIV in this document.
  - b) the tenderer has the required financial, technical and production capability necessary to perform the contract and, further, it meets the qualification criteria incorporated in the Section IX in these documents.
  - c) in case the tenderer is not doing business in India, it is duly represented by an agent stationed in India fully equipped and able to carry out the required contractual functions and duties of the supplier including after sale service, maintenance & repair etc. of the goods in question, stocking of spare parts and fast moving components and other obligations, if any, specified in the conditions of contract and/or technical specifications.
  - d) Deleted.

## 18. Documents establishing good's Conformity to TE document.

- 18.1 The tenderer shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully conform to the goods and services specified by the purchaser in the TE documents. For this purpose the tenderer shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the TE documents to establish technical responsiveness of the goods and services offered in its tender.
- 18.2 In case there is any variation and/or deviation between the goods & services prescribed by the purchaser and that offered by the tenderer, the tenderer shall list out the same in a chart form without ambiguity and provide the same along with its tender.



- 18.3 If a tenderer furnishes wrong and/or misleading data, statement(s) etc. about technical acceptability of the goods and services offered by it, its tender will be liable to be ignored and rejected in addition to other remedies available to the purchaser in this regard.
- 19. Earnest Money Deposit (EMD)**
- 19.1 Pursuant to GIT clauses 8.1 and 11.1 A (i) the tenderer shall furnish along with its tender, earnest money for amount as shown in the List of Requirements. The earnest money is required to protect the purchaser against the risk of the tenderer's unwarranted conduct as amplified under sub-clause 19.7 below.
- 19.2 The tenderers who are currently registered and, also, will continue to remain registered during the tender validity period as Micro and Small Enterprises (MSEs) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) or with National Small Industries Corporation, New Delhi shall be eligible for exemption from EMD. In case the tenderer falls in this category, it should furnish copy of its valid registration details (with MSME or NSIC, as the case may be).
- A) The MSE's Bidder to note and ensure that nature of services and goods/items manufactured mentioned in MSE's certificate matches with the nature of the services and goods /items to be supplied as per Tender.**
- B) Traders/resellers/distributors/authorized agents will not be considered for availing benefits under PP Policy 2012 for MSEs as per MSE guidelines issued by MoMSME.**
- 19.3 The earnest money shall be denominated in Indian Rupees or equivalent currencies as per GIT clause 12.2. The earnest money shall be furnished in one of the following forms:
- i) Account Payee Demand Draft
  - ii) Fixed Deposit Receipt
  - iii) Banker's cheque and
  - iv) Bank Guarantee
- 19.4 The demand draft or banker's cheque or Fixed Deposit Receipt shall be drawn on any scheduled commercial bank in India or country of the tenderer, in favour of the "**HLL Infra Tech Services Limited**" payable at New Delhi. In case of bank guarantee, the same is to be provided from any scheduled commercial bank in India or country of the tenderer as per the format specified under Section XIII in these documents.
- 19.5 The earnest money shall be valid for a period of forty-five (45) days beyond the validity period of the tender. As validity period of Tender as per Clause 20 of GIT is 120 days, the EMD shall be valid for 165 days from Techno – Commercial Tender opening date.
- 19.6 Unsuccessful tenderers' earnest money will be returned to them without any interest, after expiry of the tender validity period, but not later than thirty days after conclusion of the resultant contract. Successful tenderer's earnest money will be returned without any interest, after receipt of performance security from that tenderer.
- 19.7 Earnest Money is required to protect the purchaser against the risk of the Tenderer's conduct, which would warrant the forfeiture of the EMD. Earnest money of a tenderer will be forfeited, if the tenderer withdraws or amends its tender or impairs or derogates from the tender in any respect within the period of validity of its tender or if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged without prejudice to other rights of the purchaser. The successful tenderer's earnest money will be forfeited without prejudice to other rights of Purchaser if it fails to furnish the required performance security within the specified period.
- 19.8 In the case of Bank Guarantee furnished from banks outside India (i.e. foreign Banks), it should be authenticated and countersigned by any nationalised bank in India by way of back-to-back counter guarantee and the same should be submitted along with the bid.
- 20. Tender Validity**
- 20.1 If not mentioned otherwise in the SIT, the tenders shall remain valid for acceptance for a period of 120 days (One hundred and twenty days) after the date of tender opening prescribed in the TE document. Any tender valid for a shorter period shall be treated as unresponsive and rejected.

- 20.2 In exceptional cases, the tenderers may be requested by the purchaser to extend the validity of their tenders up to a specified period. Such request(s) and responses thereto shall be conveyed by surface mail or by fax/ telex/cable followed by surface mail. The tenderers, who agree to extend the tender validity, are to extend the same without any change or modification of their original tender and they are also to extend the validity period of the EMD accordingly. A tenderer, who may not agree to extend its tender validity after the expiry of the original validity period the EMD furnished by them shall not be forfeited.
- 20.3 In case the day up to which the tenders are to remain valid falls on/ subsequently declared a holiday or closed day for the purchaser, the tender validity shall automatically be extended up to the next working day.

## **21. Digital Signing of Tender**

- 21.1 The tenderers shall submit their tenders as per the instructions contained in GIT Clause 11. Tenders shall be uploaded with all relevant tender documents in the prescribed format. The relevant tender documents should be uploaded by an authorised person having Class 3 digital signature certificate.

## **D. SUBMISSION OF TENDERS**

### **22. Submission of e-Tenders**

- 22.1 The tender shall be submitted online only.

#### **(i) Pre-qualification and Technical compliance along with the Techno-Commercial Bid in excel format:**

- a) Scanned copies of tender processing fee and EMD
- b) Manufacturer's authorization in case bid is submitted by an Indian agent (A declaration must be attached here in case directly quoted by a manufacturer or a document establishing the relation of the Indian office with the manufacturer in case quoted by Indian office of the manufacturer).
- c) Tender Form as per Section X.
- d) Compliance of all terms and conditions of TED like- warranty, CMC, delivery period, delivery terms, payment terms, Liquidated Damages Clause, Arbitration clause, etc
- e) Declaration regarding Fall Clause and Deregistration, debarment from any Govt Dept/ Agencies
- f) Copy of PAN.
- g) Certificate of Incorporation/ or a Declaration in case the firm is being a proprietary firm.
- h) Abridged Annual report of last 03 years (Balance sheet and Profit & Loss Account) completed till December 2017, in pdf format.
- i) Name, address and details of account with respect to bidder and/or beneficiary of L/C.
- j) Quality Control Requirements as per Section VIII
- k) Performance statement along with required PO copies and its corresponding end user's satisfactory performance certificate as per section IX.
- l) Technical Bid along with clause-by-clause technical compliance statement for the quoted goods vis-à-vis the Technical specifications along with product catalogue and data sheet in the tender enquiry.
- m) The bidder should submit blank proforma invoice from the foreign manufacturer along with his technical bid, duly mentioning the specifications and code number of the parts quoted.
- n) The original proforma invoices from the foreign principal will be applicable in case of 100% subsidiary companies incorporated in India also.
- o) In case the bidder quotes an equipment of a foreign manufacturer and submits the documents as per Clause 22.1 (i) l & m from the subsidiary company of the foreign Original Equipment Manufacturer in India, the bidder must submit the Power of Attorney given to the subsidiary company by the foreign Original Equipment Manufacturer, authorizing it to do business and perform all obligations for and on behalf of the foreign manufacturer company, in India.

- p) A tenderer quoting imported goods located within India shall produce documentary evidence of the goods having been imported and already located within India (i.e. Bills of Entry for the quoted items and a self-declaration confirming that the quoted items were imported for the purpose of storage in bidder warehouse and for further sale), along with their techno-commercial bid.
- q) *The Integrity pact (At Appendix-B) shall be a part and parcel of this document and has to be signed by bidder(s) at the pre-tendering stage itself, as a pre-bid obligation and should be submitted along with the Techno-Commercial Bids. All bidders are bound to comply with the integrity pact clauses. Bids submitted without signing the integrity pact will be ab initio rejected without assigning any reason.*

**(ii) PRICE BID (ONLY ONLINE):**

- a) The tenderers must ensure that they submit the Price Bid in prescribed format uploaded along with the tender enquiry. It is the responsibility of the bidder to ensure that the contents of the format are not tampered.
- b) The tenderers must ensure that they submit the on-line tenders not later than the closing time and date specified for submission of tenders.
- c) Along with price bid recent purchase order copies for the same model and technical configuration issued by institute of National importance and/or reputed central/state government hospitals should be uploaded in pdf form for reasonability of the offered price.
- d) The bidder should submit the copy of original proforma invoice from the foreign manufacturer along with the price bid.
- e) The supplier shall justify the present quotes based on previous purchase orders for similar project executed either in India or Globally. If they quote any new model or upgraded version of earlier model, they may mention the same in their tender.
- 22.2 The tenderers must ensure that they submit the on-line tenders within the scheduled closing date & time. They shall also ensure to submit the original Tender Processing Fee and EMD within its scheduled date & time.

**23. Late Tender:**

- 23.1 There is NO PROVISION of uploading late tender beyond stipulated date & time in the e-tendering system. However, if the necessary Tender Processing Fee and EMD in original are not submitted within the scheduled time, the tender shall be declared as late tender and online tender shall not be opened and shall be ignored.

**24. Alteration and Withdrawal of Tender**

- 24.1 The tenderer is permitted to change, edit or withdraw its bid on or before the end date & time of bid opening.

**E. TENDER OPENING**

**25. Opening of Tenders**

- 25.1 The purchaser will open the e-tenders at the specified date and time and at the specified place as indicated in the NIT.  
In case the specified date of tender opening falls on / is subsequently declared a holiday or closed day for the purchaser, the tenders will be opened at the appointed time and place on the next working day.
- 25.2 Authorized representatives of the tenderers, who have submitted tenders on time, may attend the tender opening provided they bring with them letters of authority from the corresponding tenderers.

The tender opening official(s) will prepare a list of the representatives attending the tender opening. The list will contain the representatives' names & signatures and corresponding tenderers' names and addresses.

- 25.3 This being a Two - Tender system, the **Techno - Commercial Tenders** are to be opened in the first instance, at the prescribed time and date as indicated in NIT. These Tenders shall be scrutinized and evaluated by the competent committee/ authority with reference to parameters prescribed in the TE document. During the Techno - Commercial Tender opening, the tender opening official(s) will read the salient features of the tenders like brief description of the goods offered, delivery period, Earnest Money Deposit and any other special features of the tenders, as deemed fit by the tender opening official(s). Thereafter, in the second stage, the Price Tenders of only the Techno - Commercially acceptable offers (as decided in the first stage) shall be opened for further scrutiny and evaluation on a date notified after the evaluation of the Techno-Commercial tender.

## **F. SCRUTINY AND EVALUATION OF TENDERS**

### **26. Basic Principle**

- 26.1 Tenders will be evaluated on the basis of the terms & conditions already incorporated in the TE document, based on which tenders have been received and the terms, conditions etc. mentioned by the tenderers in their tenders. No new condition will be brought in while scrutinizing and evaluating the tenders.

### **27. Scrutiny of Tenders**

- 27.1 The Purchaser will examine the Tenders to determine whether they are complete, whether any computational errors have been made, whether required sureties have been furnished and, whether the documents uploaded are in legible form.
- 27.2 The Purchaser's determination of a Tender's responsiveness is to be based on the contents of the tender itself without recourse to extrinsic evidence
- 27.3 Deleted
- 27.4 The tenders will be scrutinized to determine whether they are complete and meet the essential and important requirements, conditions etc. as prescribed in the TE document. The tenders, which do not meet the basic requirements, are liable to be treated as non-responsive and will be summarily ignored.
- 27.5 The following are some of the important aspects, for which a tender shall be declared non-responsive during the evaluation and will be ignored;
- (i) Tender validity is shorter than the required period.
  - (ii) Required EMD or its exemption documents have not been provided.
  - (iii) Tenderer has not agreed to give the required performance security of required amount in an acceptable form in terms of GCC clause 5, read with modification, if any, in Section - V – "Special Conditions of Contract", for due performance of the contract.
  - (iv) Poor/ unsatisfactory past performance.
  - (v) Tenderers who stand deregistered/banned/blacklisted by any Govt. Authorities.
  - (vi) Tenderer is not eligible as per GIT Clauses 5.1 & 17.1.
  - (vii) Tenderer has not quoted for the entire quantity as specified in the List of Requirements/ BOQ for the quoted schedule.
  - (viii) Tenderer has not agreed to other essential condition(s) specially incorporated in the tender enquiry, like delivery terms, delivery schedule, terms of payment, liquidated damages clause, warranty clause, dispute resolution mechanism applicable law.
  - (ix) A tenderer quoting imported goods located within India shall produce documentary evidence of the goods having been imported and already located within India (i.e. Bills of Entry for the quoted items and a self-declaration confirming that the quoted items were imported for the purpose of storage in bidder warehouse and for further sale), along with their techno-commercial bid.

**28. Minor Informality/Irregularity/Non-Conformity**

If during the preliminary examination, the purchaser find any minor informality and/or irregularity and/or non-conformity in a tender, the purchaser may waive the same provided it does not constitute any material deviation and financial impact and, also, does not prejudice or affect the ranking order of the tenders. Wherever necessary, the purchaser will convey its observation on such 'minor' issues to the tenderer by registered/speed post etc. asking the tenderer to respond by a specified date. If the tenderer does not reply by the specified date or gives evasive reply without clarifying the point at issue in clear terms, that tender will be liable to be ignored.

**29 Discrepancies in Prices**

- 29.1 If, in the price structure quoted by a tenderer, there is discrepancy between the unit price and the total price (which is obtained by multiplying the unit price by the quantity), the unit price shall prevail and the total price corrected accordingly, unless the purchaser feels that the tenderer has made a mistake in placing the decimal point in the unit price, in which case the total price as quoted shall prevail over the unit price and the unit price corrected accordingly.
- 29.2 If there is an error in a total price, which has been worked out through addition and/or subtraction of subtotals, the subtotals shall prevail and the total corrected; and
- 29.3 If there is a discrepancy between the amount expressed in words and figures, the amount in words shall prevail, subject to sub clause 29.1 and 29.2 above.
- 29.4 If, as per the judgement of the purchaser, there is any such arithmetical discrepancy in a tender, the same will be suitably conveyed to the tenderer by registered / speed post. If the tenderer does not agree to the observation of the purchaser, the tender is liable to be ignored.

**30. Discrepancy between original and copies of Tender**

Not applicable being e-Tender.

**31. Qualification Criteria**

- 31.1 Bids which do not meet the required Qualification Criteria prescribed in Section IX, will be treated as non - responsive and will not be considered further.
- 31.2 The Purchaser reserves the right to relax the Norms on Prior Experience for Start-ups and Micro & Small Enterprises in Public Procurement.

The Start-ups are defined in Annexure-A of the "Action Plan for Start-ups in India". The same is available on the website of Department of Industrial policy and Promotion (DIPP), Ministry of Commerce & Industry.

**The Notification is available in the below link:**

[http://www.finmin.nic.in/the\\_ministry/dept\\_expenditure/ppcell/RelaxNorms\\_StartupMedEnterprise25072016.pdf](http://www.finmin.nic.in/the_ministry/dept_expenditure/ppcell/RelaxNorms_StartupMedEnterprise25072016.pdf)

**The FAQs are available in the below link:**

[http://dipp.nic.in/English/Investor/startupindia/FAQs\\_StartupIndia\\_30March2016.pdf](http://dipp.nic.in/English/Investor/startupindia/FAQs_StartupIndia_30March2016.pdf)

**Note:- Definition of Startup (only for the purpose of Government schemes)**

(Ref: Ministry of Finance Office Memorandum No. F.20/2/2014-PPD(Pt.) dated 25<sup>th</sup> July 2016.)

**32. Conversion of tender currencies to Indian Rupees**

32.1 In case the TE document permits the tenderers to quote their prices in different currencies, all such quoted prices of the responsive tenderers will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the exchange rates established by the Reserve Bank of India for similar transactions, as on the date of 'Price Tender' opening.

**33. Schedule-wise Evaluation**

33.1 In case the List of Requirements contains more than one schedule, the responsive tenders will be evaluated and compared separately for each schedule. The tender for a schedule will not be considered if the complete requirements prescribed in that schedule are not included in the tender.

**34. Comparison of Tenders**

34.1 Unless mentioned otherwise in Section – III – Special Instructions to Tenderers and Section – VI – List of Requirements, the comparison of the responsive tenders shall be carried out on Delivery Duty Paid (DDP) consignee site basis. The quoted Site Modification Work prices and Comprehensive Annual Maintenance charges (CMC) prices will also be added for comparison/ ranking purpose for evaluation. **“Net Present value (NPV) of the actual CMC price quoted for the required CMC period after the warranty period shall be considered for bid comparison and the NPV will be calculated after discounting the quoted CMC price by a discounting factor of 10% per annum.”**

**35. Additional Factors and Parameters for Evaluation and Ranking of Responsive Tenders**

35.1 Further to GIT Clause 34 above, the purchaser's evaluation of a tender will include and take into account the following:

- i) In the case of goods manufactured in India or goods of foreign origin already located in India, GST or any other taxes which will be contractually payable (to the tenderer), on the goods if a contract is awarded on the tenderer; and
- ii) in the case of goods of foreign origin offered from abroad, customs duty and other similar import duties/taxes, which will be contractually payable (to the tenderer) on the goods if the contract is awarded on the tenderer.

35.2 The purchaser's evaluation of tender will also take into account the additional factors, if any, incorporated in SIT in the manner and to the extent indicated therein.

35.3 The Purchaser reserves the right to give the price preference to small-scale sectors etc. and purchase preference to central public sector undertakings as per the instruction in vogue while evaluating, comparing and ranking the responsive tenders.

- i. In exercise of powers conferred in Section 11 of the Micro, Small and Medium Enterprises Development (MSMED) Act 2006, the Government has notified a new Public Procurement Policy for Micro & Small Enterprises effective from 1<sup>st</sup> April 2012. The policy mandates that 20% of procurement of annual requirement of goods and services by all Central Ministries / Public Sector Undertakings will be from the micro and small enterprises. The Government has also earmarked a sub-target of 4% procurement of goods & services from MSEs owned by SC/ST entrepreneurs out of above said 20% quantity.
- ii. In accordance with the above said notification, the participating Micro and Small Enterprises (MSEs) in a tender, quoting price within the band of L 1+15% would also be allowed to supply a portion of the requirement by bringing down their price to the L 1 price, in a situation where L 1 price is from someone other than an MSE. Such MSEs would be allowed to supply up to 20% of the total tendered value. In case there are more than one such eligible MSE, the 20% supply will be shared equally. Out of 20% of the quantity earmarked for supply from MSEs, 4% quantity is earmarked for procurement from MSEs owned by SC/ST entrepreneurs. However, in the event of failure of such MSEs to participate in the tender process or meet the tender requirements and the L 1 price, the 4% quantity earmarked for

MSEs owned by SC/ST entrepreneurs will be met from other participating MSEs.

- iii. The MSEs fulfilling the prescribed eligibility criteria and participating in the tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir Board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being an MSE, failing which their tender will be liable to be ignored.

**Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2012.”**

35.4 **Preference to Make in India:** As per the order issued by Department of Industrial Policy and Promotion (DIPP) vide No. P-45021/2/2017-BE-II dated 15.06.2017; the purchaser reserves the right to give preference to the local supplier. A copy of this order is enclosed at **Appendix-A** which will form a part of this TED for evaluation and ranking of bids. A local supplier (definition of ‘local supplier’ is given in clause 2 of the aforesaid order of DIPP) has to submit the following along with their tender(s) failing which their bid will be evaluated without considering such preference mentioned in the DIPP order dated 15.06.2017:

- a. The local supplier at the time of tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made.
- b. In cases of procurement for a value in excess of Rs. 10 crores. the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.

The DIPP has notified a Public Procurement (Preference to Make in India) Order, 2017 vide Order no P-45021/2/2017-B.E-II dated 15th June 2017 (Annexure -1). The procurement policy for Micro & Small Enterprises 2012 has been notified under MSMED Act, 2006 (Annexure 2). The orders mandate that purchase preference shall be given to local suppliers and MSEs in all procurement undertaken by procuring entities. General principles as per above orders and criteria fixed by MoHFW shall be followed for various scenarios for award of contract. Accordingly, the criteria of award of contract will be as under:

- a) In procurement of goods where there is sufficient local capacity and local competition and where the estimated value of procurement is Rs.50 lakhs or less, only local suppliers shall be eligible.
- b) If the estimated value of procurement of goods is more than Rs.50 lakhs and which are divisible in nature, the following procedure would apply:
  - I. In case L1 firm is a local supplier:
    - i) The L1 bidder will be awarded full quantity or 80% quantity in case MSEs quotes are within margin of price preference and also accepts L-1 prices.
    - ii) MSME bidders falling under the margin of purchase preference would be awarded upto 20% of the tendered quantity subject to matching the L-1 rate.
  - II. In case L1 firm is not a local supplier:
    - i) 50% of the tender quantity shall be awarded to L1 bidder. Thereafter, the lowest bidder among the local suppliers, will be awarded remaining 50% quantity subject to the local supplier’s quoted prices falling within margin of price preference and

- match the L1 price. In case such lowest eligible local supplier fail to match the L-1 price or accept less than the offered quantity, the next higher local supplier within the margin of purchase preference shall be invited to match the L-1 price for remaining quantity and so on, and contract shall be awarded accordingly.
- ii) The MSME bidders falling under Purchase Preference would be awarded 20% of the tendered quantity subject to matching the L-1 price.
- (c) If the estimated value of procurement of goods is more than Rs.50 lakhs and which are not divisible, the following procedure would apply:
- i). Among all qualified bids, the lowest bid will be termed as L-1. If L-1 is from a local supplier, the contract will be awarded to L-1.
  - ii). If L-1 is not from local supplier, the lowest bidder among the local suppliers, will be invited to match the L-1 price subject to local supplier's quoted prices falling within the margin of purchase preference, and the contract shall be awarded to such local supplier subject to matching the L-1 price.
  - iii). In case such lowest eligible local supplier fails to match the L-1 price, the local supplier with the next higher bid within the margin of purchase preference shall be invited to match the L-1 price and so on and contract shall be awarded accordingly. In case none of the local suppliers within the margin of purchase preference matches the L-1 price then the contract may be awarded to L-1 bidder.
- 35.5 Minimum Local Content: A supplier shall be considered as local supplier provided the minimum local content of the offered item is 50%.
- 35.6 Margin of Purchase Preference: The margin of purchase preference shall be 20%.
- 35.6 Manufacture under license/technology collaboration agreements with phased indigenization are exempted from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement/transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content
- 35.7 Verification of local content
- a. The local supplier at the time of tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made.
  - b. In cases of procurement for a value in excess of Rs. 10 crores, the local supplier shall require to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.
  - c. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating the procuring entity.
  - d. A constituted committee with internal and external experts will examine for independent verification of self-declarations and auditor's/accountant's certificates on random basis and in the case of complaints.
  - e. A fees of Rs.10000/- in the form of demand draft favouring CFO (HLL Infratech Services Limited), payable at New Delhi, is required to be deposited with complaints for verification of local content.
  - f. False declarations will be breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.
  - g. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the duration of the debarment. The debarment for such other



procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities.

### **36. Tenderer's capability to perform the contract**

- 36.1 The purchaser, through the above process of tender scrutiny and tender evaluation will determine to its satisfaction whether the tenderer, whose tender has been determined as the lowest evaluated responsive tender is eligible, qualified and capable in all respects to perform the contract satisfactorily. If, there is more than one schedule in the List of Requirements, then, such determination will be made separately for each schedule.
- 36.2 The above-mentioned determination will, interalia, take into account the tenderer's financial, technical and production capabilities for satisfying all the requirements of the purchaser as incorporated in the TE document. Such determination will be based upon scrutiny and examination of all relevant data and details submitted by the tenderer in its tender as well as such other allied information as deemed appropriate by the purchaser.

### **37. Contacting the Purchaser**

- 37.1 From the time of submission of tender to the time of awarding the contract, if a tenderer needs to contact the purchaser for any reason relating to this tender enquiry and / or its tender, it should do so only in writing.
- 37.2 In case a tenderer attempts to influence the purchaser in the purchaser's decision on scrutiny, comparison & evaluation of tenders and awarding the contract, the tender of the tenderer shall be liable for rejection in addition to appropriate administrative actions being taken against that tenderer, as deemed fit by the purchaser.

## **G. AWARD OF CONTRACT**

### **38. Purchaser's Right to accept any tender and to reject any or all tenders**

- 38.1 The purchaser reserves the right to accept in part or in full any tender or reject any or more tender(s) without assigning any reason or to cancel the tendering process and reject all tenders at any time prior to award of contract, without incurring any liability, whatsoever to the affected tenderer or tenderers.

### **39. Award Criteria**

- 39.1 Subject to GIT clause 38 above, the contract will be awarded to the lowest evaluated responsive tenderer decided by the purchaser in terms of GIT Clause 36.

### **40. Variation of Quantities at the Time of Award/ Currency of Contract**

- 40.1 At the time of awarding the contract, the purchaser reserves the right to increase or decrease by up to twenty five (25) per cent, the quantity of goods and services mentioned in the schedule(s) in the "List of Requirements" (rounded off to next whole number) without any change in the unit price and other terms & conditions quoted by the tenderer.
- 40.2 If the quantity has not been increased to the maximum of 25% of the tendered quantity at the time of awarding the contract, the purchaser reserves the right to increase the quantity further by up to the balance available twenty five (25) per cent of the tendered quantity of goods and services (rounded off to next whole number) without any change in the unit price and other terms & conditions mentioned in the contract during the currency of the contract.

### **41. Notification of Award**

- 41.1 Before expiry of the tender validity period, the purchaser will notify the successful tenderer(s) in writing, by registered / speed post or by email (to be confirmed by registered / speed post) that its tender for goods & services, which have been selected by the purchaser, has been accepted, also briefly indicating therein the essential details like description, specification and

quantity of the goods & services and corresponding prices accepted. The successful tenderer must furnish to the purchaser the required performance security within thirty days from the date of dispatch of this notification, failing which the EMD will be forfeited and the award will be cancelled. Relevant details about the performance security have been provided under GCC Clause 5 under Section IV.

41.2 The Notification of Award shall constitute the conclusion of the Contract.

#### **42. Issue of Contract**

42.1 Promptly after notification of award, the Purchaser/Consignee will mail the contract form (as per Section XVI) duly completed and signed, in duplicate, to the successful tenderer by registered / speed post.

42.2 Within twenty one days from the date of the contract, the successful tenderer shall return the original copy of the contract, duly signed and dated, to the Purchaser/Consignee by registered/speed post. The successful tenderer should also submit Proforma Invoice from the foreign principal (if applicable as per contractual price) within 21 days from the date of NOA.

42.3 The Purchaser/ Consignee reserves the right to issue the Notifications of Award consignee wise.

#### **43. Non-receipt of Performance Security, Proforma Invoice and Contract by the Purchaser/Consignee**

43.1 Failure of the successful tenderer in providing performance security, Proforma Invoice and / or returning contract copy duly signed in terms of GIT clauses 41 and 42 above shall make the tenderer liable for forfeiture of its EMD and, also, for further actions by the Purchaser/Consignee against it as per the clause 24 of GCC – Termination of default.

#### **44. Return of EMD**

44.1 The earnest money of the successful tenderer and the unsuccessful tenderers will be returned to them without any interest, whatsoever, in terms of GIT Clause 19.6.

#### **45. Publication of Tender Result**

45.1 The name and address of the successful tenderer(s) receiving the contract(s) will be mentioned in the notice board/bulletin/web site of the purchaser.

#### **46. Corrupt or Fraudulent Practices**

46.1 It is required by all concerned namely the Consignee/Tenderers/Suppliers etc to observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy, the Purchaser: -

(a) defines, for the purposes of this provision, the terms set forth below as follows:

(i) “corrupt practice” means the offering, giving, receiving or soliciting of anything of value to influence the action of a public official in the procurement process or in contract execution; and

(ii) “fraudulent practice” means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of the Purchaser, and includes collusive practice among Tenderers (prior to or after Tender submission) designed to establish Tender prices at artificial non-competitive levels and to deprive the Purchaser of the benefits of free and open competition;

(b) will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;

(c) will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a contract by the purchaser if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing the contract.

**SECTION - III****SPECIAL INSTRUCTIONS TO TENDERERS (SIT)**

Sl. No.	GIT Clause No.	Topic	SIT Provision	Page No.
A	1 to 7	Preamble	No Change	
B	8 to 10	TE documents	No Change	
C	11 to 21	Preparation of Tenders	No Change	
D	22 to 24	Submission of Tenders	Extra information	18
E	25	Tender Opening	No Change	
F	26 to 37	Scrutiny and Evaluation of Tenders	No Change	
G	38 to 45	Award of Contract	Extra Information	25

The following Special Instructions to Tenderers will apply for this purchase. These special instructions will modify/substitute/supplement the corresponding General Instructions to Tenderers (GIT) incorporated in Section II. The corresponding GIT clause numbers have also been indicated in the text below:

In case of any conflict between the provision in the GIT and that in the SIT, the provision contained in the SIT shall prevail.

**SUBMISSION OF e-TENDERS**

- (i) All the necessary documents as prescribed in the NIT shall be prepared and scanned in different files (in PDF format as prescribed) and uploaded for on-line submission of Proposal.
- (ii) Except Tender Processing Fee and EMD, all document(s)/ information(s) including the Financial Proposal (i.e. FORMAT FOR SUBMISSION OF FINANCIAL PROPOSAL) should be uploaded **online only** in the prescribed format given in the website. No other mode of submission shall be acceptable.
  - i) The prospective bidders may **scan the documents in low resolution (75 to 100 DPI)** instead of 200 DPI. The documents may be scanned for further lower resolution (if possible). This would reduce the size of the Cover and would be uploaded faster.
  - ii) The Individual file size of uploading is restricted up to 5 MB. Bidders may upload multiple files (Not exceeding 5 MB individually) & give relevant file name indicating the contents.
  - iii) The file name of price bid should match the file of the price bid format uploaded by the purchaser in the portal. This can be downloaded from the **Notes & Attachment** under **Details** of item when the event is in **Display Mode**.

**AWARD OF CONTRACT**

- (i) **The quantities in this tender (including additional quantities against the clause “Variation of Quantities at the Time of Award/ Currency of Contract”) can be used by both HLL Infra Tech Services as well as its parent company HLL Lifecare Limited.**

**SECTION - IV**  
**GENERAL CONDITIONS OF CONTRACT (GCC)**

**TABLE OF CLAUSES**

Sl. No.	Topic	Page No.
1	Application	29
2	Use of contract documents and information	29
3	Patent Rights	29
4	Country of Origin	29
5	Performance Security	29
6	Technical Specifications and Standards	30
7	Packing and Marking	30
8	Inspection, Testing and Quality Control	31
9	Terms of Delivery	32
10	Transportation of Goods	32
11	Insurance	32
12	Spare parts	33
13	Incidental services	33
14	Distribution of Dispatch Documents for Clearance/Receipt of Goods	33
15	Warranty	34
16	Assignment	35
17	Sub Contracts	35
18	Modification of contract	36
19	Prices	36
20	Taxes and Duties	36
21	Terms and mode of Payment	36
22	Delivery	39
23	Liquidated Damages	40
24	Termination for default	40
25	Termination for insolvency	40
26	Force Majeure	40
27	Termination for convenience	41
28	Governing language	41
29	Notices	41
30	Resolution of disputes	42
31	Applicable Law	42
32	Withholding and Lien	42
33	General/Miscellaneous Clauses	42

**1. Application**

1.1 The General Conditions of Contract incorporated in this section shall be applicable for this purchase to the extent the same are not superseded by the Special Conditions of Contract prescribed under Section V, List of requirements under Section VI and Technical Specification under Section VII of this document.

**2. Use of contract documents and information**

2.1 The supplier shall not, without the purchaser's prior written consent, disclose the contract or any provision thereof including any specification, drawing, sample or any information furnished by or on behalf of the purchaser in connection therewith, to any person other than the person(s) employed by the supplier in the performance of the contract emanating from this TE document. Further, any such disclosure to any such employed person shall be made in confidence and only so far as necessary for the purposes of such performance for this contract.

2.2 Further, the supplier shall not, without the purchaser's prior written consent, make use of any document or information mentioned in GCC sub-clause 2.1 above except for the sole purpose of performing this contract.

2.3 Except the contract issued to the supplier, each and every other document mentioned in GCC sub-clause 2.1 above shall remain the property of the purchaser and, if advised by the purchaser, all copies of all such documents shall be returned to the purchaser on completion of the supplier's performance and obligations under this contract.

**3. Patent Rights**

3.1 The supplier shall, at all times, indemnify and keep indemnified the purchaser, free of cost, against all claims which may arise in respect of goods & services to be provided by the supplier under the contract for infringement of any intellectual property rights or any other right protected by patent, registration of designs or trademarks. In the event of any such claim in respect of alleged breach of patent, registered designs, trademarks etc. being made against the purchaser, the purchaser shall notify the supplier of the same and the supplier shall, at his own expenses take care of the same for settlement without any liability to the purchaser.

**4. Country of Origin**

4.1 All goods and services to be supplied and provided for the contract shall have the origin in India or in the countries with which the Government of India has trade relations.

4.2 The word "origin" incorporated in this clause means the place from where the goods are mined, cultivated, grown, manufactured, produced or processed or from where the services are arranged.

4.3 The country of origin may be specified in the Price Schedule

**5. Performance Security**

5.1 Within twenty one (21) days from date of the issue of notification of award by the Purchaser/Consignee, the supplier, shall furnish performance security to the Purchaser/Consignee for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations, which is initially valid for a period of minimum six months plus number of months under warranty from the date of Notification of Award

5.2 The Performance security shall be denominated in Indian Rupees or in the currency of the contract as detailed below:

It shall be in any one of the forms namely Account Payee Demand Draft or Fixed Deposit Receipt drawn from any Scheduled bank in India or Bank Guarantee issued by a Scheduled bank in India, in the prescribed form as provided in section XV of this document in favour of the Purchaser/Consignee. The validity of the Fixed Deposit receipt or Bank Guarantee will be for a period up to sixty (60) days beyond Warranty Period.

- 5.3 In the event of any failure /default of the supplier with or without any quantifiable loss to the government including furnishing of consignee wise Bank Guarantee for CMC security as per Proforma in Section XV, the amount of the performance security is liable to be forfeited. The Administration Department may do the needful to cover any failure/default of the supplier with or without any quantifiable loss to the Government.
- 5.4 In the event of any amendment issued to the contract, the supplier shall, within fifteen (15) days of issue of the amendment, furnish the corresponding amendment to the Performance Security (as necessary), rendering the same valid in all respects in terms of the contract, as amended.
- 5.5 The supplier shall enter into Annual Comprehensive Maintenance Contract as per the 'Contract Form – B' in Section XVI with respective consignees, 3 (three) months prior to the completion of Warranty Period. The CMC will commence from the date of expiry of the Warranty Period.
- 5.6 Subject to GCC sub – clause 5.3 above, the Purchaser/Consignee will release the Performance Security without any interest to the supplier on completion of the supplier's all contractual obligations including the warranty obligations, extension of time (with or without Liquidated Damages) & after receipt of Consignee wise bank guarantee for CMC security in favour of Head of the Hospital/ Institute/ Medical College of the consignee as per the format in Section XV.

## **6. Technical Specifications and Standards**

- 6.1 The Goods & Services to be provided by the supplier under this contract shall conform to the technical specifications and quality control parameters mentioned in 'Technical Specification' and 'Quality Control Requirements' under Sections VII and VIII of this document.

## **7. Packing and Marking**

- 7.1 The packing for the goods to be provided by the supplier should be strong and durable enough to withstand, without limitation, the entire journey during transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. As and if necessary, the size, weights and volumes of the packing cases shall also take into consideration, the remoteness of the final destination of the goods and availability or otherwise of transport and handling facilities at all points during transit up to final destination as per the contract.
- 7.2 The quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall strictly comply with the requirements as provided in Technical Specifications and Quality Control Requirements under Sections VII and VIII and in SCC under Section V. In case the packing requirements are amended due to issue of any amendment to the contract, the same shall also be taken care of by the supplier accordingly.
- 7.3 Packing instructions:

Unless otherwise mentioned in the Technical Specification and Quality Control Requirements under Sections VII and VIII and in SCC under Section V, the supplier shall make separate packages for each consignee (in case there is more than one consignee mentioned in the contract) and mark each package on three sides with the following with indelible paint of proper quality:

- a. contract number and date
- b. brief description of goods including quantity
- c. packing list reference number
- d. country of origin of goods
- e. consignee's name and full address and
- f. supplier's name and address

## **8. Inspection, Testing and Quality Control**

- 8.1 The purchaser and/or its nominated representative(s) will, without any extra cost to the purchaser, inspect and/or test the ordered goods and the related services to confirm their conformity to the contract specifications and other quality control details incorporated in the contract. The purchaser shall inform the supplier in advance, in writing, the purchaser's programme for such inspection and, also the identity of the officials to be deputed for this purpose. "The cost towards the transportation, boarding and lodging will be borne by the purchaser and/or its nominated representative(s) for the first visit. In case the goods are rejected in the first instance and the supplier requests for re-inspection, and if same is accepted by purchaser/consignee/PSA/PA, all subsequent inspections shall be at the cost of the supplier. The expense will be to and fro Economy Airfare, Local Conveyance, Boarding and Lodging of the inspection team for the inspection period."
- 8.2 The Technical Specification and Quality Control Requirements incorporated in the contract shall specify what inspections and tests are to be carried out and, also, where and how they are to be conducted. If such inspections and tests are conducted in the premises of the supplier or its subcontractor(s), all reasonable facilities and assistance, including access to relevant drawings, design details and production data, shall be furnished by the supplier to the purchaser's inspector at no charge to the purchaser.
- 8.3 If during such inspections and tests the contracted goods fail to conform to the required specifications and standards, the purchaser's inspector may reject them and the supplier shall either replace the rejected goods or make all alterations necessary to meet the specifications and standards, as required, free of cost to the purchaser and resubmit the same to the purchaser's inspector for conducting the inspections and tests again.
- 8.4 In case the contract stipulates pre-despatch inspection of the ordered goods at supplier's premises, the supplier shall put up the goods for such inspection to the purchaser's inspector well ahead of the contractual delivery period, so that the purchaser's inspector is able to complete the inspection within the contractual delivery period.
- 8.5 If the supplier tenders the goods to the purchaser's inspector for inspection at the last moment without providing reasonable time to the inspector for completing the inspection within the contractual delivery period, the inspector may carry out the inspection and complete the formality beyond the contractual delivery period at the risk and expense of the supplier. The fact that the goods have been inspected after the contractual delivery period will not have the effect of keeping the contract alive and this will be without any prejudice to the legal rights and remedies available to the purchaser under the terms & conditions of the contract.
- 8.6 The purchaser's/consignee's contractual right to inspect, test and, if necessary, reject the goods after the goods' arrival at the final destination shall have no bearing of the fact that the goods have previously been inspected and cleared by purchaser's inspector during pre-despatch inspection mentioned above.  
"On rejection, the supplier shall remove such stores within 14 days of the date of intimation of such rejection from the consignee's premises. If such goods are not removed by the supplier within the period mentioned above, the purchaser/consignee may remove the rejected stores and either return the same to the supplier at his risk and cost by such mode of transport as purchaser/consignee may decide or dispose of such goods at the suppliers risk to recover any expense incurred in connection with such disposals and also the cost of the rejected stores if already paid for."
- 8.7 Goods accepted by the purchaser/consignee and/or its inspector at initial inspection and in final inspection in terms of the contract shall in no way dilute purchaser's/consignee's right to reject the same later, if found deficient in terms of the warranty clause of the contract, as incorporated under GCC Clause 15.
- 8.8 Principal/ Foreign supplier shall also have the equipment inspected by recognised/ reputed agency like SGS, Lloyd, Bureau Veritas, TUV prior to despatch at the supplier's cost and furnish necessary certificate from the said agency in support of their claim.

**9. Terms of Delivery**

- 9.1 Goods shall be delivered by the supplier in accordance with the terms of delivery and as per the delivery period specified in the schedule of requirement. Please note that the time shall be the essence of the contract.

**10. Transportation of Goods**

- 10.1 Instructions for transportation of imported goods offered from abroad:

The supplier shall not arrange part-shipments and/or transshipment without the express/prior written consent of the purchaser. The supplier is required under the contract to deliver the goods under CIP (Named port of destination) terms; the shipment shall be made by Indian flag vessel or by vessels belonging to the conference lines in which India is a member country through India's forwarding agents/coordinators. In case the forwarding agent/coordinators are unable to provide timely adequate space in Indian flag vessel or by vessels belonging to the conference lines, the supplier shall arrange shipment through any available vessel to adhere to the delivery schedule given in the contract.

In case of airlifting of imported goods offered from abroad, the same will be done only through the National Carrier i.e. Air India wherever applicable. In case the National Carrier is not available, any other airlines available for early delivery may be arranged.

- 10.2 Instructions for transportation of domestic goods including goods already imported by the supplier under its own arrangement:

In case no instruction is provided in this regard in the SCC, the supplier will arrange transportation of the ordered goods as per its own procedure.

**11. Insurance:**

- 11.1 Unless otherwise instructed in the SCC, the supplier shall make arrangements for insuring the goods in favour of HLL Infratech services Limited or as directed by HITES against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the following manner:

- i) in case of supply of domestic goods on Consignee site basis, the supplier shall be responsible till the entire stores contracted for arrival in good condition at destination. The transit risk in this respect shall be covered by the Supplier by getting the stores duly insured for an amount equal to 110% of the value of the goods from ware house to ware house (consignee site) on all risk basis. The insurance cover shall be obtained by the Supplier and should be valid till 3 months after the receipt of goods by the Consignee.
- ii) in case of supply of the imported goods on CIP Named port of Destination Basis, the additional extended Insurance (local transportation and storage) would be borne by the Supplier or its Indian Subsidiary/Indian agent from the port of entry to the consignee site for a period including 3 months beyond date of delivery for an amount equal to 110% of the overall expenditure to be incurred by the purchaser from ware house to ware house (consignee site) on all risk basis.

If the equipment is not commissioned and handed over to the consignee within 3 months, the insurance will have to be extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the consignee. In case the delay in the installation and commissioning is due to handing over of the site to the supplier by the consignee, such extensions of the insurance will still be done by the supplier, but the insurance extension charges at actuals will be reimbursed. In case of non-availability of insurance as per requirement and any incident /damage / loss occur during the transit / storage of consignment under any circumstance as mentioned above, the cost and risk shall be borne by the supplier only.



**12. Spare parts**

12.1 If specified in the List of Requirements and in the resultant contract, the supplier shall supply/provide any or all of the following materials, information etc. pertaining to spare parts manufactured and/or supplied by the supplier:

- a) The spare parts as selected by the Purchaser/Consignee to be purchased from the supplier, subject to the condition that such purchase of the spare parts shall not relieve the supplier of any contractual obligation including warranty obligations; and
- b) In case the production of the spare parts is discontinued:
  - i) Sufficient advance notice to the Purchaser/Consignee before such discontinuation to provide adequate time to the purchaser to purchase the required spare parts etc., and
  - ii) The supplier shall be responsible for undertaking the supply of any such spare part for the proper up keeping of equipment for a period of 10 years including the warranty and CMC periods.

12.2 Supplier shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the goods so that the same are used during warranty and CMC period.

**13. Incidental services**

13.1 Subject to the stipulation, if any, in the SCC (Section – V), List of Requirements (Section – VI) and the Technical Specification (Section – VII), the supplier shall be required to perform the following services.

- a. Installation & commissioning, Supervision and Demonstration of the goods
- b. Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
- c. Training of Consignee's Doctors, Staff, operators etc. for operating and maintaining the goods
- d. Supplying required number of operation & maintenance manual for the goods

**14. Distribution of dispatch documents for clearance/receipt of goods**

The supplier shall send all the relevant despatch documents well in time to the Purchaser/Consignee to enable the Purchaser/Consignee clear or receive (as the case may be) the goods in terms of the contract.

Unless otherwise specified in the SCC, the usual documents involved and the drill to be followed in general for this purpose are as follows.

A) For Domestic Goods, including goods already imported by the supplier under its own arrangement

Within 24 hours of despatch, the supplier shall notify the purchaser, consignee, and others concerned if mentioned in the contract, the complete details of despatch and also supply the following documents to them by registered post / speed post / courier (or as instructed in the contract):

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Two copies of packing list identifying contents of each package;
- (iii) Certificate of origin for goods of foreign origin;
- (iv) Insurance Certificate as per GCC Clause 11.
- (v) Manufacturers/Supplier's warranty certificate & In-house inspection certificate.

## B) For goods imported from abroad

Within 24 hours of despatch, the supplier shall notify the purchaser, consignee, and others concerned if mentioned in the contract, the complete details of despatch and also supply the following documents to them by registered post / speed post (or as instructed in the contract). Any delay or demurrage occurred during the customs clearance on account of the non-availability of technical support/ clarifications /documents from the supplier shall be borne by the supplier:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Original and four copies of the negotiable clean, on-board Bill of Lading/Airway bill, marked freight prepaid and four copies of non-negotiable Bill of Lading/Airway bill;
- (iii) Four Copies of packing list identifying contents of each package;
- (iv) Manufacturer's/Supplier's warranty certificate;
- (v) Inspection Certificate for the despatched equipment issued by recognized/ reputed agency like SGS, Lloyd, BUREAU VERITAS, TUV prior to despatch
- (vi) Manufacturer's own factory inspection report;
- (vii) Certificate of origin
- (viii) Port of Loading;
- (ix) Port of Discharge and
- (x) Expected date of arrival.

**Note: 1 In case of sea shipment minimum 14 days demurrage free period to be allotted and instructed to the shipping lines by the supplier/beneficiary.**

**2. Necessary instruction to be given by the beneficiary/ supplier to the Shipping line / airline/ agent / Console to file the IGM in the name of M/s. HLL Infratech Services Limited only.**

**3. In case of air shipments soft copy of Airwaybill, Invoice and Packing list with catalogue of shipment has to be submitted to HITES prior to landing of shipment.**

## 15. Warranty:

- The supplier warrants comprehensively that the goods supplied under the contract is new, unused and incorporate all recent improvements in design and materials unless prescribed otherwise by the purchaser in the contract. The supplier further warrants that the goods supplied under the contract shall have no defect arising from design, materials (except when the design adopted and/or the material used are as per the Purchaser's/Consignee's specifications) or workmanship or from any act or omission of the supplier, that may develop under normal use of the supplied goods under the conditions prevailing in India.
- The warranty shall remain valid for 60 months from the date of installation & commissioning with a regular updates of newer technology as and when evolved followed by a CMC for a period of 5 (Five) Years for all the equipment after the goods or any portion thereof as the case may be, have been delivered to the final destination and installed and commissioned at the final destination and accepted by the purchaser/ consignee in terms of the contract, unless specified otherwise in the SCC.
- No conditional warranty will be acceptable.
- Warranty as well as Comprehensive Maintenance contract will be inclusive of all accessories and Site Modification work and it will also cover the following wherever applicable:-
  - Any kind of motor.
  - Plastic & Glass Parts against any manufacturing defects.
  - All kind of sensors.

- All kind of coils, probes and transducers.
  - Printers and imagers including laser and thermal printers with all parts.
  - UPS including the replacement of batteries.
  - Air-conditioners
  - Replacement and repair will be under taken for the defective goods.
  - All kinds of painting, civil, HVAC and electrical work
  - Proper marking has to be made for all spares for identification like printing of installation and repair dates.
- 15.3 In case of any claim arising out of this warranty, the Purchaser/Consignee shall promptly notify the same in writing to the supplier. The period of the warranty will be as per G.C.C clause number 15.2 above irrespective of any other period mentioned elsewhere in the bidding documents.
- 15.4 Upon receipt of such notice, the supplier shall, within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis respond to take action to repair or replace the defective goods or parts thereof, free of cost, at the ultimate destination. The supplier shall take over the replaced parts/goods after providing their replacements and no claim, whatsoever shall lie on the purchaser for such replaced parts/goods thereafter. The penalty clause for non-rectification will be applicable as per tender conditions
- 15.5 In the event of any rectification of a defect or replacement of any defective goods during the warranty period, the warranty for the rectified/replaced goods shall be extended till the completion of the original warranty period of the main equipment.
- 15.6 If the supplier, having been notified, fails to respond to take action to repair or replace the defect(s) within 8 hours on a 24(hrs) X 7 (days) X 365 (days) basis, the purchaser may proceed to take such remedial action(s) as deemed fit by the purchaser, at the risk and expense of the supplier and without prejudice to other contractual rights and remedies, which the purchaser may have against the supplier.
- 15.7 During Warranty period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the installation for preventive maintenance of the goods
- 15.8 The Purchaser/Consignee reserve the rights to enter into Annual Comprehensive Maintenance Contract between Consignee and the Supplier for the period as mentioned in Section VII, Technical Specifications after the completion of warranty period.
- 15.9 The supplier along with its Indian Agent and the CMC provider shall ensure continued supply of the spare parts for the machines and equipment supplied by them to the purchaser for 10 years from the date of installation and handing over.
- 15.10 The Supplier along with its Indian Agent and the CMC Provider shall always accord most favoured client status to the Purchaser vis-à-vis its other Clients/Purchasers of its equipment/ machines/ goods etc. and shall always give the most competitive price for its machines/ equipment supplied to the Purchaser/Consignee.

## **16. Assignment**

- 16.1 The Supplier shall not assign, either in whole or in part, its contractual duties, responsibilities and obligations to perform the contract, except with the Purchaser's prior written permission.

## **17. Sub Contracts**

- 17.1 The Supplier shall notify the Purchaser in writing of all sub contracts awarded under the contract if not already specified in its tender. Such notification, in its original tender or later, shall not relieve the Supplier from any of its liability or obligation under the terms and conditions of the contract.
- 17.2 Sub contract shall be only for bought out items and sub-assemblies.
- 17.3 Sub contracts shall also comply with the provisions of GCC Clause 4 ("Country of Origin").

**18. Modification of Contract**

- 18.1 If necessary, the purchaser may, by a written order given to the supplier at any time during the currency of the contract, amend the contract by making alterations and modifications within the general scope of contract in any one or more of the following:
- i. Specifications, drawings, designs etc. where goods to be supplied under the contract are to be specially manufactured for the purchaser,
  - ii. Mode of packing,
  - iii. Incidental services to be provided by the supplier
  - iv. Mode of despatch,
  - v. Place of delivery, and
  - vi. Any other area(s) of the contract, as felt necessary by the purchaser depending on the merits of the case.
- 18.2 In the event of any such modification/alteration causing increase or decrease in the cost of goods and services to be supplied and provided, or in the time required by the supplier to perform any obligation under the contract, an equitable adjustment shall be made in the contract price and/or contract delivery schedule, as the case may be, and the contract amended accordingly. If the supplier doesn't agree to the adjustment made by the Purchaser/Consignee, the supplier shall convey its views to the Purchaser/Consignee within twenty-one days from the date of the supplier's receipt of the Purchaser's/Consignee's amendment / modification of the contract.

**19. Prices**

- 19.1 Prices to be charged by the supplier for supply of goods and provision of services in terms of the contract shall not vary from the corresponding prices quoted by the supplier in its tender and incorporated in the contract except for any price adjustment authorised in the SCC.

**20. Taxes and Duties**

- 20.1 Supplier shall be entirely responsible for all taxes, duties, fees, levies etc. incurred until delivery of the contracted goods to the purchaser.
- 20.2 Further instruction, if any, shall be as provided in the SCC.

**21. Terms and mode of payment****21.1 Payment Terms**

Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

**TERMS AND MODE OF PAYMENT****A) Payment for Domestic Goods Or Foreign Origin Located Within India.**

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

**a) On delivery:**

Seventy Five percent (75%) payment of the contract price shall be paid on receipt of goods in good condition and upon the submission of the following documents subject to recovery of LD, if any:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount
- (ii) Two copies of packing list identifying contents of each package
- (iii) Inspection certificate issued by the nominated Inspection agency, if any

- (iv) Insurance Certificate as per GCC Clause 11
- (v) Certificate of origin for imported goods
- (vi) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee

**b) On Acceptance:**

Balance Twenty Five percent (25%) payment would be made against 'Final Acceptance Certificate' as per Section XVIII of goods to be issued by the consignees subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise. FAC needs to be issued by the designated consignee after installation, commissioning, testing and one to two weeks of successful trail run of the equipment.

**B) Payment For Imported Goods:**

Payment for foreign currency portion shall be made in the currency as specified in the contract in the following manner:

**a) On Shipment:**

Seventy Five percent (75%) of the net CIP price (CIP price less Indian Agency commission) of the goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country and upon submission of documents specified hereunder:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Original and four copies of the negotiable clean, on-board Bill of Lading/ Airway bill, marked freight pre paid and four copies of non-negotiable Bill of Lading/Airway bill;
- (iii) Four Copies of packing list identifying contents of each package;
- (iv) Insurance Certificate as per GCC Clause 11 and documents also to be submitted for payment of LC confirming that dispatch documents has already been sent to all concerned as per the contract within 24 hours;
- (v) Manufacturer's/Supplier's warranty certificate;
- (vi) Manufacturer's own factory inspection report and
- (vii) Certificate of origin by the chamber of commerce of the concerned country;
- (viii) Inspection Certificate for the dispatched equipment issued by recognized/ reputed agency like SGS, Lloyd, BEAURU VARITUS and TUV prior to despatch.
- (ix) Consignee Receipt Certificate as per Section XVII in original issued by the authorized representative of the consignee

**b) On Acceptance:**

Balance payment of Twenty Five percent (25%) of net CIP price of goods would be made against 'Final Acceptance Certificate' as per Section XVIII to be issued by the consignees through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the Foreign Principal in a bank in his country, subject to recoveries, if any. FAC need to be issued by the designated consignee after installation, commissioning, testing and one to two weeks of successful trail run of the equipment.

- c) Payment of Incidental Costs till consignee site & Incidental Services** (including Installation & Commissioning, Supervision, Demonstration and Training) will be paid in Indian Rupees to the Indian Agent on proof of final installation, commission and acceptance of equipment by the consignee.

**d) Payment of Indian Agency Commission:**

Indian Agency commission will be paid to the manufacturer's agent in the local currency for an amount in Indian rupees indicated in the relevant Price Schedule (as per prevailing rate of exchange ruling on the date of Contract) and shall not be subject to further escalation / exchange variation. This is payable against submission of a certificate from the principal supplier that they have realised full and final settlement against their supply.

**C) Payment of Site Modification Work, if any:**

Site Modification Work payment will be made to the bidder/ manufacturer's agent of its Indian Office in Indian rupees as indicated in the relevant Price Schedule (as per prevailing rate of exchange ruling on the date of Contract) and shall not be subject to further escalation / exchange variation. This will be paid on proof of final installation, commission and acceptance of equipment by the consignee

**D) Payment for Annual Comprehensive Maintenance Contract Charges:**

The consignee will enter into CMC with the supplier at the rates as stipulated in the contract. The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by the consignee on receipt of bank guarantee for an amount equivalent to 2.5 % of the cost of the equipment as per contract in the prescribed format given in Section XV valid till 2 months after expiry of entire CMC period.

- 21.2 The supplier shall not claim any interest on payments under the contract.
- 21.3 Where there is a statutory requirement for tax deduction at source, such deduction towards income tax and other tax as applicable will be made from the bills payable to the Supplier at rates as notified from time to time.
- 21.4 Irrevocable & non – transferable LC shall be opened by the Purchaser. However, if the supplier requests specifically to open confirmed LC, the extra charges would be borne by the supplier. If LC is required to be extended and/or amended for reasons not attributable to the purchaser/consignee, the charges thereof shall be borne by the supplier.
- 21.5 The payment shall be made in the currency / currencies authorised in the contract.
- 21.6 The supplier shall send its claim for payment in writing, when contractually due, along with relevant documents etc., duly signed with date, to respective consignees.
- 21.7 While claiming payment, the supplier is also to certify in the bill that the payment being claimed is strictly in terms of the contract and all the obligations on the part of the supplier for claiming that payment has been fulfilled as required under the contract.
- 21.8 While claiming reimbursement of duties, taxes etc. (like custom duty and/or GST or any other taxes) from the Purchaser/Consignee, as and if permitted under the contract, the supplier shall also certify that, in case it gets any refund out of such taxes and duties from the concerned authorities at a later date, it (the supplier) shall refund to the Purchaser/Consignee forthwith.
- 21.9 In case where the supplier is not in a position to submit its bill for the balance payment for want of receipted copies of Inspection Note from the consignee and the consignee has not complained about the non-receipt, shortage, or defects in the supplies made, balance amount will be paid by the paying authority without consignee's receipt certificate after three months from the date of the preceding part payment for the goods in question, subject to the following conditions:
- (a) The supplier will make good any defect or deficiency that the consignee (s) may report within six months from the date of despatch of goods.
  - (b) Delay in supplies, if any, has been regularized.
  - (c) The contract price where it is subject to variation has been finalized.
  - (d) The supplier furnishes the following undertakings:

“I/We, \_\_\_\_\_ certify that I/We have not received back the Inspection Note duly receipted by the consignee or any communication from the purchaser or the consignee about non-receipt, shortage

or defects in the goods supplied. I/We \_\_\_\_\_ agree to make good any defect or deficiency that the consignee may report within three months from the date of receipt of this balance payment.

## **22. Delivery**

- 22.1 The supplier shall deliver the goods and perform the services under the contract within the time schedule specified by the Purchaser/Consignee in the List of Requirements and as incorporated in the contract. The time for and the date of delivery of the goods stipulated in the schedule shall be deemed to be of the essence of the contract and the delivery must be completed not later than the date (s) as specified in the contract.
- 22.2 Subject to the provision under GCC clause 26, any unexcused delay by the supplier in maintaining its contractual obligations towards delivery of goods and performance of services shall render the supplier liable to any or all of the following sanctions:
- (i) imposition of liquidated damages,
  - (ii) forfeiture of its performance security and
  - (iii) termination of the contract for default.
- 22.3 If at any time during the currency of the contract, the supplier encounters conditions hindering timely delivery of the goods and performance of services, the supplier shall promptly inform the Purchaser/Consignee in writing about the same and its likely duration and make a request to the Purchaser/Consignee for extension of the delivery schedule accordingly. On receiving the supplier's communication, the Purchaser/Consignee shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.
- 22.4 When the period of delivery is extended due to unexcused delay by the supplier, the amendment letter extending the delivery period shall, inter alia contain the following conditions:
- (a) The Purchaser/Consignee shall recover from the supplier, under the provisions of the clause 23 of the General Conditions of Contract, liquidated damages on the goods and services, which the Supplier has failed to deliver within the delivery period stipulated in the contract.
  - (b) That no increase in price on account of any ground, whatsoever, including any stipulation in the contract for increase in price on any other ground and, also including statutory increase in or fresh imposition of customs duty and/or GST or on account of any other tax or duty which may be levied in respect of the goods and services specified in the contract, which takes place after the date of delivery stipulated in the contract shall be admissible on such of the said goods and services as are delivered and performed after the date of the delivery stipulated in the contract.
  - (c) But nevertheless, the Purchaser/Consignee shall be entitled to the benefit of any decrease in price on account of reduction in or remission of customs duty and/or GST or any other duty or tax or levy or on account of any other grounds, which takes place after the expiry of the date of delivery stipulated in the contract.
- 22.5 The supplier shall not dispatch the goods after expiry of the delivery period. The supplier is required to apply to the Purchaser/Consignee for extension of delivery period and obtain the same before despatch. In case the supplier dispatches the goods without obtaining an extension, it would be doing so at its own risk and no claim for payment for such supply and / or any other expense related to such supply shall lie against the purchaser.

- 22.6.1 Passing of Property:
- 22.6.2 The property in the goods shall not pass to the purchaser unless and until the goods have been delivered to the consignee in accordance with the conditions of the contract.
- 22.6.3 Where there is a contract for sale of specific goods and the supplier is bound to do something to the goods for the purpose of putting them into a deliverable state the property does not pass until such thing is done.
- 22.6.4 Unless otherwise agreed, the goods remain at the supplier's risk until the property therein is transferred to the purchaser.

### **23. Liquidated damages**

- 23.1 Subject to GCC clause 26, if the supplier fails to deliver or install /commission any or all of the goods or fails to perform the services within the time frame(s) incorporated in the contract including opening of office in India as per the undertaking given in the qualification criteria, the Purchaser/Consignee shall, without prejudice to other rights and remedies available to the Purchaser/Consignee under the contract, deduct from the contract price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods, installation, commissioning and/or services until actual delivery or performance subject to a maximum of 10% of the contract price. *Since the Liquidated damages are in virtue of non-performance of services, it will attract GST or any other applicable taxes which in turn shall be deducted from the bidder.* Once the maximum is reached Purchaser/Consignee may consider termination of the contract as per GCC 24.

During the above-mentioned delayed period of supply and / or performance, the conditions incorporated under GCC sub-clause 22.4 above shall also apply.

### **24. Termination for default**

- 24.1 The Purchaser/Consignee, without prejudice to any other contractual rights and remedies available to it (the Purchaser/Consignee), may, by written notice of default sent to the supplier, terminate the contract in whole or in part, if the supplier fails to deliver any or all of the goods or fails to perform any other contractual obligation(s) within the time period specified in the contract, or within any extension thereof granted by the Purchaser/Consignee pursuant to GCC sub-clauses 22.3 and 22.4.
- 24.2 In the event of the Purchaser/Consignee terminates the contract in whole or in part, pursuant to GCC sub-clause 24.1 above, the Purchaser/Consignee may procure goods and/or services similar to those cancelled, with such terms and conditions and in such manner as it deems fit and the supplier shall be liable to the Purchaser/Consignee for the extra expenditure, if any, incurred by the Purchaser/Consignee for arranging such procurement.
- 24.3 Unless otherwise instructed by the Purchaser/Consignee, the supplier shall continue to perform the contract to the extent not terminated.

### **25. Termination for insolvency**

- 25.1 If the supplier becomes bankrupt or otherwise insolvent, the purchaser reserves the right to terminate the contract at any time, by serving written notice to the supplier without any compensation, whatsoever, to the supplier, subject to further condition that such termination will not prejudice or affect the rights and remedies which have accrued and / or will accrue thereafter to the Purchaser/Consignee.

### **26. Force Majeure**

- 26.1 Notwithstanding the provisions contained in GCC clauses 22, 23 and 24, the supplier shall not be liable for imposition of any such sanction so long the delay and/or failure of the supplier in fulfilling its obligations under the contract is the result of an event of Force Majeure.
- 26.2 For purposes of this clause, Force Majeure means an event beyond the control of the supplier and not involving the supplier's fault or negligence and which is not foreseeable and not



brought about at the instance of , the party claiming to be affected by such event and which has caused the non – performance or delay in performance. Such events may include, but are not restricted to, wars or revolutions, hostility, acts of public enemy, civil commotion, sabotage, fires, floods, explosions, epidemics, quarantine restrictions, strikes excluding by its employees , lockouts excluding by its management, and freight embargoes.

- 26.3 If a Force Majeure situation arises, the supplier shall promptly notify the Purchaser/Consignee in writing of such conditions and the cause thereof within twenty one days of occurrence of such event. Unless otherwise directed by the Purchaser/Consignee in writing, the supplier shall continue to perform its obligations under the contract as far as reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
- 26.4 If the performance in whole or in part or any obligation under this contract is prevented or delayed by any reason of Force Majeure for a period exceeding sixty days, either party may at its option terminate the contract without any financial repercussion on either side.
- 26.5 In case due to a Force Majeure event the Purchaser/Consignee is unable to fulfil its contractual commitment and responsibility, the Purchaser/Consignee will notify the supplier accordingly and subsequent actions taken on similar lines described in above sub-paragraphs.

## **27. Termination for convenience**

- 27.1 The Purchaser/Consignee reserves the right to terminate the contract, in whole or in part for its (Purchaser's/Consignee 's) convenience, by serving written notice on the supplier at any time during the currency of the contract. The notice shall specify that the termination is for the convenience of the Purchaser/Consignee. The notice shall also indicate interalia, the extent to which the supplier's performance under the contract is terminated, and the date with effect from which such termination will become effective.
- 27.2 The goods and services which are complete and ready in terms of the contract for delivery and performance within thirty days after the supplier's receipt of the notice of termination shall be accepted by the Purchaser/Consignee following the contract terms, conditions and prices. For the remaining goods and services, the Purchaser/Consignee may decide:
- a) To get any portion of the balance completed and delivered at the contract terms, conditions and prices; and / or
  - b) To cancel the remaining portion of the goods and services and compensate the supplier by paying an agreed amount for the cost incurred by the supplier towards the remaining portion of the goods and services.

## **28. Governing language**

- 28.1 The contract shall be written in English language following the provision as contained in GIT clause 4. All correspondence and other documents pertaining to the contract, which the parties exchange, shall also be written accordingly in that language.

## **29. Notices**

- 29.1 Notice, if any, relating to the contract given by one party to the other, shall be sent in writing by speed post/ Regd. Post or by email. The procedure will also provide the sender of the notice, the proof of receipt of the notice by the receiver. The addresses of the parties for exchanging such notices will be the addresses as incorporated in the contract.
- 29.2 The effective date of a notice shall be either the date when delivered to the recipient or the effective date specifically mentioned in the notice, whichever is later.

### 30. Resolution of disputes

- 30.1 If dispute or difference of any kind shall arise between the Purchaser/Consignee and the supplier in connection with or relating to the contract, the parties shall make every effort to resolve the same amicably by mutual consultations.
- 30.2 If the parties fail to resolve their dispute or difference by such mutual consultation within twenty-one days of its occurrence, then, unless otherwise provided in the SCC, either the Purchaser/Consignee or the supplier may give notice to the other party of its intention to commence arbitration, as hereinafter provided the applicable arbitration procedure will be as per the Arbitration and Conciliation Act, 1996 of India or amendments thereof. In the case of a dispute or difference arising between the Purchaser/Consignee and a domestic Supplier relating to any matter arising out of or connected with the contract, such dispute or difference shall be referred to the sole arbitrator appointed by CEO (HITES). The award of the arbitrator shall be final and binding on the parties to the contract subject to the provision that the Arbitrator shall give reasoned award in case the value of claim in reference exceeds Rupees One lakhs (Rs. 1,00,000/-)
- 30.3 Venue of Arbitration: The venue of arbitration shall be the place from where the contract has been issued, i.e., New Delhi/NCR, India.
- 30.4 Jurisdiction of the court will be from the place where the tender enquiry document has been issued, i.e., New Delhi, India

### 31. Applicable Law

The contract shall be governed by and interpreted in accordance with the laws of India for the time being in force.

### 32. Withholding and Lien in respect of sums claimed

Whenever any claim for payment arises under the contract against the supplier the purchaser shall be entitled to withhold and also have a lien to retain such sum from the security deposit or sum of money arising out of under any other contract made by the supplier with the purchaser, pending finalization or adjudication of any such claim. It is an agreed term of the contract that the sum of money so withheld or retained under the lien referred to above, by the purchaser, will be kept withheld or retained till the claim arising about of or under the contract is determined by the Arbitrator or by the competent court as the case may be, and the supplier will have no claim for interest or damages whatsoever on any account in respect of such withholding or retention.

### 33. General/ Miscellaneous Clauses

- 33.1 Nothing contained in this Contract shall be constructed as establishing or creating between the parties, i.e. the Supplier/its Indian Agent/CMC Provider on the one side and the Purchaser on the other side, a relationship of master and servant or principal and agent.
- 33.2 Any failure on the part of any Party to exercise right or power under this Contract shall not operate as waiver thereof.
- 33.3 The Supplier shall notify the Purchaser/Consignee /the Government of India of any material change would impact on performance of its obligations under this Contract.
- 33.4 Each member/constituent of the Supplier/its Indian Agent/CMC Provider, in case of consortium shall be **jointly and severally liable** to and responsible for all obligations towards the Purchaser/Consignee/Government for performance of contract/services including that of its Associates/Sub Contractors under the Contract.
- 33.5 The Supplier/its Indian Agent/CMC Provider shall at all times, indemnify and keep indemnified the Purchaser/Government of India against all claims/damages etc. for any

infringement of any Intellectual Property Rights (IPR) while providing its services under CMC or the Contract.

- 33.6 The Supplier/its Agent/CMC Provider shall, at all times, indemnify and keep indemnified the Purchaser/Consignee/Government of India against any claims in respect of any damages or compensation payable in consequences of any accident or injury sustained or suffered by its employees or agents or by any other third party resulting from or by any action, omission or operation conducted by or on behalf of the supplier/its associate/affiliate etc.
- 33.7 All claims regarding indemnity shall survive the termination or expiry of the contract.
- 33.8 If any provisions of this tender enquiry or a contract formed on the basis of this tender enquiry are invalid or void under any of the existing provisions of Indian law, then such provisions will not affect other provisions of this tender enquiry/ contract.

**SECTION – V**

**SPECIAL CONDITIONS OF CONTRACT (SCC)**

The following Special Conditions of Contract (SCC) will apply for this purchase. The corresponding clauses of General Conditions of Contract (GCC) relating to the SCC stipulations have also been incorporated below.

These Special Conditions will modify/ substitute/ supplement the corresponding (GCC) clauses. Whenever there is any conflict between the provision in the GCC and that in the SCC, the provision contained in the SCC shall prevail.

The warranty and CMC period will be as mentioned in the list of requirement as per section VI of the tender enquiry.

**SECTION - VI****LIST OF REQUIREMENTS****Part I**

<b>SI No</b>	<b>RFx No</b>	<b>Equipment</b>	<b>Department</b>	<b>Quantity</b>	<b>Warranty</b>	<b>CMC</b>
1	3000003442	Grossing Station-Advance Model	Pathology & Lab Medicine	1	5 years	5 years
2	3000003443	Automated Media Pouring Station Including Stacker and Perparator	Microbiology	1	5 years	5 years
3	3000003444	Fully Automatic Chemiluminescence Analyser	Microbiology	1	5 years	5 years
4	3000003445	Advance Robotic Rehabilitation System	Physical Medicine & Rehabilitation (PMR)	1	5 years	5 years
5	3000003446	Automated Platelet Aggregometer	Transfusion Medicine & Blood Bank (TM& BB)	1	5 years	5 years
6	3000003447	Refrigrated Centrifuge	Transfusion Medicine & Blood Bank (TM& BB)	1	5 years	5 years
7	3000003448	532nm Green laser with slit lamp (Photoocoagulative)	Ophthalmology	1	5 years	5 years
8	3000003449	Inmmunoassay Analyser	Biochemistry	1	5 years	5 years
9	3000003450	Echocardiography machine	Anesthesiology	1	5 years	5 years
10	3000003451	Surgical Microscope	Burn & Plastic Surgery	1	5 years	5 years
11	3000003452	Transcranial Magnetic Stimulator (TMS) with paired Pulse	Neurology	1	5 years	5 years
12	3000003453	Automatic Function Test (AFT) with Tilt Table	Physiology	1	5 years	5 years
13	3000003454	128 Channel EEG ERP System	Physiology	1	5 years	5 years
14	3000003455	Dual Energy X-Ray Absorptiometry	Endocrinology	1	5 years	5 years
15	3000003456	Laporoscopic Surgery Set	General Surgery	1	5 years	5 years

**Part II: Required Delivery Schedule:****a) For Indigenous goods or for imported goods if supplied from India:**

75 days from date of Notification of Award to delivery at consignee site or 30 days from the date of site handover, whichever is later. The date of delivery will be the date of delivery at consignee site. Tenderers may quote earliest delivery period.

Installation and commissioning shall be done within 45 days of receipt of the stores/ goods at site.

**b) For Imported goods directly from foreign:**

90 days from the date of opening of L/C to deliver at port of destination or 30 days from handing over the site, whichever is later. The date of delivery will be the date on which the consignment reaches the Port of Destination. (Tenderers may quote the earliest delivery period).

Installation and commissioning shall be done within 45 days of receipt of the stores/ goods at site.

If the delivery gets delayed due to site related issues, the supplier must get the Site hand over date duly vetted by the consignee.

(The supplier has to ensure the site readiness from the Director/MS of respective consignee/Executing agency before dispatching the equipment. Any delay attributable to site readiness of individual institutes shall be communicated to M/s. HLL Infra Tech Services Limited in writing, for extension of delivery period, with proof from respective Institutes).

Layout drawing for approval, valid Performance Security and Proforma Invoice (in case of LC opening) are to be submitted within 21 days from the date of release of NOA.

Site Readiness means that the site is ready in all aspects for successful delivery, installation and commissioning.

**Note:**

- i) Supplier has to submit clear documents for opening of LC to HLL Infra Tech Services Limited within 21 days of placement of order. Any delay will be treated as non-performance and Liquidated Damages shall be levied.
- ii) In case of multiple LC are opened in favour of multiple manufacturers, the delivery period for all the items under the contract shall be counted from the date of opening of the first LC only.
- iii) Indigenous goods or imported goods if supplied from India (offered in INR) which are linked with supply of directly imported goods, are to be supplied within the contractual delivery period as stated in para b) above.
- iv) Since the supplier is not responsible for custom clearing and forwarding the goods to consignee site, the time taken for the same shall not be counted for computation of LD. However, time taken by the supplier to rectify the short comings of any document for custom clearing the goods to be counted in the above delivery period.

For delayed delivery and/ or installation and commissioning liquidated damages will get applied as per GCC clause 23.

**Part III: Scope of Incidental Services:**

Installation & Commissioning, Supervision, Demonstration, Trial run and Training etc. as specified in GCC Clause 13

**Part IV:** Site Modification Work (if any) as per details in Technical Specification.

**Part V:** Warranty period as per details in general technical specification and as specified in Part I above. Warranty period will be 60 months from the date of installation, commissioning and acceptance.

Comprehensive Maintenance Contract (CMC) as per details in Technical Specification as specified in part I above

**Part VI:**

**Required Terms of Delivery and Destination:**

**a) For Indigenous goods or for imported goods if supplied from India:**

At Consignee Site(s)

**b) For Imported goods directly from abroad:**

The foreign tenderers are required to quote their rates on CIP Named Port of Destination Basis giving breakup of the price as per the Proforma prescribed in the Price Schedule. Purchaser will place the order on CIP Named Port of Destination basis.

**Insurance (local transportation and storage) would be extended and borne by the Supplier or its Indian Subsidiary/Agent from ware house to the consignee site for a period including 3 months beyond date of delivery.**

**Destination/Consignee details:**

A list of Consignee is given in Section XIX. The goods mentioned at Part-I in this section are intended to be supplied to the following hospitals/medical institutes. However, order may be placed for any hospital/institute across India.

**Section – VII**  
**Technical Specifications**

**Item sl. no. 01;**  
**Grossing Station**

1. Should be constructed from high-quality stainless steel.
2. The elevating switch should be located on the front of station and allows the unit to elevate from 32” to 44” (81.3 cm to 111.8 cm).
3. A large 19” x 8 ¼” x 14” (48.3 cm x 22.2 cm x 35.6 cm) sink with radiused corners should be built in the station.
4. The sink also houses a mixing faucet with a gooseneck spout. The mixing faucet should be operated by foot controls, and comes with built in vacuum breaker protection.
5. Vacuum breaker protected water supply preventing contaminated water from getting into the potable water or the natural drinking water.
6. The water temperature should controlled by the mixing valve located inside the sink cabinet.
7. Should have 0.65 Horse Power disposal. When activated, the heavy duty motor runs on a continuous basis and has a manual reset overload switch if needed.
8. The dissecting area rinse should be provided with a constant flow of water to remove debris from the work area to the sink. The rinse can be operated by foot controls.
9. A polyethylene dissecting board should be included to assist with specimen grossing. The durable white surface provides an excellent background for cutting specimens.
10. Grossing Stations should have facility to be ducted to an outside ventilation system for removal of hazardous fumes, vapors and odor.
11. The GFIC waterproof electrical receptacles should be available on quoted model and are located on the front of each station for easy accessibility.
12. A flex arm halogen light is mounted on the station to direct focused light over the work area. The flexible arm allows the light to be adjusted in a variety of positions.
13. The recessed halogen lighting provides proper lighting across the entire work area for better illumination.
14. A stainless steel, braided, handheld spray nozzle with thumb control should be located in the sink area near the mixing faucet. Depress the thumb control to start the water flow. A manual shut off valve should be located under the sink in case of an emergency. Easily accessible control panel.
15. Should have fluorescent magnifier lighting
16. Should be supplied with formalin container with spigot for user convenience while grossing procedures.
17. A magnetic instrument holder should be mounted on the front of the station to store and organize common tools within arm’s reach
18. A standard c-fold paper towel holder should be available
19. Should consist of perforated exhaust grid on the equipment
20. Should have facility for end rinse with control valve
21. Should have camera Digital SLR Camera with 18-55 lens, CMOS Sensor, 16 GB card for recording good quality of photographs with computer interphase (18.0 Mega Pixel) with HDMI cable
22. Should have an eyewash assembly
23. Should have voice activated dictation



24. Should have microphone on flex arm ,230VAC
25. Turnkey installation
26. Should be USFDA /CE/ISO/BIS approved system

**Item sl. no. 02**  
**Automated Culture Media Preparator**

Automated, microprocessor controlled culture media preparator complete for preparing, sterilizing and dispensing media with following items and features:

1. Media preparation system:
  - a. Autoclavable stainless steel cuvette of capacity 1-20 liters or more,
  - b. Sterilization cycle variable from 1000C to 125C for 10-30 minutes as per requirement
  - c. Provision for preparing at least 5 different culture media in a day
  - d. Provision for temperature cycles for preparation of chocolate blood agar and Incorporation of oddities like blood, serum, antibiotics and other selective agents
  - e. Media pouring temperature range of 25°C to 80°C
  - f. Temperature accuracy:  $\pm 0.1^{\circ}\text{c}$
  - g. Preparation time for 5 liters of medium not more than 90 minutes
  - h. LED/LCD display for programme settings and monitoring progress of the procedure
  - i. Selectable stirring to 50-200RPM forward and reverse direction
  - j. Automatic cleaning /draining of inner chamber
  - k. interface 2 x RS232, Ethernet, AUX contact, RS232
  - l. CE / US-ASME certified
  - m. Local service engineer back up
2. Culture Media dispenser system:
  - a. Dosing range 1-75 ml
  - b. Should be able to accommodate culture plates of size 55 and 90mm culture plates from any standard manufacturer
  - c. Twelve (12) additional tubing sets and filling nozzles
  - d. Facility for printing of following on each plate:
    - i. Name of department
    - ii. Name of culture media
    - iii. Date
    - iv. Batch No.
  - e. Cooling plate and UV light in the poring unit for petridishes
3. The instrument should include all devices, equipment and software required for good working of the system
4. Compatible voltage stabilizer
5. Rates of printing ink should be quoted which will be frozen for period of warranty and CMC
6. CMC and Warranty should include all parts of the machine including printing system and maintenance of printer also will be the responsibility of media preparatory supplier but not of any third party.
7. All the civil work required for installation of machine shall be the responsibility of supplier

**Item sl. no. 03;**

**Chemiluminescence based fully automated immunoassay for Infectious Markers Testing:**

1. Fully Automated immunodiagnostic system, shall be based on enhanced chemiluminescence technology, walkaway, high throughput system for diagnosis of all kinds of Infectious diseases and Immunological disorder including Auto-immune diseases with latest acceptable Technology and reagents should be acceptable by DCGI.
2. The instrument should provide comprehensive process check that performs, monitors and verifies each step throughout sample and assay processing.
3. It should have continuous loading capacity of 30 samples.
4. It should have barcode reader to read multiple barcode types.
5. It should have a capability to do the assay in continuous, random, batch & stat mode.
6. System should have On board dilution facility and On board reagent refrigerator.

**2. Sample Handling**

- (i) It should have the capacity to accept various types of sample container like primary, secondary tubes and micro sample cups for sampling purposes.
- (ii) It should have access to the samples during operations.
- (iii) It should have the facility for clot detection, bubble detection, check viscosity, sample level and short samples to ensure accuracy preventing erroneous results due to improper samples.
- (iv) It should have an ability to do on board dilution and reflex dilution for high and abnormal samples.
- (v) It should have the disposable tip sampling system to overcome the carryover and cross contamination probability OR Should have Ultrasonic probe clinging without regular tips.

**3. Reagent Management**

- (i) Reagent loading should be possible during running of instrument without bringing it to standby.
- (ii) The on-board reagent stability should be minimum 2 months.
- (iii) It may have continuous random access to loading and unloading reagents.
- (iv) It should have the compact, integrated reagent pack with all components.
- (v) Lowest turnaround time for sample processing and result will be preferred.
- (vi) It should have the capability of inbuilt inventory management system by tracking all the reagents and supplies automatically.

**4. Calibration and Quality Control**

- (i) It should have the calibration stability of at least 25-30 days for each parameter to decrease reagents consumption.
- (ii) It shall have multiple lot calibration capabilities and calibration curve transition facility"
- (iii) It shall have the QC package system to monitor the quality of result obtained.

## 5. Data Management

- (i) It should have the self diagnosis and error recovery system with on-board operator guides for efficient trouble shooting purpose.
  - (ii) It shall be compatible to the laboratory information system for on data storage facility for a min. of 5000 reports.
  - (iii) It shall have online status for worksheet, samples, reagents, tips, quality controls.
6. Waste Management: it shall have the facility to collect both liquid and solid waste.
  7. Company should provide 500 test reagent for HIV, HbsAg, HCV each free of cost.
  8. Equipment should be complete to start for day one.
  9. The system software must have the capability to diagnose the hardware performance using dedicated software with the provision to log all activities for full traceability.
  10. The firm should provide the details of after sales and service and application backup.
  11. Demonstration and onsite training of staff up to their satisfaction by the application experts is an absolute must.
  12. The firm must have an application specialist and service engineer in the respective state.
  13. The firm must quote the cost of complete set of all routine tests including start up and shut down consumption of reagents required by the system.
  14. A certificate to be provided to the effect that shut down period of the machine must not exceed for more than 24 hours and back up option in case of equipment break down.
  15. The equipment and reagents / kits should be provide European CE-IVD and US-FDA certification.
  16. The firm will supply the UPS with one hour backup system along with system free of cost.
  17. The equipment should be able to run on the existing electrical provision.

### **Item sl. no. 04:** **Advance Robotic Rehabilitation System**

#### Specifications for Advance Robotic Rehabilitation System

Advance Robotic Rehabilitation System for Rehabilitation of Lower and Upper Limb of Patients in Trauma & Neurological Conditions

System should comprise of:

#### **A. ROBOTIC GAIT TRAINING SYSTEM WITH FOLLOWING SPECIFICATION**

1. The unit should be equipped with an electrically driven gait orthosis operated in combination with a treadmill and body weight support system, controlled via PC.
2. The unit should allow intensive gait training for patients with walking disabilities.
3. The unit enables training with a physiologic gait pattern and with active patient participation.
4. The unit should be equipped with Augmented Performance Feedback in which specific gait exercises can be performed in a playful way.
5. The unit should allow for training progression and ability matched challenge by reducing the support given to the patient gradually during the rehabilitation process.

6. The unit should be equipped with assessment tools to analyze and document the patient's progress.
7. The unit should be single therapist operated, easy to use and should relieve therapist from the strain of guiding legs; enabling him to focus on the optimum therapy of the patient.
8. The unit should be upgradable to train small children with femur length 21-35 cm and pelvic width 17-28 cm, on same unit with an interchangeable pair of paediatric orthosis. The paediatric unit should also have Augmented Performance Feedback software specially designed for use in paediatric patients. (optional)
9. Application and efficacy of use of the unit should have been proven by minimum of 250 studies from independent research groups in pathologies like stroke, spinal cord injury, traumatic brain injury, multiple sclerosis, Morbus Parkinson and in patients with cerebral palsy.
10. There should be sufficient Evidence available from trials with the unit showing gait quality improvement, sustained long term benefits, increase in aerobic capacity as a result of training with the unit.

### **Robotic gait orthosis**

1. The unit should be equipped with a driven gait orthosis (gait-robot) for intensive gait training
2. The gait robot should have electrical drives in knee and hip joints
3. Ankle joint dorsiflexion should be supported passively by foot lifters.
4. The gait robot should be mechanically adjustable to individual patient anatomy.
5. The centers of rotation of the joints of the gait orthosis should be adjustable to be congruent with the centers of rotation of the patient's joints for allowing optimum gait trajectories.
6. The upper leg length of the adult gait orthosis should be adjustable between 350 and 470 mm
7. Should also have adjustable pelvis width

### **Robotic orthosis Module for pelvis movements**

8. The unit should contain a Pelvis module with an electrical drive for combined lateral translation and transverse rotation to enable lateral weight shift over stance leg.
9. The Pelvis module should contain an integrated force sensor.
10. The unit should provide Driven lateral translation of body weight support, synchronized to the pelvis movement.
11. The unit should contain slidable cuffs to allow for hip abduction and adduction movements associated with the lateral translation and transverse rotation movements of the pelvis.

### **Treadmill**

12. The unit should be equipped with a treadmill to suitable for manual and Robotic gait training .
13. The treadmill should be operated together with the robotic gait orthosis for performing robotic gait training and should also be operable as a stand-alone without orthosis for performing manual treadmill training..

14. The speed of the treadmill should be adjustable from 0.5-3.2 km/h, during robotic gait training and upto 0.5-10 km/h during manual gait training
15. The unit should facility to increase treadmill speed during robotic gait training to 4.0 km/h for research studies.
16. Surface Dimensions of the treadmill should be minimum (L x W) 140 x 65 cm.

### **Body Weight Support System**

17. The unit should be equipped with a patient lift to allow the lifting of a patient from wheel chair.
18. The unit should have facility of static and dynamic unloading,
19. The unit should allow static unloading of minimum 135 kg.
20. Should have the facility of continuous adjustable unloading of the patient's weight between 0 and 85 kg in the dynamic condition.
21. During the dynamic condition, the unit should promote a physiological vertical movement of the body to enable sensory stimulation of the foot soles.
22. Should have electronically operated unloading of the unit using remote control
23. Body weight system should be equipped with inbuilt electronic weighing scale for real time display of approximate static patient weight and the dynamic weight support.
24. system software should store unloading values set during a training session for reference and records

### **Unit design**

25. The unit should have a ramp to allow access of the patient in the wheel chair.
26. The Unit should allow easy lifting and setting up the patient in the device without turning the patient around
27. The unit should be equipped with easy to reach handrails for the patient. Handrails should be adjustable in width and height according to patient's body dimensions.
28. The unit should have easy to handle design allowing the therapist to access the unit from front and side.
29. The unit should have adjustable seats for the therapist and should allow observation of the patients gait by therapist without any visual obstacles.
30. The unit should be compact and should not consume too much space of a room.

### **User Interface and software**

31. The unit should be equipped with user-friendly touch screen to carry put adjustments and for parameter display
32. The software of the unit should allow adjustment of range of motion of the right & left hip joints and knee individually in order to adjust to asymmetries of neurologic patients.
33. The unit software should allow adjustment of the orthosis speed between robotic gait orthosis and treadmill.

- 
34. The software should allow adjustment of gait velocity specific changes in stride length and stride frequency in order to allow for natural timing of the walking movement.
  35. The unit should have facility to store all measured parameters and patient notes during the individual training sessions for review and records
  36. The software should allow the therapist to continuously access patient history and patient personal notes during the entire training

### **Feedback**

37. The unit should be equipped with a patient feedback screen for visualization and real-time feedback of the patients' performance.
38. The system should have facility to measure gait performance via force sensors.
39. The system should continuously record knee and hip joint activity and should display values for the swing and the stance phase of walking, as well as for the right and left leg.
40. The unit should provide visualization of the gait performance in graphs and other sign.
41. The unit should provide a motivating and instructive functional feedback for the subject in virtual environments with an appropriate choice of environments
42. The unit should provide Augmented Performance Feedback exercises with adjustable settings for subjects specifically focused on
 

i) Attention and motivation	ii) Endurance	iii) Activity timing.
iv) Movement pattern	v) Gait symmetry	vii) step length.
43. Result of Augmented Performance Feedback exercises should allow the therapist to create individual therapy plan for each patient for higher muscle activation and effort during walking.
44. The unit should provide enhanced assist-as-needed support to the subject to enable free timing of movements while ensuring a physiologic gait pattern.
45. The unit should have facility to automatically reward patient activity with higher walking speeds.
46. The unit should have Gait Variation facility to introduce small variations in gait pattern even for patient who require maximal Guidance Force.
47. Difficulty levels and exercise duration of the unit should be adjustable to patients cognitive, gait phase dependent and side specific abilities.
48. The unit provides Augmented Performance Feedback allows the therapist choose between different Avatars for the exercises (Man/ Woman/Boy/Girl/Elvis/Robot)

### **Active training**

49. For active training the system should have facility of guidance force control to allow continuous adjustable level of gait assistance for each leg between full and zero guidance to encourage patient to work harder.
50. Should have facility to select the level of guidance individually for left and right leg of the neurologic patient

51. The system should contain pre-programmable training parameters to automatically vary training parameters as per need.
52. The unit should have speed and unloading training programs and should also allow interval or pyramid training.

### **Assessment Tools**

53. The unit should have assessment tools for following assessment of neurological patients
  - i) Stiffness in the patient's hip and knee joints.
  - ii) Isometric force generated by the patient.
  - iii) Patient's hip and knee- passive range of motion.
54. The unit should have facility to visualize assessment results on the patient screen.

### **Research & reporting Tools**

55. The unit should have a research function which allow to collect a wide range of training data, to record data with atleast 10 signals , to allow the therapist to choose and create his own data sets for recording for later export and analysis.
56. The unit should have facility to display data of single session, summarize all training data as therapy progression format and generate individual patient therapy reports
57. It should be possible to export data of the device usage (hours, session, patients) for records and analysis

### **Connectivity and remote support**

58. It should be possible to print reports from the unit into the clinic network.
59. It should be possible to access the unit from remote location for clinical and technical support and software updates.

### **Safety Mechanisms**

60. The unit should be equipped with wireless remote control that must be pressed by the user to confirm that the system is running safely and correctly.
61. The unit should have clearly red marked emergency stop button.
62. The system should detect undesired movements and stop automatically if the range of movement is beyond a certain threshold or if a patient is stumbling.
63. The system should contain a torque monitoring system to automatically stops the machine if the torque is beyond the normal range or if a patient generates too high forces, which could be generated through the patient's spasticity.
64. The body weight support system of the unit should be equipped with an emergency release feature.

### **System should be supplied complete with following Orthopedic material**

65. Patients' harnesses (adult) in four different sizes for different sized patients- 1 no each

66. Cuffs for the upper, middle and lower legs in seven different sizes - 1 set
67. Foot lifter - 1 set.
68. The harnesses should be washable at 60 deg C in the washing machine.
69. Cushioning material for back, gluteus and shins- 1 set
70. Wall mount Orthopedic storage device.

### **Size and dimensions**

71. The unit should be compact to be easily installed in a room of size 5.5m L X 3.5m W X 2.6m H
72. The unit should be extendable in height to accommodate patients with more height
73. The unit should be suitable for training of patients with height of upto 200 cm

### **Certification and compliances**

74. It should be USFDA registered device with verifiable documents
75. Should be electromagnetic compatibility Class A device (CISPR 11).
76. The unit should meet the requirements of EN 60601-1-2., meet the protective class SK 1.
77. The unit should be a class II a device according to the European medical device directive 93/42/EC.

### **System Performance**

78. The system should be in wide use all over the world with minimum 200 installations.
79. There should be sufficient published literature regarding Application and efficacy of use of the unit should have been proven by minimum of 50 studies from independent research groups in pathologies like stroke, spinal cord injury, traumatic brain injury, multiple sclerosis, Parkinson and in patients with cerebral palsy. There should be sufficient Evidence available from trials with the unit showing functional improvement, sustained long term benefits, increase capacity as a result of training with the unit.
80. The vendor should demonstrate the quoted system anywhere in India at own expenses to the hospital doctors. Demonstration is a must

## **B) SELF DRIVEN UPPER EXTREMITY FUNCTIONAL THERAPY SYSTEM WITH FOLLOWING SPECIFICATIONS**

1. Instrumented arm orthosis with integrated weight compensation mechanism with Augmented feedback
2. The Arm Weight Support should be based on an exoskeleton with integrated spring mechanism for Arm Weight Support.
3. The Arm Weight support should be designed to counterbalance (middle length setting / horizontal position)
  - forearm weight: 0.7 kg - 2.4 kg
  - upper arm weight: 0.5 kg - 3.8 kg



- 
4. The size exoskeleton of the exoskeleton should be adjustable to different arm sizes of adult persons; minimum settings are
    - forearm lengths (elbow to handgrip): 290-390 mm
    - upper arm lengths (shoulder to elbow): 220-310 mm
  5. The exoskeleton should contain the (lockable) distal degrees of freedom for pro Supination and wrist flexion / extension
  6. The exoskeleton should be equipped with at least 7 joint angle sensors to detect movements in all relevant degrees of freedom
  7. The exoskeleton should contain a pressure sensitive handle to detect trace amounts of grasping activity of a patient
  8. The exoskeleton should be usable for the left and the right arm
  9. The exoskeleton should contain braces to position the patients' arms
  10. The braces should be eudermic and washable
  11. The exoskeleton should be mounted on an electric lifting column (Adjustable Range: 400 mm)
  12. The exoskeleton should be mounted on a trolley and be movable by one person
  13. The electric lifting column should be operated by a remote control (should not be wireless for safety)
  14. The Augmented Feedback should contain at least 20 different functional exercises or games
  15. The Augmented Feedback should contain functional exercises in a 1D,2D and 3 Dimensional workspace.
  16. The library of Augmented Feedback functional exercises should be filterable according to therapy goals and the joints involved
  17. The Augmented Feedback workspace for 2D and 3D exercises should be automatically personalized according to the physiological volume reachable by the patient in 3D.
  18. The Augmented Feedback should contain reach to grasp type exercises (involving Arm and Hand treatment at the same time)
  19. The Augmented Feedback should contain a user database to manage individual patients' therapy plans
  20. The Augmented Feedback should automatically progress through all exercises in patient's individual therapy plan, without staff intervention
  21. The Augmented Feedback should store all therapy activities and contain export routines to access raw data
  22. The user database should support at least 500 patients data
  23. The system should have facility of automatic backup of user data
  24. The system should contain inbuilt software recovery tools
  25. The Augmented Feedback should contain assessment capabilities to quantify

- i. Patients' coordination ability
  - ii. Patients' range of motion in each degree of freedom.
  - iii. Physiological volume reachable by the patient in 3D.
26. The level of difficulty of therapy exercises in the Augmented Feedback should be adjustable
  27. Within each level of difficulty in the Augmented Feedback, there should be multiple exercise levels to allow for gradual progression in exercise challenge.
  28. The level of visual detail of therapy exercises in the Augmented Feedback should be adjustable to patient's visual or cognitive impairment.
  29. The Therapy system should be adaptable to the individual patients' active range of motion in the impaired arm
  30. The Augmented Feedback should contain reporting capabilities to automatically generate reports in PDF format with patients' exercise scores and assessment results over time and with information about device usage for the hospital managers.
  31. The system should be operated on Normal Indian Power supply 230 V 50/60 Hz.
  32. The system should contain PC and a 24" screen for visualization of the exercises
  33. The system should be capable of software updates via USB stick
  34. A user training by a clinical application specialist should be contained in the scope of delivery
  35. The Supplier shall guarantee the equipment for a period of 5 years
  36. Wear parts should be contained in the scope of delivery.
  37. The therapy system should also be supplied complete with a hand module for training in grasping function for patients with impaired hand functions
  38. The hand module should support a hand opening movement with a non-powered spring mechanism  
The hand module should work in plug-and-play manner during therapy sessions. Hand Module should also have Augmented Feedback containing exercises specifically designed for combined hand and arm control.

### **System Performance**

39. Application and efficacy of use of the unit should have been proven by minimum of 50 studies from independent research groups in pathologies like stroke, spinal cord injury, traumatic brain injury, multiple sclerosis, Parkinson and in patients with cerebral palsy.
40. There should be sufficient Evidence available from trials with the unit showing functional improvement, sustained long term benefits, increase upper extremity capacity as a result of training with the unit.
41. The system should be in wide use all over the world with minimum 500 installations.
42. The vendor should demonstrate the quoted system anywhere in India at own expenses to the hospital doctors. Demonstration is a must

**Certification and compliances**

43. It should be USFDA registered device with verifiable documents
44. The unit should have CE market clearance and should meet the requirements of EN 60601-1 standards

**Installation & Training**

The vendor should undertake to provide free installation and training for technical and clinical staff at site of installation by their engineer including operation and upkeep of the equipment.

**Warranty**

- a) The unit should be covered under warranty for a period of 5 years
- b) Compatible software updates should be provided free of charge during the warranty period.
- c) The unit's manufacturer should provides remote technical support and clinical support for questions via email and phone
- d) The manufacturer should provide a clinical knowledge platform with latest information about products, state of the art clinical research and the possibility to exchange with other professionals in the community.

**Item sl. no. 05:**  
**Platelet Aggregometer**

1. Platelet Aggregometer working on the principle of Platelet Rich plasma (Optical turbidimetric)/ Whole blood Impedance)
2. Should have minimum two channels with optical aggregation - Aggregation in PRP, with sample volume: 250µl or less. Impedance aggregation - Aggregation in whole blood preferably with a sample volume of 500µl.
3. It should be possible to upgrade the two channel model into four channels, in future.
4. Should gave adjustable stirrer speed between 400 to 1200 RPM in 10-RPM steps with "stirrer stopped" position, and error detection to prevent operation, if not within  $\pm 10$  RPM
5. Should offer flexibility to use both regular / disposable probes based on application needs.
6. At least six wells for incubating samples at 37 deg C ( $\pm 1$  deg C) before running the test shall be Available
7. Should have LCD display, one per channel, for display of': (a) Heater block - temperature in 'C (b) Stirring speed in RPM (c)operating mode (optical or Impedance) (d)warning message
8. Computerized data analysis with appropriate system software with CD-Writer for convenient storage and retrieval of data, colour monitor and colour inkjet printer should be included. Software package (windows XP, or the higher version applicable one) should be included.

9. System should include an internal computer interface and comprehensive software for measuring up to two samples for simultaneous aggregation
10. Software on-board should provide for real time display of two channels of aggregation, Computation of amplitude, slope, log time and area under the curve, Storage of reagent data for tracking test value, demography details, for later recall
11. Should be FDA and/or CE approved
12. System should be quoted complete with appropriate stabiliser or UpS with one hour battery backup.
13. The recommended maintenance kit, for trouble free operation and maintenance of the system for a minimum period of five years should be quoted.
14. The consumables including reagents should be available for supply in local currency and preferably within 48 hours after ordering.
15. Start up kit, comprising appropriate liquid handling systems, reagents and consumables for installation and demonstration shall be included. Demonstration and onsite training of staff by the application experts is an absolute must.
16. The after sales service should be available preferably within 48 hours and with trained service personnel based in Bhubaneswar/Kolkata only. The bidder has to produce a certificate assuring that they have an application specialist based in Bhubaneswar/Kolkata who will be attending any test related problem within reasonable time to ensure zero down time.
17. Warranty for 5 years & CMC price for next 5 years, which should include cost of labour spares, entire equipment.
18. The Firm should provide price for all the reagents and consumables required for next 5 years.
19. Evaluation of Tender: The tender will be evaluated on the basis of cost of the equipment along with the price of the reagents panel (for ADP, arachidonic acid, epinephrine, collagen and ristocetin) and other consumables, if necessary for 100 tests.

**Item sl. no. 06;**  
**Refrigerated centrifuge**

**Technical specification of Refrigerated centrifuge**

1. For separation of components from whole blood.
2. Microprocessor controlled
3. Programmable memory: memory with temper proof facility.
4. Stainless steel chamber: easy to clean, corrosion resistant with provision of both drain and condensed water collection container. Stainless steel should be of Medical grade.
5. CFC free refrigerant.
6. Swinging bucket rotor: with metal buckets, 6x2000 ml, with or without wind shield, suitable adapters for 12/16 blood bags of 350ml and 450 ml or preferably 16 blood bags of 350 ml and 450 ml capacity.

7. Removable plastic cups to hold single/ double/ triple/ quadruple blood bags with partition in every bucket.
8. Insert with hook adapter to spin buffy coat or small volume of blood and balancing weights for inserts.
9. Equipped with automatic lid lock.
10. Centrifugal force: 5000-6000g
11. Speed variation: microprocessor controlled rotor speed to within 10 rpm of set value.  
Acceleration and deceleration profiles shall be available.
12. Temperature range -10 deg C to +40 deg C
13. Microprocessor controlled rotor temperature within r.o c of set temperature regardless of centrifuge speed.
14. Programmable time: 0-99 minutes with minimum resolution of 1 minute.
15. Digital display of temperature, speed and time. Minimum no. of 3 digit resolution.
16. Motor imbalance detection: automatic shut down of centrifuge if rotor load is out of balance with appropriate indicator. Should incorporate alarms for imbalance detection lid interlock, over temperature, rotor over speed.
17. Power requirement: Power requirement:220/240 volts, 50 Hz. Single phase or Three phase as per machine requirement. . ,
18. The equipment shall be suitable for operation from 0 deg to 40 deg C at 90% relative humidity. Electronic circuitry shall be tropicalised for this ambient condition.
19. Noise level within 60 decibels.
20. The equipment shall have lockable castors.
21. Protection of data: in event of power interruption or complete failure, data should remain stored.
22. Should have a provision for external connectivity and should have software which should be compatible with hospital information system of respective AIIMS and /or Blood Bank software any interfacing required must be provided by the firm.
23. It shall have a security lock to prevent unintentional switch off and also unauthorized opening of the equipment.
24. At least 5 year warranty period with 5 years CMC after expiry of warranty period. Free of cost annual calibration in the warranty period.
25. Automatic line voltage corrector/ voltage stabilizer:
  - (i) A line voltage corrector of appropriate rating (10 KVA or as per the requirement of equipment) should form part of standard configuration.
  - (ii) Copper wound single phase automatic line voltage corrector conforming to IS: 9815(PLI)/94 with latest amendments or equivalent international standards fitted with a voltmeter and switch to indicate output/ input voltage.
  - (iii) Input voltage: 140-280 V,50 Hz, output voltage:  $220 \pm 10 \%$ .
  - (iv) Input output voltmeter and amperemeter. Protection for high low voltage cut off, overload and

short circuit protection.

- (v) Equipment should be supplied with 2 meter cord at input and fitted with plugs of appropriate rating.
- (vi) Make of the line voltage corrector shall be indicated.

26. Certifications: Product Certification: ISO 9001, ISO 14001 accreditation, (CE certified as class II A medical device by a notified body as per medical device directive 93/42/EEC), MED/CERT to DIN EN ISO 13485-2003 and FDA registrations. Please note that just plain CE certificate will not suffice.

**Item sl. no. 07;**  
**532nm Green laser with slit lamp (Photoocoagulative)**

**Specification for solid state photo coagulator (Green laser) with slit lamp**

1. Laser source should be diode-pumped, frequency doubled solid-state laser having 532 nm operating wavelength.
2. Should have a aiming laser wave length of 635 nm, max power 1mW.
3. Red aiming beam should have adjustable intensity.
4. Portable design, light weight. Not more than 10kg.
5. Exposure time: 0.01 to 4.0 sec.
6. Spot size: 50 to 1000 microns, continuously variable, Para focal.
7. Auto pulse adjustable from 100-6000 ms pulse intervals.
8. Repetition rate up to 10Hz.
9. Should have forced air cooling system.
10. Unit should have foot switch, safety glass and transport case.
11. Attachment: -Slit lamp (with at least 3 step magnification).
12. Slit lamp adapter with micromanipulator.
13. Should have provision to connect two doctors filter.
14. Should have provision for remote interlock.
15. Should have dual laser ports.
16. Should have electric requirement of 220 V.
17. Should have control in footswitch.
18. Should have a output power ranging from 30mV to 2000mV (2W).
19. Should have a exposure time of 10ms-2000ms and continuous wave.
20. Should have power source for LIO on the console.
21. Microscope filter attachment with Endo LASER delivery.(optional).
22. Suitable laser contact lenses (i.e. Goldman 3 mirrors, mainster PRP 165 deg).

**CE Certified**

**Item sl. no. 08**

**Fully Automated Chemiluminescence Immunoassay Analyser**

1. Should be based on the principle of Chemiluminescence Immunoassay analyser.
2. The system should be using chemiluminescence technology with very high sensitivity and linearity
3. It should be fully automated continuous random access analyzer with a throughput of 200 tests or more per hour.
4. As it will be for hospital and clinical laboratory it should be able to detect concentrations below, in and above the range of normal clinical levels.
5. The samples should be serum, plasma, urine, CSF and other body fluids.
6. The test types will include hormones, tumour markers, antibodies, to infectious diseases and autoimmune diseases, allergy, cardiac, nutrition markers and drugs to be monitored.
7. There should be random sampling with minimum sample volume ranging from 10-150 microliter with minimum dead volume.
8. Provision for paediatric sample cup
9. Clot detection, clot management and crash prevention facility should be available.
10. Use of disposable cups, cuvettes and tips.
11. There should be provision for onboard dilution with automatic calculation.
12. There should be provision for on board refrigeration and stability of reagents on board for at least 15 days
13. 65 or more reagents / test parameters onboard at a time with provision of continuous loading of samples
14. Auto inventory on board
15. The pack size of reagents should range from 20 to 500.
16. The reagents should be ready to use
17. Shelf life of reagents should be a minimum of 6 months
18. Provision for both bar code reading and manual entry.
19. Floor model
20. Provision for running emergency tests.
21. Low maintenance, the recurring cost of washing, maintenance and consumables (other than reagents) used per test should be specified.
22. Accessories, reagents calibrator and control: company shall provide a list of accessories, reagents, calibrator and control to be used for running the instrument. Manufacturing Company if have their own system reagents, controls and calibrators then the price list for the same should be enclosed with the price bid.
23. Noise generated should be less than 50 decibels
24. Supplied With UPS with a minimum of 60 minutes backup
25. Operating Software: Windows or compatible.

26. Program should be compatible to our LIS and HIS for easy reporting System should have facility for bidirectional interface for LIS/HIS.
27. Program should have access to report retrieval, statistics and storage for data should be up to 1 year or more.
28. System should have memory of minimum 50,000 patient test results.
29. System should have onboard maintenance logs.
30. System should have high resolution touch screen LCD with mouse.
31. Attached Laser Printer with real time individual sample reporting facility
32. Technology should be latest and not more than 2 years in market.
33. Warranty should be for at least 5 years or more.
34. Closed or open system
35. Price for AIIMS for test parameters should be comparable with the existing rates.
36. The risk assessment of triple and quadruple tests and other risk assessment applications should be in built or if not available should be compatible with those sold by other companies.
37. Training for the staff should be provided on site by company using their reagents.
38. Provision for providing service for troubleshooting, breakdown etc within 12 hours. Service engineer and dealer should be at Bhubaneswar.
39. The equipment to be supplied should have US FDA / European CE certification.

**Item sl. no. 09**  
**Echocardiography machine**

<b>1</b>	<b>Description of Function</b>
1.1	Echocardiography System is required to study the anatomical abnormalities and blood flow in the heart and associated vessels.
<b>2</b>	<b>Operational Requirements</b>
2.1	<b>Latest generation Electronic Phased array Colour Doppler system with Minimum 50000 Electronic independent channels. System should be DICOM ready and capable of being interfaced with HIS/RIS/ PACS.</b>
2.3	Frequency compounding or better technology for better resolution and penetration.
<b>3</b>	<b>Technical Specifications</b>
3.1	<b>Latest generation Electronic Phased array Colour Doppler system with Minimum 50,000 Electronic independent channels.</b>
3.2	256 grey shades for sharp contrast resolutions
<b>4</b>	<b>Transducers ( frequency tolerance :±2MHz)</b>
a.	Adult phased array transducer with frequency from 1-5 MHz:
b	Pediatric phased array transducer with frequency from 3-8 MHz
c	Convex Transducer for abdominal ultrasound with 1-5 MHz
d.	Vascular Linear probe with frequency form 4-12 MHz
e	<b>Separate TEE probe for adult and pediatric echocardiography</b>
<b>5</b>	<b>Harmonic Imaging- System should have following modes in harmonic with separate setting for:</b>
a.	Tissue Harmonic



c.	Harmonic Angio
6	Harmonic imaging capability in Adult Cardiac and linear Probe.
7	Gain control in two dimensions for additional level of flexibility to image quality control.
8	Real time high frequency 2D for higher resolution and low frequency Doppler for higher sensitivity in all probes
9	<b>Frame rate should be 500 FPS or more</b>
10	Steerable PW/CW in all Phased Array probes.
11	High-definition acoustic zoom for enlarging sections of 2D and Colour flow images with more acoustic information for greater clarity and detail while maintaining an optimal frame rate.
12	Colour M-Mode
13	Modes —2D, M-Mode, Steerable PW/CW Doppler, Colour Doppler, and High Definition Colour flow <b>Anatomical M Mode, Tissue Doppler, Automated speckle tracking with strain rate and volumes display</b>
14	Monitor should be 15" or more, high-resolution colour Monitor.
15	Tilt and Swivel monitor should be able to view in all angles and all light conditions.
16	Tissue Colorization (B-Colour) for improved contrast resolution
17	Application software for Adult and Peripheral Vascular and Trans oesophageal applications. (All application package should be built into the system)
18	Cine loop memory- more than 120MB of memory or equivalent cine loop memory in frames/ sec.
a.	High Frame rate review for better clarity of playback images study in slow motion.
b.	Quad loop with memory for pre and post image comparison of any procedure.
c.	Memory- 256 frames or more in quad loop. M Mode & Doppler Scroll Memory-40seconds or more.
d.	Frame grabber facility for post analysis.
19	Various maps for pre and post processing.
20	ECG trigger facility.
21	User defined system and application presets for multi-user department.
22	Dedicated integrated dynamic stress echo package for flexible user defined protocols with stacked sub loops facility and contrast stress protocol usable for stress echocardiography.
23	<b>Tissue movement colorization with 2D quantification/2D strain possibly for IHD/CAD/Heart failure patient</b>
24	Three or more active transducer ports.
25	Colour Map resolution up to 128 levels.
26	Facility for high definition digital acquisition, review and editing of complete patient studies.
27	<b>PC based Peripheral system comprising of dedicated computer of standard make, at least 5TB storage space (Hard disc) with 8 GB RAM or more with a Microprocessor speed of more than 3.00 GHz, interfaced with the echocardiography machine with DVD writer and a high quality Colour Laser printer.CD/DVD produced should be playable on any system.</b>
28	Online UPS of suitable rating with voltage regulation and spike protection for 30 min back up.
<b>29</b>	<b>Standards, Safety and Training</b>
29.1	Should have European CE with 4 digit notifying body number/USFDA Approved certificate to be provided.
30	<b>Added Para: The system should be upgradable to contrast agent imaging</b>
<b>SI No</b>	<b>BOQ</b>
1	Colour Doppler Echocardiography System with 2D facility , as specified
2	Digital Storage and Retrieval device
3	Adult Cardiac probe
4	Linear probe

5	Pediatric Cardiac probe
6	Adult Convex Probe
7	Multi plane TEE Probe for Adult echocardiography
8	Paediatric TEE probe
9	Colour Print Paper- 500 sheets
10	ECG Cable
11	Laser Colour Printer
12	Suitable UPS for a 30 minute backup for whole system.
13	Standalone PC (Windows based) with suitable DICOM viewer
14	CD
15	DVD

**Item sl. no. 10**

**Surgical Microscope**

1. Working distance	<b>200- 500 mm</b> continuously variable through motorized multifocal lens, activated through Hand or Foot switch and through control panel. Manually adjustable override.
2. Magnification range	<b>1.5x-21x</b> with 12.5 X eyepiece.
3. Field of view	Minimum field of View <b>15 mm -150 mm</b> with 10x eyepiece
4. Focusing	Motorized via multifocal lens activated through Hand or Foot switch & Touch screen control panel. Manually adjustable override.
5. Eyepieces	12.5 X eyepieces
6. Light Source	300W <b>Xenon illumination</b> with integrated 300W <b>Xenon back-up</b> and Single button Automatic lamp Change over
7. Illumination Field Diameter :	Should have Built in automatic zoom-synchronized illumination field diameter, with manual override and reset feature. <b>Automatic Iris Control or Auto Iris</b>
8. Automated illumination control	Should have Automated Illumination. Brightness control should be linked to working distance. <b>Focus Light Link or Bright Care.</b>  Illumination also can be controlled by hand switch or-foot switch.
9. Binocular tube	<b>0 -180 deg Tilttable</b> binocular tube (must) Tube with 12.5X eyepiece with PD adjustment knob.
10. Balancing :	<b>One touch auto-balance</b> for the whole microscope.
11. Suspension system & Height:	<b>Counter weight suspension system</b> (Hydraulic suspension system will not consider). True overhead capability not less than <b>7Kg</b> . Height in Stand by position not less than <b>1880 mm</b> .
12. X-Y Movement	XY robotic movement in 3 axes (variable speed)
13. Control Unit	Single window operation. (Means one control panel can operate all the microscope, Camera, Display & recording

	Unit, along with Fluorescence(Optional).)
14. Beam Splitter	Fully integrated Beam Splitter (not Visible from outside)
15. Camera	<b>Integrated 3CCD HD Video Camera</b> (Camera not Visible from outside) So that maximum resolution will display & record.
16. Display	HD Medical grade touchscreen display system attached with the microscope body. (No External Monitor / detachable monitor will be acceptable)
17. Recording	Fully HD 1080p Recording system with <b>integrated HDD.</b>
18. Image Injection	Microscope should have facility of image injection of any external image like MRI, CT Scan & Endoscope Image as a HD mode etc.
19. Face to Face attachment	Built in face to face attachment with 0 -180 deg Tilttable binocular (must) Tube with 12.5X eyepiece with PD adjustment knob
20. Footswitch	14 Function or More functional footswitch control for microscope
21. ICG Systems	The microscope should be up gradable to ICG systems for vascular patency test.
22. After Sale & Service	After Sales & Service will be provided by the <b>Manufacturer company.</b> Should have sufficient stoke of spare in their local office.

**Item sl. no. 11**  
**Transcranial Magnetic Stimulator (TMS) with paired Pulse**

1. It should be a highly effective non-invasive repetitive biphasic magnetic stimulator for bilateral cortical or peripheral and central nerve pathways stimulation.
2. The system should be able to deliver output power as per the following.
  - a. should deliver 100% power at 41-42 Hz
  - b. should deliver 50% output at 96-100 Hz.
3. The system should have
  - a. Integrated color touch screen display.
  - b. Integrated 2 channel Motor Evoked Potential.
  - c. Should be programmable to deliver Single pulse or Repetitive train of sessions.
  - d. Theta Burst Protocol.
4. The pulse train interval should be user selectable with 0.1s increment and pulse width of 400  $\mu$ s with maximum upper frequency of 100 Hz
5. The machine should be provided with EMG Interface module.

6. The system should be supplied with following Accessories:
  - a. 70mm double Air-cooled coil- 1 no.
  - b. 70mm double Air film- 1 no. The coil should have integrated air circulation system which should deliver up to 3000 pulses at 120 % motor threshold at 10 Hz stimulation (Biphasic waveform) over a 30-minute period and remains within allowed temperature limits.
  - c. 70 mm double Placebo coil- 1 no.
  - d. 360-degree articulated coil Stand for positioning coil at any position on patient.
  - e. In built stand for the main unit
7. Store and retrieve setups and protocols
8. Transfer data to external devices via USB or serial port
9. Advanced Power Pulse for 40 % additional stimulation power
10. Versatile In/Out triggers for EEG, EMG and EP equipment
11. Should have facility for charge delay for TMS Synchronized EEG.
12. Should be supplied with patient comfortable chair
13. The equipment should also have provision to attach other coil such as 90mm high power circular coil, 110mm double cone coil with rTMS system and the rates of these coils should be quoted separately.
14. Should have an intelligent thermal protection for temperature management of the coils.
15. Automatic sequence set-up and user-definable protocols
16. Should be Upgradable for interleaved TMS/fMRI study
17. Should have safety certificate from competent authority CE/ ISO/ FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
18. Warranty & CMC as per tender terms.

**Item sl. no. 12**

**Automatic Function Test (AFT) with Tilt Table**

**Technical Specifications:**

The system should be able to record and analysis

1. Non-invasive beat to beat continuous Blood pressure monitoring in Human with Systolic, Diastolic and other hemodynamic parameters like CO, SV, TPR, Baro-reflex sensitivity, Blood pressure variability etc.) By finger arterial pressure measurement.
2. Continuous ECG monitoring and recording with real time HRV Analysis.
3. At least four channel Universal Bio-amplifier for ECG, EMG, EEG recordings & analysis.
4. Dynamometer to study handgrip strength profile with balance board for static posturography studies.

5. Hand Held Non Invasive Pulse Tonometer for Vascular function Testing.
6. Number of channels: 8 or more channels high speed USB Data Acquisition system.
7. High sensitivity and sampling rate of 10 KHz or more per channel (aggregate speed).
8. ADC resolution = 16 bits or more.
9. The Beat to beat Blood pressure monitor should be supplied with HCU and Finger cuffs of 3 different sizes (i.e. small, medium & large) and should have accuracy: 1% of full scale (max. 3 mmHg) , zeroing automatic.
10. Transducers for Finger PPG, Respiration Sensor, Airflow, GSR, Temperature measurement, Hand dynamometer and Wireless heart rate belt transducer and other accessories required.
11. Licensed Software: It should have various automatic analysis modules for ECG, HRV, Blood pressure, Cardiac output, Peak analysis, spike histogram etc.
12. Online & offline analysis with various export options for common formats.
13. The software should be provided with a 5 year of free updates and upgrades.
14. Computer system:- Intel Core i7 , Genuine Windows 8 professional, 20 inch LED Monitor ,4GB RAM , 1 TB Hard Drive, DVD. Facility for internet connectivity, Laser printer and 2 KVA UPS along with trolley and working table.
15. Motorized electric tilt table:-
  - Safe working load and lifting capacity (from minimum height ) of 180 kgs
  - Large wheel design, with central locking & steering facility.
  - Individual banking castors with electric height and tilting operation with hand switch control and tilt angle inclinometer as standard.
  - Adjustable angle dual foot boards – positive and negative (+15 to -30degrees).
  - Lowers to wheelchair height for ease of patient transfer.
16. Proper demonstrations to be carried out before finalizing.
17. Onsite training should be provided by the company experts for 7-10 days.
18. 5 years warranty.
19. CE, ISO, FDA and other safety certificates must be provided.
20. The system should have world-wide installation, acceptance and recognition in published research papers globally. Performance certificates should be provided from the user using the system in India /abroad.

**Item sl. no. 13**  
**128 Channel EEG ERP System**

**Technical Specifications**

**A) EEG Amplifier:**

1. The 128-Channel EEG System must have CE certification under medical device directive.
2. The 128-channel EEG amplifier must meet the below technical specs
  - a) Input Noise:  $< 0.9 \mu\text{V}$  @ 1000 Hz sampling rate
  - b) A/D Resolution: 24 bit
  - c) Dynamic Range:  $\pm 200\text{mV}$
  - d) Bandwidth: DC to 2,000 Hz (for all 128 channels)
  - e) Sampling rate: 1000 Hz for all 128 channels
  - f) Input impedance:  $> 1 \text{ G Ohm}$
  - g) CMRR  $> 90 \text{ dB}$
  - h) Digital Inputs – 16 bit
  - i) DC coupling
3. The 128-Channel EEG Amplifier must be a single base unit amplifier. No ganging/stacking of amplifiers allowed
4. The 128-Channel EEG Amplifier must have an on-board Intel Microprocessor and Embedded Linux OS.
5. The 128-Channel EEG Amplifier must be a fiber-optic/Giga Ethernet peripheral.

**B) Electrodes, Nets/Caps and affiliated accessories:**

6. Total application time of 128-channel electrode net/cap must be less than 15 minutes. Live demonstration may be required to evaluate this capability.
7. Coverage for the inferior /ventral surface of the brain is required for accurate source modeling; this means there must be adequate electrode coverage below the eyebrows, around the ears, cheeks, etc.
8. Abrasion-free, gel-free (saline electrolyte based) EEG recording capability. Very low daily operating and maintenance cost. All the consumables needed for a minimum of 100 recordings must be included in the offer.
9. Open structured net design preferred with dense and even distribution of electrodes for superior comfort as well as to conform different head shapes, sizes and hair types.
10. A total of at least 3 different net/caps sizes must be included in the offer to fit Adult Indian head sizes ranging 51 to 58 cm. All 3 caps/nets must be warranted for at least 5 years.
11. Each net/cap must come with its own set of dedicated electrodes to avoid wear and tear in the long run. So a minimum of 384 EEG electrodes must be included in the offer. And all these electrodes as well as caps must be covered under warranty for a minimum of 5 years,

**C) EEG/ERP Software:**

12. Two licenses of full EEG/ERP acquisition, review and analysis software must be included in the offer. Analysis tools must at least comprise of the following -

---

Filtering, Segmentation, Artifact Detection, Bad channel Replacement, Averaging, Average Referencing, Baseline Correction, Grand Averaging, Statistic Extraction, etc. The software must also allow flexible data import/export. Export to Matlab plugins such as EEGLab, FieldTrip, etc. must be easily achieved. Furthermore, export to EDF+ and simple Binary format must be possible. It must also be possible to call Mat lab/EEG Lab routines from within the review and analysis software environment.

#### **D) Source Estimation Software:**

13. In addition to the complete EEG/ERP Acquisition, Review and Analysis software licenses, a full license for source localization capabilities must be included in the offer. The source localization must be based on distributed modeling algorithms such as LAURA, LORETA, SLORETA, etc.
14. The Source estimation software must provide various age and gender specific high resolution FDM Atlases.
15. Further, the software must be upgradeable to allow individual MRI import capability.

#### **E) Computer, Monitoring and Isolation Transformer for EEG System:**

16. Two 27" iMacs must be provided with the offer for EEG/ERP data acquisition and source analysis. The minimum specs are:

3.2 GHz quad-core Intel Core i5, Turbo Boost up to 3.6GHz  
8GB 1867MHz DDR3 SDRAM – two 4GB  
1TB Serial ATA Drive @ 7200 rpm  
AMD Radeon R9 M380 with 2GB video memory.

17. Two 600-Watt Hospital Grade Isolation Transformers must be included with the offer; one for the EEG system and the other for the ERP Stimulus Presentation Workstation.

#### **F) Stimulus Presentation System (ERP):**

18. A complete stimulus Presentation workstation system based on E-Prime 2.0 Professional software or comparable software must be offered. The package must include everything that is needed to configure the system in a typical 2-room ERP laboratory setup. The stimulus presentation workstation and EEG system must communicate via TCP/IP protocol and millisecond timing accuracy must be guaranteed by software based Network Time Protocol (NTP).
19. Two licenses of E-prime 2.0 Professional software or two licenses of comparable stimulus presentation software should be included in the offer.
20. A high performance Desktop computer must be included with the stimulus presentation system. Please note that this is a separate machine from EEG acquisition computer.
21. Two 21" LCD Monitors must be included; one for the participant and the other for experimenter room. Again, these are separate monitors from the EEG acquisition screen. These monitors must be tested for fixed and minimal timing offset.
22. Appropriate video splitter, switches and cables must be included to allow both the monitors to be connected /controlled through a single desktop machine.
23. Experimental control hardware such as response pad, serial response box, microphone, etc. that is needed for typical visual as well as auditory ERP experiments must be included.

24. A complete A V Device package to allow millisecond timing verification of the auditory and visual ERP experiments must be included in the offer.
25. Necessary training must be provided with the on-site visit.

**G) On-site Installation/Training visit:**

26. A minimum of 3-day on-site install/training visit must be provided with the EEG/ERP system. Local scientific and engineering support must be available.

**H) Warranty and support:**

27. Five years of warranty for the 128-channel EEG amplifier must be included in the main offer.
28. Five years of warranty must be provided on all the electrodes and caps that are included in the main offer.
29. Five years of technical support must be included in the main offer.

**Item sl. no. 14**

**Dual Energy X-Ray Absorptiometry(DEXA) System**

**Scanner Hardware and Acquisition Technology.**

- High Resolution one pass isocentric True Fan Beam Scan Technique .
- Energy switching technique using 140/100kv Peak Energies.
- Internal Reference system for Pixel by Pixel Data automatic continuous calibration. Operator calibration not required.
- Motorized Table and motorized rotating c-arm.
- Minimum 216 or more Multi Element High Resolution Detector Array gadolinium based or better scintillator technology used in modern CT Devices.
- High frequency oil cooled x- ray source.

**Minimum Computer Hardware**

- Computer workstation with minimum Dual core 3 GHz. With Window 7 Professional. Or higher version.
- 250 GB hard drive. 2GB RAM or more.
- High Resolution minimum 19" or more wide screen LCD Monitor with suitable color DeskJet printer.
- 2 Nos. of external hard Drive of minimum 1 TB or more to store patients data.

**Quality Assurance & Phantoms**

- Automated daily procedure, multiple system checks with automatic pass/fail results. Automatic trending graph of QA variables.
- Anthropomorphic Spine Phantom for Daily Quality Control assures long term stability
- **Whole Body Phantom with following minimum configuration of better:**
  1. It should be designed to monitor the reliability of BMD and Body Composition.
  2. The phantom is composed of material to represent the thickness and percentage of bone, lean and fat mass in the human physiological range.



3. It should have layers of high density polyethylene and polyvinylchloride simulate fat and lean tissue respectively.
4. It should also have layer of aluminum to represent bone.

### **Radiation Dose.**

- Skin entrance dose for a typical AP Spine or Femur Scan in 0.07mGy. Skin entrance dose for typical AP or Lateral Instant Vertebral Assessment High Definition image is 0.05Gy.
- Scatter dose less than 10 µGy/h at 1 meter.

### **Reference Data**

- Reference Data n > 8000 FDA approved
- Default NHANES III standardized database
- The Bone and Mineral Density in Childhood Study (BMDCS) Reference Database, Caucasian, sex and age matched.

### **Scan Time**

- 10 sec for Hip & Spine or less
- 5 Minute Whole Body BMD Scan or less with Automated 9 region analysis protocol for 9 region analysis Protocol for BMD Composition Analysis.

### **Scan Window**

- Apex operating system.
- Facility for the supine lateral spine with baseline compensation.
- FRAX (WHO 10 years fracture risk assessment)
- BMD 10 second Spine and Hip acquisition.
- Better than 1.0% in vivo precision.
- L1-L4 Standard Spine Analysis.
- Five Region Proximal Femur Analysis (Total Femur, Femoral Neck, Inter-Trochanteric, Acu view automated Dual Femur.
- Instant Vertebral Assessment High Definition.
- Trabecular bone scoring.
- Visual, Quantitative and Semi – quantitative analysis of vertebral bodies:
- Computer – aided fracture assessment tool.
- Quantitative Morphometry with 6 point vertebral height measurements.
- Forearm analysis with automated analysis of Ultra Distal, 33% Shaft, Mid – Shaft and Total ROI for Ulna, Radius and Global measurements.
- Sub – Region BMD and Body Composition Analysis.
- Advanced Body composition Analysis with Visceral Fat Assessment IVA HD with image Pro High Resolution Imaging Capability.
- Calculation and display of advanced and Ganoids sub – regions.
- Reflection – ability to reflect data for limbs when there are large are differences between right and limb.
- Scoliotic Spine Analysis.
- Pediatric Spine, Hip, and Whole Body Analysis.
- Small animal.

- Abdominal aortic calcifications.
- ISCD Complaint Standardized BMD Reporting.
- Single/Dual Energy Image Display.
- Serial Scan Comparison.
- Software and Hardware upgrade available to protect from obsolescence.

### **Application Training**

- 2-3 days of comprehensive Application training by a certified Radiologic Technologist.

### **Item sl. no. 15** **Laparoscopic Surgery Set**

- 1) 3D Telescope & camera:
  - a. The company should quote the latest model available.
  - b. Should have dual lens technology for left and right image signals.
  - c. The 3D Scope should be 10 mm in diameter with either High resolution flexi-Tip Dual lens Chip on tip (CCD Based) Technology with 0-80 Deg in all four directions.

Or

With a camera head and 1 no's rigid scope each for 10 mm 0 deg, & 30 Deg for 3D Laparoscope.

- d. Should have a facility for synchronizing and processing of both left and right image signals form camera processor to transmit 3D image.
- e. Should provide full HD imaging. The resolution of output should be at least 1920x1080p HD.
- f. Output provided should be compatible with the offered 3D monitor.
- g. Should have provision for switch over between 3D and 2D observation mode from the 3D telescope for easy maneuvering.
- h. The 3D scope should be focus free with depth of field of at least 20-90 mm with a wide field of view of at least 60-90 degree.
- i. Should also have a facility to connect routine HD camera heads 3 CCD/CMOS for surgery for use with non-video telescopes for routine 2D applications when 3D is not desire.
- j. 3CCD/CMOS full HD camera should be supplied along with the system with following Specifications:
  - Pixels: 1920x1080
  - Lens: Integrated optical Zoom, F=15-31 mm
  - Control buttons on camera Head: 3(2 of them Freely Programmable).
  - AGC: Microprocessor controlled.
- k. System should have special offer filter for observation of capillary vessels and fine patterns in the superficial layer of mucosa for early detection of lesions.

1. There should not be any image rotation and distortion associated with the rotation of 3 D Telescope/Camera head on 30°Lens.

2) Monitor:

- a. The 3D monitor should be full HD with 32 inch or above in size and should have 1920\*1200 pixel resolution.
- b. The 3D monitor should have:
  - I. At least two 3D /HD-SDI, DVI-D inputs for 3D image display.
  - II. At least two 3d/HD-SDI output for 3D image transmission.
- c) The 3D monitor should have brightness adjustment feature while switching between 2D/3D images.

3) Light source:

- a. Should have powerful 300 W xenon OR LED light source equivalent to 300 W xenon.
- b. Should have special filter light for observation of capillary vessels and fine patterns in the
- c. Superficial layer of mucosa for early detection of lesions” .
- d. The company should provide at least one spare xenon bulb and free of cost maintenance of light source.
- e. The necessary cables and adaptors should be provided with the light source.

4) Dual channel HD 3D video recording device:

- a. Dual channel 3D HD with memory of 500 GB to record, store and transmit 3D videos and images. The recording system should have facility to overlay the voice of surgeon on the video for teaching purpose.
- b. Should have touch screen with size 7 inch or more.

5) Insufflator: CO2 ELECTRONIC INSUFFLATOR 45L/Min or More

- a. “High flow mode with flow performance 45l/min or more
- b. Optional and acoustic alarm signals in the event of patient overpressure
- c. Should have automatic/manual smoke evacuation facility
- d. Electronic color screen

Technical Specification:

- Gas flow: 0 to 40 l/min or more.
- Pressure: 0-25 mmHg (4000Pa) or more
- System should be US FDA/European CE certified
- All items should be from same manufacturer

Essential accessories:

The system should be supplied with:

- A) 20 nos. of light weight polarized glasses.
- B) Additional 10 nos. of clip on type glasses for spectacle wearing surgeons
- C) Necessary cables to connect all equipment's.
- D) OEM Trolley to house all equipment's.

Others:

- The system should be compatible and should adapt to the existing operation theatre system.
- Physical demonstration of the quoted model is mandatory.
- The complete installation including accessories is the responsibility of the supplying firm. The equipment should be fully functional and compatible with the existing electrical and other equipment's of the main operation theatre of the institute.

## GENERAL TECHNICAL SPECIFICATIONS

### GENERAL POINTS:

1. Warranty:

- a) Five years Comprehensive Warranty as per Conditions of Contract of the TE document for complete equipment (including Batteries for UPS, other vacuumatic parts wherever applicable) Warranty period will be 5 years from the date of installation, commissioning and Site Modification Work from the date of satisfactory installation, commissioning, trial run & handing over of equipment to Hospital/Institution/Medical College.
- b) 95% up time Warranty of complete equipment with extension of Warranty period by double the downtime period on 24 (hrs) X 7 (days) X 365 (days) basis.
- c) **All software updates should be provided free of cost during Warranty period.**

2. After Sales Service:

After sales service centre should be available at the city of Hospital/Institution/Medical College on 24 (hrs) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 8 hrs. The service should be provided directly by Tenderer/Indian Agent. Undertaking by the Principals that the spares for the equipment shall be available for at least 10 years from the date of supply.

3. Training:

On Site training to Doctors/ Technicians/ staff is to be provided by Principal/ Indian Agents (if they have the requisite know-how) for operation and maintenance of the equipment to the satisfaction of the consignee. The same will be in line with the training modalities as specified in general technical specification.

4. Annual Comprehensive Maintenance Contract (CMC) of subject equipment with Site Modification Work:

- a) The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual of the manufacturer, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years on yearly basis for complete equipment (including Batteries for UPS, other vacuumatic parts wherever applicable) and Site Modification Work (if any). The supplier shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, but at least once in six months during the CMC period
- b) The cost of CMC may be quoted along with taxes applicable on the date of Tender Opening. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
- c) Cost of CMC will be added for Ranking/Evaluation purpose. The same will be taken at Net Present Value with a 10% discounting factor each year.
- d) The payment of CMC will be made on six monthly basis after satisfactory completion of said period, duly certified by end user on receipt of bank guarantee for 2.5 % of the cost

of the equipment as per Section XV valid till 2 months after expiry of entire CMC period.

- e) There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- f) During CMC period, the supplier is required to visit at each consignee's site at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
- g) All software updates should be provided free of cost during CMC.
- h) Failure of the above [4. e) to 4. g)] by the supplier, may lead to the forfeiture of the Bank Guarantee for Annual CMC.
- i) The payment of CMC will be made as stipulated in GCC Clause 21.

#### 5. Site Modification Work:

Site Modification Work is indicated in the technical specification of the respective items, wherever required. The Tenderer shall examine the existing site where the equipment is to be installed, in consultation with HOD of Hospital/Institution/Medical College concerned. Site Modification Work details of each Hospital/Institution/Medical College are given at the end of Technical Specification. The Tenderer to quote prices indicating break-up of prices of the Machine and Site Modification Work of each Hospital/Institution/Medical College. The Site Modification Work costs to be quoted in Indian Rupee will be added for Ranking Purpose. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

The Site Modification Work should completely comply with AERB requirement, if any.

**Note 1:** Tenderer's attention is drawn to GIT clause 18 and GIT sub-clause 11.1 A (iii). The tenderer is to provide the required details, information, confirmations, etc. accordingly failing which it's tender is liable to be ignored.

**Note 2:** General: Bidders are requested to make sure that they should attach the list of equipment for carrying out routine and preventive maintenance wherever asked for and should make sure that Electrical Safety Analyzer / Tester for Medical equipment to periodically check the electrical safety aspects as per BIS Safety Standards IS-13540 which is also equivalent to IEC electrical safety standard IEC-60601 is a part of the equipment s. If the Electrical Safety Analyzer/Tester is not available they should provide a commitment to get the equipment checked for electrical safety compliance with Electronic Regional Test Labs / Electronics Test and Development Centres across the country on every preventive maintenance call.

**Note 3:** Adequate training of personnel and non-locked open software and standard interface interoperability conditions for networked equipment in hospital management information system (HMIS)

The successful tenderer will be required to undertake to provide at his cost technical training for personnel involved in the use and handling of the equipment on site at the institute immediately after its installation. The company shall be required to train the institute personnel onsite for a minimum period of 1 month

All software updates should be provided free of cost during warranty period and CMC period

**Section – VIII**  
**Quality Control Requirements**

(Proforma for equipment and quality control employed by the manufacturer(s))

Tender Reference No.

Date of opening

Time

Name and address of the Tenderer:

Note: All the following details shall relate to the **manufacturer(s)** for the goods quoted for.

- 01 Name of the manufacturer
  - a. full postal address
  - b. full address of the premises
  - e. telephone number
  - f. fax number
  
- 02 Plant and machinery details:
  
- 03 Manufacturing process details:
  
- 04 Monthly (single shift) production capacity of goods quoted for
  - a. normal
  - b. maximum
  
- 05 Total annual turn-over (value in Rupees)
  
- 06 Quality control arrangement details
  - a. for incoming materials and bought-out components
  - b. for process control
  - c. for final product evaluation
  
- 07 Test certificate held
  - a . type test
  - b . BIS/ISO certification
  - c . any other
  
- 08 Details of staff
  - a. technical
  - b skilled
  - c unskilled

**Signature and seal of the Tenderer**

**Section – IX**  
**Qualification Criteria**

1. The tenderer must be a manufacturer. In case the manufacturer does not quote directly, they may authorize an agent as per proforma of Manufacturer authorization form as given in the tender enquiry document to quote and enter into a contractual obligation.
- 2(a) The Manufacturer should have supplied and installed in last **Five** years from the date of Tender Opening, at least 25% of the quoted quantity (rounded off to next whole number) of the similar equipment meeting major parameters of technical specification which is functioning satisfactorily.
- 2(b) The Tenderers quoting as authorized representative of the manufacturer meeting the above criteria should have executed at least one contract in the last five years from the date of tender opening of medical equipment anywhere in India of the same manufacturer.
3. The bidders/ firms identifying as MSME and or start-up firms are exempted from fulfilling criteria at S. No. 2 (a) and 2(b) stated above. However, this does not exempt any bidder/ firm/ manufacturer from fulfilling the quality requirements.

*Note: “If the bidder is a MSME, it shall declare in the bid document the Udyog Aadhar Memorandum Number issued to it under the MSMED Act, 2006. If a MSME bidder do not furnish the UAM Number along with bid documents, such MSME unit will not be eligible for the benefits available under Public Procurement Policy for MSEs Order 2012.”*

**NOTE:**

1. The tenderer shall give an affidavit as under:  
**“We hereby certify that if at any time, information furnished by us is proved to be false or incorrect, we are liable for any action as deemed fit by the purchaser in addition to forfeiture of the earnest money.”**
2. In support of 2 (a) & 2 (b), the Tenderer shall furnish Performance statement in the enclosed Proforma ‘A’.  
 The manufacturer (Tenderer)/ Indian Agent shall furnish Satisfactory Performance Certificate in respect of above, duly translated in English and duly notarized in the country of origin, alongwith the tender.
3. The Tenderer shall furnish a brief write-up, packed with adequate data explaining and establishing his available capacity/capability (both technical and financial) to perform the Contract (if awarded) within the stipulated time period, after meeting all its current/present commitments. The Tenderer shall also furnish details of Equipment and Quality Control in the enclosed Section VIII.
4. Notwithstanding anything stated above, the Purchaser reserves the right to assess the Tenderer’s capability and capacity to perform the contract satisfactorily before deciding on award of Contract, should circumstances warrant such an assessment in the overall interest of the Purchaser.
- 5. The bidder should submit the manufacturer’s production capacity, meeting the quantity requirement and delivery schedule requirement of this tender document.**
6. The Purchaser reserves the right to ask for a free demonstration of the quoted equipment at a pre determined place acceptable to the purchaser for technical acceptability as per the tender specifications, before the opening of the Price Tender.



**PROFORMA 'A'**  
**PROFORMA FOR PERFORMANCE STATEMENT**  
(For the period of last five years)

Tender Reference No. : \_\_\_\_\_

Date of opening : \_\_\_\_\_

Time : \_\_\_\_\_

Name and address of the Tenderer : \_\_\_\_\_

Name and address of the manufacturer : \_\_\_\_\_

Order placed by (full address of Purchaser/ Consignee)	Order number and date	Description and quantity of ordered goods and services	Value of order (Rs.)	Date of completion of Contract		Remarks indicating reasons for delay if any	Have the goods been functioning Satisfactorily (attach documentary proof)**
				As per contract	Actual		
1	2	3	4	5	6	7	8

We hereby certify that if at any time, information furnished by us is proved to be false or incorrect, we are liable for any action as deemed fit by the purchaser in addition to forfeiture of the earnest money.

**Signature and seal of the Tenderer**

- \*\* The documentary proof will be a certificate from the consignee/ end user with cross-reference of order no. and date in the certificate along with a notarized certification authenticating the correctness of the information furnished.**
- \*\* The bidders are requested to submit the latest purchase order copies supplied to AIIMS, PGIMER, JIPMER, Institute of National importance for the specific model quoted along with the price bid.**

**Section – X**

**TENDER FORM**

Date\_\_\_\_\_

To  
CEO  
HLL Infra Tech Services Limited  
Procurement and Consultancy Division  
B-14 A, Sector -62, Noida -201307, Uttar Pradesh.

Ref. Your TE document No. \_\_\_\_\_ dated \_\_\_\_\_

We, the undersigned have examined the above mentioned TE document, including amendment/corrigendum No. \_\_\_\_\_, dated \_\_\_\_\_ (*if any*), the receipt of which is hereby confirmed. We now offer to supply and deliver\_\_\_\_\_ (*Description of goods and services*) in conformity with your above referred document **for the sum as shown in the price schedules attached herewith and made part of this tender**. If our tender is accepted, we undertake to supply the goods and perform the services as mentioned above, in accordance with the delivery schedule specified in the List of Requirements.

We further confirm that, if our tender is accepted, we shall provide you with a performance security of required amount in an acceptable form in terms of GCC clause 5, read with modification, if any, in Section - V – “Special Conditions of Contract”, for due performance of the contract.

We agree to keep our tender valid for acceptance as required in the GIT clause 20, read with modification, if any in Section - III – “Special Instructions to Tenderers” or for subsequently extended period, if any, agreed to by us. We also accordingly confirm to abide by this tender up to the aforesaid period and this tender may be accepted any time before the expiry of the aforesaid period. We further confirm that, until a formal contract is executed, this tender read with your written acceptance thereof within the aforesaid period shall constitute a binding contract between us.

We further understand that you are not bound to accept the lowest or any tender you may receive against your above-referred tender enquiry.

We confirm that we do not stand deregistered/banned/blacklisted by any Govt. Authorities.

We confirm that we fully agree to the terms and conditions specified in above mentioned TE document, including amendment/ corrigendum if any

(Signature with date)

(Name and designation)

Duly authorised to sign tender for and on  
behalf of

**SECTION – XI**  
**PRICE SCHEDULE**

Price to be filled in the relevant field of Price Format in Excel provided in the e-tendering portal.

**SECTION – XII**  
**QUESTIONNAIRE**

**Fill up the Techno-Commercial Compliance Sheet Bid provided in spreadsheet (Excel file) and upload in the C-Folder**

1. The tenderer should furnish specific answers to all the questions/issues mentioned in the Techno-Commercial Compliance Sheet. In case a question/issue does not apply to a tenderer, the same should be answered with the remark “not applicable”.
2. Wherever necessary and applicable, the tenderer shall enclose certified scanned copy as documentary proof/ evidence to substantiate the corresponding statement.
3. In case a tenderer furnishes a wrong or evasive answer against any of the question/issues, their tender is liable to be ignored.

Note: *The documents like Priced Proforma Invoice (Single Proforma Invoice from Manufacturer’s indicating uniform unit rates) and List of Consumables with prices can be uploaded in the Notes & Attachment under Rfx information (Please note, in the separate Notes & Attachment provided under Rfx information and not in the C-Folder Notes & Attachments).*

**SECTION – XIII**

**BANK GUARANTEE FORM FOR EMD**

Whereas \_\_\_\_\_ (hereinafter called the “Tenderer”) has submitted its quotation dated \_\_\_\_\_ for the supply of \_\_\_\_\_ (hereinafter called the “tender”) against the purchaser’s tender enquiry No. \_\_\_\_\_ Know all persons by these presents that we \_\_\_\_\_ of \_\_\_\_\_ (Hereinafter called the “Bank”) having our registered office at \_\_\_\_\_ are bound unto \_\_\_\_\_ (hereinafter called the “Purchaser) in the sum of \_\_\_\_\_ for which payment will and truly to be made to the said Purchaser, the Bank binds itself, its successors and assigns by these presents. Sealed with the Common Seal of the said Bank this \_\_\_\_\_ day of \_\_\_\_\_ 20\_\_\_\_. The conditions of this obligation are:

- 1) If the Tenderer withdraws or amends, impairs or derogates from the tender in any respect within the period of validity of this tender.
- 2) If the Tenderer having been notified of the acceptance of his tender by the Purchaser during the period of its validity:-

fails or refuses to furnish the performance security for the due performance of the contract or fails or refuses to accept/execute the contract or if it comes to notice that the information/documents furnished in its tender is incorrect, false, misleading or forged

We undertake to pay the Purchaser up to the above amount upon receipt of its first written demand, without the Purchaser having to substantiate its demand, provided that in its demand the Purchaser will note that the amount claimed by it is due to it owing to the occurrence of one or both the two conditions, specifying the occurred condition(s).

This guarantee will remain in force for a period of forty-five days after the period of tender validity and any demand in respect thereof should reach the Bank not later than the above date.

.....  
(Signature with date of the authorised officer of the Bank)

.....  
Name and designation of the officer

.....  
Seal, name & address of the Bank and address of the Branch

**SECTION – XIV**

**MANUFACTURER’S AUTHORISATION FORM**

CEO  
HLL Infra Tech Services Limited  
Procurement and Consultancy Division  
B-14 A, Sector -62, Noida -201307, Uttar Pradesh.

Dear Sir,

Ref: Your TE document No \_\_\_\_\_ dated \_\_\_\_\_

We, \_\_\_\_\_ who are proven and reputable manufacturers of \_\_\_\_\_ (*name and description of the goods offered in the tender*) having factories at \_\_\_\_\_, hereby authorise Messrs \_\_\_\_\_ (*name and address of the agent*) to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also state that we are not participating directly in this tender for the following reason(s):  
\_\_\_\_\_  
(*please provide reason here*).

We further confirm that no supplier or firm or individual other than Messrs. \_\_\_\_\_ (*name and address of the above agent*) is authorised to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

We also hereby extend our full warranty, CMC as applicable as per clause 15 of the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document.

We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorised agent

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly”

Yours faithfully,

[*Signature with date, name, designation and Email*]  
for and on behalf of Messrs \_\_\_\_\_  
[*Name & address of the manufacturers*]

Note:

- (1) *This letter of authorisation should be on the letter head of the manufacturing firm and should be signed by a person competent and having the power of attorney to legally bind the manufacturer.*
- (2) *Original letter may be sent.*
- (3) *The purchaser reserves the right to verify this document with its signatory.*

**SECTION – XV**

**BANK GUARANTEE FORM FOR PERFORMANCE SECURITY/ CMC SECURITY**

CEO  
HLL Infra Tech Services Limited  
Procurement and Consultancy Division  
B-14 A, Sector -62, Noida -201307, Uttar Pradesh.

WHEREAS \_\_\_\_\_ (Name and address of the supplier) (Hereinafter called “the supplier”) has undertaken, in pursuance of contract no \_\_\_\_\_ dated \_\_\_\_\_ to supply (description of goods and services) (herein after called “the contract”).

AND WHEREAS it has been stipulated by you in the said contract that the supplier shall furnish you with a bank guarantee by a scheduled commercial bank recognised by you for the sum specified therein as security for compliance with its obligations in accordance with the contract;

AND WHEREAS we have agreed to give the supplier such a bank guarantee;

NOW THEREFORE we hereby affirm that we are guarantors and responsible to you, on behalf of the supplier, up to a total of. \_\_\_\_\_ (Amount of the guarantee in words and figures), and we undertake to pay you, upon your first written demand declaring the supplier to be in default under the contract and without cavil or argument, any sum or sums within the limits of (amount of guarantee) as aforesaid, without your needing to prove or to show grounds or reasons for your demand or the sum specified therein.

We hereby waive the necessity of your demanding the said debt from the supplier before presenting us with the demand.

We further agree that no change or addition to or other modification of the terms of the contract to be performed there under or of any of the contract documents which may be made between you and the supplier shall in any way release us from any liability under this guarantee and we hereby waive notice of any such change, addition or modification.

This guarantee shall be valid till such time to cover two months beyond the warranty period from the date of Notification of Award i.e. up to \_\_\_\_\_ (indicate date).

.....  
(Signature with date of the authorised officer of the Bank)

.....  
Name and designation of the officer

.....  
Seal, name & address of the Bank and address of the Branch

**SECTION – XVI****CONTRACT FORM - A****CONTRACT FORM FOR SUPPLY, INSTALLATION, COMMISSIONING, HANDING OVER, TRIAL RUN, TRAINING OF OPERATORS & WARRANTY OF GOODS**

(Address of the Purchaser/Consignee  
Office issuing the contract)

Contract No \_\_\_\_\_ dated \_\_\_\_\_

**This is in continuation to this office's Notification of Award No \_\_\_\_\_ dated \_\_\_\_\_**

1. Name & address of the Supplier: \_\_\_\_\_
2. Purchaser's TE document No \_\_\_\_\_ dated \_\_\_\_\_ and subsequent Amendment No \_\_\_\_\_, dated \_\_\_\_\_ (if any), issued by the purchaser
3. Supplier's Tender No \_\_\_\_\_ dated \_\_\_\_\_ and subsequent communication(s) No \_\_\_\_\_ dated \_\_\_\_\_ (if any), exchanged between the supplier and the purchaser in connection with this tender.
4. In addition to this Contract Form, the following documents etc, which are included in the documents mentioned under paragraphs 2 and 3 above, shall also be deemed to form and be read and construed as integral part of this contract:
  - (i) General Conditions of Contract;
  - (ii) Special Conditions of Contract;
  - (iii) List of Requirements;
  - (iv) Technical Specifications;
  - (v) Quality Control Requirements;
  - (vi) Tender Form furnished by the supplier;
  - (vii) Price Schedule(s) furnished by the supplier in its tender;
  - (viii) Manufacturers' Authorisation Form (if applicable for this tender);
  - (ix) Purchaser's Notification of Award

Note: The words and expressions used in this contract shall have the same meanings as are respectively assigned to them in the conditions of contract referred to above. Further, the definitions and abbreviations incorporated under clause 1 of Section II – 'General Instructions to Tenderers' of the Purchaser's TE document shall also apply to this contract.

5. Some terms, conditions, stipulations etc. out of the above-referred documents are reproduced below for ready reference:
  - (i) Brief particulars of the goods and services which shall be supplied/ provided by the supplier are as under:

Schedule No.	Brief description of goods/services	Accounting unit	Quantity to be supplied	Unit Price	Total price	Terms of delivery

Any other additional services (if applicable) and cost thereof: \_\_\_\_\_

Total value (in figure) \_\_\_\_\_ (In words) \_\_\_\_\_

- (ii) Delivery schedule



- (iii) Details of Performance Security
- (iv) Quality Control
  - (a) Mode(s), stage(s) and place(s) of conducting inspections and tests.
  - (b) Designation and address of purchaser's inspecting officer
- (v) Destination and despatch instructions
- (vi) Consignee, including port consignee, if any

- 6. Warranty clause
- 7. Payment terms
- 8. Paying authority

\_\_\_\_\_  
(Signature, name and address  
of the Purchaser's/Consignee's authorised official)  
For and on behalf of \_\_\_\_\_

Received and accepted this contract  
(Signature, name and address of the supplier's executive  
duly authorised to sign on behalf of the supplier)  
For and on behalf of \_\_\_\_\_  
(Name and address of the supplier)  
(Seal of the supplier)

Date: \_\_\_\_\_

Place: \_\_\_\_\_

**CONTRACT FORM – B****CONTRACT FORM FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT**

Annual CM Contract No. \_\_\_\_\_

dated \_\_\_\_\_

Between

(Address of Head of Hospital )

And

(Name &amp; Address of the Supplier)

**Ref: Contract No. \_\_\_\_\_ dated \_\_\_\_\_ (Contract No. & date of Contract for supply, installation, commissioning, handing over, Trial run, Training of operators & warranty of goods)**

In continuation to the above referred contract

1. The Contract of Annual Comprehensive Maintenance is hereby concluded as under: -

1	2	3	4					5
Schedule No.	Brief description of goods	Quantity. (Nos.)	Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*.					Total Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x (4a+4b+4c+4d+4e) ]
			1 <sup>st</sup>	2 <sup>nd</sup>	3 <sup>rd</sup>	4 <sup>th</sup>	5 <sup>th</sup>	
			a	b	c	d	e	

Total value (in figure) \_\_\_\_\_ (In words) \_\_\_\_\_

- The CMC commence from the date of expiry of all obligations under Warranty i.e. from \_\_\_\_\_ (date of expiry of Warranty) and will expire on \_\_\_\_\_ (date of expiry of CMC)
- The cost of Annual Comprehensive Maintenance Contract (CMC) which includes preventive maintenance, labour and spares, after satisfactory completion of Warranty period may be quoted for next 5 years as contained in the above referred contract on yearly basis for complete equipment (including X ray tubes, Helium for MRI, Batteries for UPS, other vacuumatic parts, \_\_\_\_\_ & \_\_\_\_\_) and Site Modification Work (if any).
- There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- During CMC period, the supplier shall visit at each consignee's site for preventive maintenance including testing and calibration as per the manufacturer's service/ technical/ operational manual. The supplier shall visit each consignee site as recommended in the manufacturer's manual, but at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
- All software updates should be provided free of cost during CMC.

7. The bank guarantee valid till \_\_\_\_\_ [(fill the date) 2 months after expiry of entire CMC period] for an amount of Rs. \_\_\_\_\_ [(fill amount) equivalent to 2.5 % of the cost of the equipment as per contract] shall be furnished in the prescribed format given in Section XV of the TE document, along with the signed copy of Annual CMC within a period of 21 (twenty one) days of issue of Annual CMC failing which the proceeds of Performance Security shall be payable to the Purchaser/Consignee.
8. If there is any lapse in the performance of the CMC as per contract, the proceeds Annual CMC bank guarantee for an amount of Rs. \_\_\_\_\_ (equivalent to 2.5 % of the cost of the equipment as per contract) shall be payable to the Consignee.
9. **Payment terms:** The payment of Annual CMC will be made against the bills raised to the consignee by the supplier on six monthly basis after satisfactory completion of said period, duly certified by the HOD concerned. The payment will be made in Indian Rupees.
10. **Paying authority:** \_\_\_\_\_ (name of the consignee i.e. Hospital authorised official)

\_\_\_\_\_  
(Signature, name and address  
of Hospital authorised official)  
For and on behalf of \_\_\_\_\_

Received and accepted this contract.  
(Signature, name and address of the supplier's executive  
duly authorised to sign on behalf of the supplier)  
For and on behalf of \_\_\_\_\_  
(Name and address of the supplier)  
(Seal of the supplier)

Date: \_\_\_\_\_

Place: \_\_\_\_\_

**SECTION – XVII**

**CONSIGNEE RECEIPT CERTIFICATE**

(To be given by consignee's authorized representative)

The following store (s) has/have been received in good condition:

- 1) Contract No. & date : \_\_\_\_\_
- 2) Supplier's Name : \_\_\_\_\_
- 3) Consignee's Name & Address with  
telephone No. & Fax No. : \_\_\_\_\_
- 4) Name of the item supplied : \_\_\_\_\_
- 5) Quantity Supplied : \_\_\_\_\_
- 6) Date of Receipt by the Consignee : \_\_\_\_\_
- 7) Name and designation of Authorized  
Representative of Consignee : \_\_\_\_\_
- 8) Signature of Authorized Representative of  
Consignee with date : \_\_\_\_\_
- 9) Seal of the Consignee : \_\_\_\_\_

**SECTION – XVIII**  
**Proforma of Final Acceptance Certificate by the Consignee**

**No** \_\_\_\_\_

**Date** \_\_\_\_\_

**To**

M/s \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Subject:** Certificate of commissioning of equipment /plant.

This is to certify that the equipment (s)/plant(s) as detailed below has/have been received in good conditions along with all the standard and special accessories and a set of spares (subject to remarks in Para no.02) in accordance with the contract/technical specifications. The same has been installed and commissioned.

- (a) Contract No \_\_\_\_\_ dated \_\_\_\_\_
- (b) Description of the equipment (s)/plants: \_\_\_\_\_
- (c) Equipment (s)/ plant(s) nos.: \_\_\_\_\_
- (d) Quantity: \_\_\_\_\_
- (e) Bill of Loading/Air Way Bill/Railway Receipt/ Goods Consignment Note no \_\_\_\_\_ dated \_\_\_\_\_
- (f) Name of the vessel/Transporters: \_\_\_\_\_
- (g) Name of the Consignee: \_\_\_\_\_
- (h) Date of site hand-over to the supplier by consignee: \_\_\_\_\_
- (i) Date of commissioning and proving test: \_\_\_\_\_

**Details of accessories/spares not yet supplied and recoveries to be made on that account.**

Sl. No.	Description of Item	Quantity	Amount to be recovered

The proving test has been done to our entire satisfaction and operators have been trained to operate the equipment (s)/plant(s).

The supplier has fulfilled its contractual obligations satisfactorily ## or

The supplier has failed to fulfil its contractual obligations with regard to the following:

- a) He has not adhered to the time schedule specified in the contract in dispatching the documents/ drawings pursuant to ‘Technical Specifications’.
- b) He has not supervised the commissioning of the equipment (s)/plant(s) in time, i.e. within the period specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment (s)/plant(s).
- c) The supplier as specified in the contract has not done training of personnel.

The extent of delay for each of the activities to be performed by the supplier in terms of the contract is.....

The amount of recovery on account of non-supply of accessories and spares is given under Para no.02

The amount of recovery on account of failure of the supplier to meet his contractual obligations is \_\_\_\_\_ (here indicate the amount).

*(Signature)*

*(Name)*

*(Designation with stamp)*

**## Explanatory notes for filling up the certificate:**

- i) He has adhered to the time schedule specified in the contract in dispatching the documents/drawings pursuant to 'Technical Specification'.
- ii) He has supervised the commissioning of the equipment (s)/plant(s) in time, i.e. within the time specified in the contract from date of intimation by the Purchaser/Consignee in respect of the installation of the equipment (s)/plant(s).
- iii) Training of personnel has been done by the supplier as specified in the contract.
- iv) In the event of documents/drawings having not been supplied or installation and commissioning of the equipment (s)/plant(s) having been delayed on account of the supplier, the extent of delay should always be mentioned in clear terms.

## Section – XIX

## Consignee List

Sl. No.	Name of Hopsital and Address	Consignee Code	State	Airport	Dry Port/ Seaport
1	The Director, All India Institute of Medical Science, AIIMS-Bhubaneswar, Near BijuPatnaik Police Academy, Village-Sijua, Bhubaneswar- 751019, Orissa	Bhubaneswar	Odisha	KOLKAT A	KOLKATA

**NB: The consignee will ensure timely issue of NMIC, CDEC, Octroi Exemption Certificates, Road Permits & Entry Tax Exemption Certificates, wherever applicable, to the suppliers.**

**APPENDIX – A**

No. P-45021/2/2017-B.E.-II  
Government of India  
Ministry of Commerce and Industry  
Department of Industrial Policy and Promotion  
\*\*\*\*

Dated 15<sup>th</sup> June, 2017  
Udyog Bhawan, New Delhi

To

All Central Ministries/Departments/CPSUs/All concerned

**ORDER**

**Subject: Public Procurement (Preference to Make in India), Order 2017**

**Whereas** it is the policy of the Government of India to encourage 'Make in India' and promote manufacturing and production of goods and services in India with a view to enhancing income and employment, and

**Whereas** procurement by the Government is substantial in amount and can contribute towards this policy objective, and

**Whereas** local content can be increased through partnerships, cooperation with local companies, establishing production units in India or Joint Ventures (JV) with Indian suppliers, increasing the participation of local employees in services and training them,

**Now therefore the following Order is issued :**

1. This Order is issued pursuant to Rule 153 (iii) of the General Financial Rules 2017.
2. **Definitions:** For the purposes of this Order:

*'Local content'* means the amount of value added in India which shall, unless otherwise prescribed by the Nodal Ministry, be the total value of the item procured (excluding net domestic indirect taxes) minus the value of imported content in the item (including all customs duties) as a proportion of the total value, in percent.

*'Local supplier'* means a supplier or service provider whose product or service offered for procurement meets the minimum local content as prescribed under this Order or by the competent Ministries / Departments in pursuance of this order.

*'L1'* means the lowest tender or lowest bid or the lowest quotation received in a tender, bidding process or other procurement solicitation as adjudged in the evaluation process as per the tender or other procurement solicitation.

*'margin of purchase preference'* means the maximum extent to which the price quoted by a local supplier may be above the L1 for the purpose of purchase preference.

*'Nodal Ministry'* means the Ministry or Department identified pursuant to this order in respect of a particular item of goods or services.

.....Contd. p.2/-



'Procuring entity' means a Ministry or department or attached or subordinate office of, or autonomous body controlled by, the Government of India and includes Government companies as defined in the Companies Act.

3. **Requirement of Purchase Preference:** Subject to the provisions of this Order and to any specific instructions issued by the Nodal Ministry or in pursuance of this Order, purchase preference shall be given to local suppliers in all procurements undertaken by procuring entities in the manner specified hereunder:
- a. In procurement of goods in respect of which the Nodal Ministry has communicated that there is sufficient local capacity and local competition, and where the estimated value of procurement is Rs. 50 lakhs or less, only local suppliers shall be eligible. If the estimated value of procurement of such goods is more than Rs. 50 lakhs, the provisions of sub-paragraph b or c, as the case may be, shall apply.
  - b. In the procurements of goods which are not covered by paragraph 3a and which are divisible in nature, the following procedure shall be followed:
    - i. Among all qualified bids, the lowest bid will be termed as L1. If L1 is from a local supplier, the contract for full quantity will be awarded to L1.
    - ii. If L1 bid is not from a local supplier, 50% of the order quantity shall be awarded to L1. Thereafter, the lowest bidder among the local suppliers, will be invited to match the L1 price for the remaining 50% quantity subject to the local supplier's quoted price falling within the margin of purchase preference, and contract for that quantity shall be awarded to such local supplier subject to matching the L1 price. In case such lowest eligible local supplier fails to match the L1 price or accepts less than the offered quantity, the next higher local supplier within the margin of purchase preference shall be invited to match the L1 price for remaining quantity and so on, and contract shall be awarded accordingly. In case some quantity is still left uncovered on local suppliers, then such balance quantity may also be ordered on the L1 bidder.
  - c. In procurements of goods not covered by sub-paragraph 3a and which are not divisible, and in procurement of services where the bid is evaluated on price alone, the following procedure shall be followed:
    - i. Among all qualified bids, the lowest bid will be termed as L1. If L1 is from a local supplier, the contract will be awarded to L1.
    - ii. If L1 is not from a local supplier, the lowest bidder among the local suppliers, will be invited to match the L1 price subject to local supplier's quoted price falling within the margin of purchase preference, and the contract shall be awarded to such local supplier subject to matching the L1 price.
    - iii. In case such lowest eligible local supplier fails to match the L1 price, the local supplier with the next higher bid within the margin of purchase preference shall be invited to match the L1 price and so on and contract shall be awarded accordingly. In case none of the local suppliers within the margin of purchase preference matches the L1 price, then the contract may be awarded to the L1 bidder.

.....Contd. p.3/-

4. **Exemption of small purchases:** Notwithstanding anything contained in paragraph 3, procurements where the estimated value to be procured is less than Rs. 5 lakhs shall be exempt from this Order. However, it shall be ensured by procuring entities that procurement is not split for the purpose of avoiding the provisions of this Order.
5. **Minimum local content:** The minimum local content shall ordinarily be 50%. The Nodal Ministry may prescribe a higher or lower percentage in respect of any particular item and may also prescribe the manner of calculation of local content.
6. **Margin of Purchase Preference:** The margin of purchase preference shall be 20% .
7. **Requirement for specification in advance:** The minimum local content, the margin of purchase preference and the procedure for preference to Make in India shall be specified in the notice inviting tenders or other form of procurement solicitation and shall not be varied during a particular procurement transaction.
8. **Government E-marketplace:** In respect of procurement through the Government E-marketplace (GeM) shall, as far as possible, specifically mark the items which meet the minimum local content while registering the item for display, and shall, wherever feasible, make provision for automated comparison with purchase preference and without purchase preference and for obtaining consent of the local supplier in those cases where purchase preference is to be exercised.
9. **Verification of local content:**
  - a. The local supplier at the time of tender, bidding or solicitation shall be required to provide self-certification that the item offered meets the minimum local content and shall give details of the location(s) at which the local value addition is made.
  - b. In cases of procurement for a value in excess of Rs. 10 crores, the local supplier shall be required to provide a certificate from the statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.
  - c. Decisions on complaints relating to implementation of this Order shall be taken by the competent authority which is empowered to look into procurement-related complaints relating to the procuring entity.
  - d. Nodal Ministries may constitute committees with internal and external experts for independent verification of self-declarations and auditor's/ accountant's certificates on random basis and in the case of complaints.
  - e. Nodal Ministries and procuring entities may prescribe fees for such complaints.
  - f. False declarations will be in breach of the Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151 (iii) of the General Financial Rules along with such other actions as may be permissible under law.
  - g. A supplier who has been debarred by any procuring entity for violation of this Order shall not be eligible for preference under this Order for procurement by any other procuring entity for the

..... Contd.p.4/-

duration of the debarment. The debarment for such other procuring entities shall take effect prospectively from the date on which it comes to the notice of other procurement entities, in the manner prescribed under paragraph 9h below.

- h. The Department of Expenditure shall issue suitable instructions for the effective and smooth operation of this process, so that:
- i. The fact and duration of debarment for violation of this Order by any procuring entity are promptly brought to the notice of the Member-Convenor of the Standing Committee and the Department of Expenditure through the concerned Ministry /Department or in some other manner;
  - ii. on a periodical basis such cases are consolidated and a centralized list or decentralized lists of such suppliers with the period of debarment is maintained and displayed on website(s);
  - iii. in respect of procuring entities other than the one which has carried out the debarment, the debarment takes effect prospectively from the date of uploading on the website(s) in the such a manner that ongoing procurements are not disrupted.

**10. Specifications in Tenders and other procurement solicitations:**

- a. Every procuring entity shall ensure that the eligibility conditions in respect of previous experience fixed in any tender or solicitation do not require proof of supply in other countries or proof of exports.
- b. Procuring entities shall endeavour to see that eligibility conditions, including on matters like turnover, production capability and financial strength do not result in unreasonable exclusion of local suppliers who would otherwise be eligible, beyond what is essential for ensuring quality or creditworthiness of the supplier.
- c. Procuring entities shall, within 2 months of the issue of this Order review all existing eligibility norms and conditions with reference to sub-paragraphs 'a' and 'b' above.
- d. If a Nodal Ministry is satisfied that Indian suppliers of an item are not allowed to participate and/ or compete in procurement by any foreign government, it may, if it deems appropriate, restrict or exclude bidders from that country from eligibility for procurement of that item and/ or other items relating to that Nodal Ministry. A copy of every instruction or decision taken in this regard shall be sent to the Chairman of the Standing Committee.
- e. For the purpose of sub-paragraph 10 d above, a supplier or bidder shall be considered to be from a country if (i) the entity is incorporated in that country, or (ii) a majority of its shareholding or effective control of the entity is exercised from that country; or (iii) more than 50% of the value of the item being supplied has been added in that country. Indian suppliers shall mean those entities which meet any of these tests with respect to India."

.....Contd.p.5/-

11. **Assessment of supply base by Nodal Ministries:** The Nodal Ministry shall keep in view the domestic manufacturing / supply base and assess the available capacity and the extent of local competition while identifying items and prescribing minimum local content or the manner of its calculation, with a view to avoiding cost increase from the operation of this Order.
12. **Increase in minimum local content:** The Nodal Ministry may annually review the local content requirements with a view to increasing them, subject to availability of sufficient local competition with adequate quality.
13. **Manufacture under license/ technology collaboration agreements with phased indigenization:** While notifying the minimum local content, Nodal Ministries may make special provisions for exempting suppliers from meeting the stipulated local content if the product is being manufactured in India under a license from a foreign manufacturer who holds intellectual property rights and where there is a technology collaboration agreement / transfer of technology agreement for indigenous manufacture of a product developed abroad with clear phasing of increase in local content.
14. **Powers to grant exemption and to reduce minimum local content:** Ministries /Departments of Government of India and the Boards of Directors of Government companies or autonomous bodies may, by written order,
  - a. reduce the minimum local content below the prescribed level;
  - b. reduce the margin of purchase preference below 20% ;
  - c. exempt any particular item or procuring or supplying entities or class or classes of items or procuring or supplying entities from the operation of this Order or any part of the Order.A copy of every such order shall be marked to the Member-Convenor of the Standing Committee constituted under this Order.
15. **Directions to Government companies:** In respect of Government companies and other procuring entities not governed by the General Financial Rules, the administrative Ministry or Department shall issue policy directions requiring compliance with this Order.
16. **Standing Committee:** A standing committee is hereby constituted with the following membership:
  - Secretary, Department of Industrial Policy and Promotion—Chairman
  - Secretary, Commerce—Member
  - Secretary, Ministry of Electronics and Information Technology—Member
  - Joint Secretary (Public Procurement), Department of Expenditure—Member
  - Joint Secretary (DIPP)—Member-ConvenorThe Secretary of the Department concerned with a particular item shall be a member in respect of issues relating to such item. The Chairman of the Committee may co-opt technical experts as relevant to any issue or class of issues under its consideration.

..... Contd.p.6/-

-6-

17. **Functions of the Standing Committee:** The Standing Committee shall meet as often as necessary but not less than once in six months. The Committee
- shall oversee the implementation of this order and issues arising therefrom, and make recommendations to Nodal Ministries and procuring entities.
  - shall annually assess and periodically monitor compliance with this Order
  - shall identify Nodal Ministries and the allocation of items among them for issue of notifications on minimum local content
  - may require furnishing of details or returns regarding compliance with this Order and related matters
  - may, during the annual review or otherwise, assess issues, if any, where it is felt that the manner of implementation of the order results in any restrictive practices, cartelization or increase in public expenditure and suggest remedial measures
  - may examine cases covered by paragraph 13 above relating to manufacture under license/ technology transfer agreements with a view to satisfying itself that adequate mechanisms exist for enforcement of such agreements and for attaining the underlying objective of progressive indigenization
  - may consider any other issue relating to this Order which may arise.
18. **Removal of difficulties:** Ministries /Departments and the Boards of Directors of Government companies may issue such clarifications and instructions as may be necessary for the removal of any difficulties arising in the implementation of this Order.
19. **Ministries having existing policies:** Where any Ministry or Department has its own policy for preference to local content approved by the Cabinet after 1<sup>st</sup> January 2015, such policies will prevail over the provisions of this Order. All other existing orders on preference to local content shall be reviewed by the Nodal Ministries and revised as needed to conform to this Order, within two months of the issue of this Order.
20. **Transitional provision:** This Order shall not apply to any tender or procurement for which notice inviting tender or other form of procurement solicitation has been issued before the issue of this Order.

(B. S. Nayak)  
Under Secretary to Government of India  
Ph. 23061257

**APPENDIX-B**  
**INTEGRITY PACT**

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

**PRE-CONTRACT INTEGRITY PACT**

This Pre-Contract Integrity Pact (herein after called the Integrity Pact) is made on \_\_\_\_\_ day of the month of February 2012,

**Between**

HLL Infra Tech Services Ltd. [HITES], a wholly owned subsidiary company of M/s. HLL Lifecare Ltd. a Government of India Enterprise with registered office at HLL Bhavan, Poojappura, Thiruvananthapuram 695 012, Kerala, India. (Hereinafter called "HITES", which expression shall mean and include, unless the context otherwise requires, his successors in office and assigns) of the First Party.

**And**

M/s \_\_\_\_\_ with office at \_\_\_\_\_ represented by Shri \_\_\_\_\_, Chief Executive Officer (hereinafter called the "BIDDER/Seller"/Contractor which expression shall mean and include, unless the context otherwise requires, his successors and permitted assigns) of the Second Party.

**Preamble**

[Both HITES and BIDDER referred above are jointly referred to as the Parties]

HITES intends to award, under laid down organizational procedures, Purchase orders / contract/s against Tender /Work Order /Purchase Order No. .... HITES desires full compliance with all relevant laws and regulations, and the principles of economic use of resources, and of fairness and transparency in its relations with its Bidder/s and Contractor/s.

**NOW, THEREFORE,**

To avoid all forms of corruption by following a system that is fair, transparent and free from any influence/prejudiced dealings prior to, during and subsequent to the currency of the contract to be entered into with a view to:-

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

1. Enable HITES to obtain the desired materials/ stores/equipment/ work/ project done at a competitive price in conformity with the defined specifications by avoiding the high cost and the distortionary impact of corruption on public procurement; and
2. Enable the BIDDER to abstain from bribing or indulging in any corrupt practice in order to secure the contract by providing assurance to them that their competitors will also abstain from bribing and other corrupt practices and HITES will commit to prevent corruption, in any form, by its officials by following transparent procedures.

The parties hereto hereby agree to enter into this Integrity Pact and agree as follows:

**Clause.1. Commitments of HITES**

- 1.1 HITES undertakes that HITES and/or its Associates (i.e. employees, agents, consultants, advisors, etc.) will not demand, take a promise for or accept, directly or through intermediaries, any bribe, consideration, gift, reward, favour or any material or immaterial benefit or any other advantage from the BIDDER, either for themselves or for any person, organization or third party related to the contract in exchange for an advantage in the bidding process, bid evaluation, contracting or implementation process related to the contract.
- 1.2 HITES will, during the tender process / pre-contract stage, treat all BIDDERS with equity and reason, and will provide to all BIDDERS the same information and will not provide any such information or additional information, which is confidential in any manner, to any particular BIDDER which could afford an advantage to that particular BIDDER in comparison to other BIDDERS in relation to tendering process or during the contract execution.
- 1.3 All the officials of HITES will report to Chief Vigilance Officer of HITES (CVO), any attempted or completed breaches of the above commitments as well as any substantial suspicion of such a breach.
- 1.4 HITES will exclude from the process all known prejudiced persons and persons who would be known to have a connection or nexus with the prospective bidder.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

- 1.5 If the BIDDER reports to HITES with full and verifiable facts any misconduct on the part of HITES's Associates (i.e. employees, agents, consultants, advisors, etc.) and the same is prima facie found to be correct by HITES, necessary disciplinary proceedings, or any other action as deemed fit, including criminal proceedings may be initiated by HITES. Further, such an Associate may be debarred from further dealings related to the contract process. In such a case, while an enquiry is being conducted by HITES the proceedings under the contract would not be stalled.

**Clause 2. Commitments of BIDDERS/ CONTRACTORS**

2. The BIDDER commits itself to take all measures necessary to prevent corrupt practices, unfair means and illegal activities during any stage of its bid or during any pre-contract or post-contract stage in order to secure the contract or in furtherance to secure it and in particular commit itself to the following:-

- 2.1 The BIDDER will not offer, directly or indirectly (i.e. employees, agents, consultants, advisors, etc.) any bribe, gift, consideration, reward, favour, any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of HITES, connected directly or indirectly with the bidding process, or to any person, organization or third party related to the contract in exchange for any advantage in the bidding, evaluation, contracting and implementation of the contract.

- 2.2 The BIDDER further undertakes that it has not given, offered or promised to give, directly or indirectly any bribe, gift, consideration, reward, favour, any material or immaterial benefit or other advantage, commission, fees, brokerage or inducement to any official of HITES or otherwise in procuring the contract or forbearing to do or having done any act in relation to obtaining or execution of the contract or any other contract with the Government for showing or forbearing to show favour or disfavor to any person in relation to the contract or any other contract with the Government.

- 2.3 The BIDDER will not engage in collusion, price fixing, cartelization, etc. with other counterparty(s).



HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

- 2.4 The counterparty will not pass to any third party any confidential information entrusted to it, unless duly authorized by HITES.
- 2.5 The counterparty will promote and observe ethical practices within its Organization and its affiliates.
- 2.6 BIDDER shall disclose the name and address of agents and representatives and Indian BIDDERS shall disclose their foreign principals or associates.
- 2.7 The counterparty will not make any false or misleading allegations against HITES or its Associates.
- 2.8 BIDDERS shall disclose the payments to be made by them to agents/brokers or any other intermediary, in connection with this bid/contract.
- 2.9 The BIDDER further confirms and declares to HITES that the BIDDER is the original manufacture/integrator/authorized government sponsored export entity of the defense stores and has not engaged any individual or firm or company whether Indian or foreign to intercede, facilitate or in any way to recommend to HITES or any of its functionaries, whether officially or unofficially to award the contract to the BIDDER, nor has any amount been paid, promised or intended to be paid to any such individual, firm or company in respect of any such intercession, facilitation or recommendation.
- 2.10 The BIDDER while presenting the bid or during pre-contract negotiations or before signing the contract, shall disclose any payments he has made, is committed to or intends to make to officials of HITES or their family members, agents, brokers or any other intermediaries in connection with the contract and the details of services agreed upon for such payments.
- 2.11 The BIDDER will not accept any advantage in exchange for any corrupt practice, unfair means and illegal activities.
- 2.12 The BIDDER commits to refrain from giving any complaint directly or through any other manner without supporting it with full and verifiable facts.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

2.13 If the BIDDER or any employee of the BIDDER or any person acting on behalf of the BIDDER, either directly or indirectly, is a relative of any of the officers of HITES, or alternatively, if any relative of an officer of HITES has financial interest/stake in the BIDDER's firm, the same shall be disclosed by the BIDDER at the time of filing of tender.

The term 'relative' for this purpose would be as defined in Section 6 of the Companies Act 1956.

2.14 The BIDDER shall not lend to or borrow any money from or enter into any monetary dealings or transactions, directly or indirectly, with any employee of HITES.

2.15 The BIDDER will not collude with other parties interested in the contract to impair the transparency, fairness and progress of the bidding process, bid evaluation, contracting and implementation of the contract, and will not enter into any undisclosed agreement or understanding with other Bidders, whether formal or informal. This applies in particular to prices, specifications, certifications, subsidiary contracts, submission or non-submission of bids or any other actions to restrict competitiveness or to introduce cartelization in the bidding process.

2.16 The BIDDER will not commit any offence under the relevant Indian Penal Code, 1860 or Prevention of Corruption Act, 1988; further the Bidder(s)/ Contractor(s) will not use improperly, for purposes of competition or personal gain, or pass on to others, any information or document provided by the HITES as part of the business relationship, regarding plans, technical proposals and business details, including information contained or transmitted electronically. The BIDDER also undertakes to exercise due and adequate care lest any such information is divulged.

2.17 The BIDDER will not instigate third persons to commit offences outlined above or be an accessory to such offences.

2.18 The Bidder(s)/Contractors(s) of foreign origin shall disclose the name and address of the Agents/representatives in India, if any. Similarly the Bidder(s)/Contractors(s) of Indian Nationality shall furnish the name and address of the foreign Principal(s), if any.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

**Clause.3. Previous contravention and Disqualification from tender process and exclusion from future contracts**

**3.1**

The BIDDER declares that no previous contravention occurred in the last three years immediately before signing of this Integrity Pact, with any other company in any country in respect of any corrupt practices envisaged hereunder or with any Public Sector Enterprise in India or any Government Department in India that could justify BIDDER's exclusion from the tender process

**3.2**

The BIDDER agrees that if it makes incorrect statement on this subject, BIDDER can be disqualified from the tender process or the contract, if already awarded, can be terminated for such reason.

If BIDDER before award or during execution has committed a contravention through a violation of Clause 2, above or in any other form such as to put his reliability or credibility in question, HITES is entitled to disqualify the BIDDER from the tender process.

**Clause .4. Equal treatment of all Bidders / Contractors / Subcontractors**

**4.1**

The Bidder(s)/ Contractor(s) undertake(s) to demand from his Subcontractors a commitment in conformity with this Integrity Pact.

**4.2**

HITES will enter into agreements with identical conditions as this one with all Bidders and Contractors.

**4.3**

HITES will disqualify from the tender process all bidders who do not sign this Pact or violate its provisions.

**Clause .5. Consequences of Violation / Breach**

5.1 Any breach of the aforesaid provision by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER) shall entitle HITES to take all or any one of the following action, wherever required:-

HLL Infra Tech Services I.td.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

- i. To Immediately call off the pre-contract negotiations without assigning any reason or giving any compensation to the BIDDER. However, the proceedings with the other BIDDER(s) would continue.
- ii. If BIDDER commits violation of Integrity Pact Policy during bidding process, he shall be liable to compensate HITES by way of liquidated damages amounting to a sum equivalent to 5% to the value of the offer or the amount equivalent to Earnest Money Deposit/Bid Security, whichever is higher.
- iii. In case of violation of the Integrity Pact after award of the contract, HITES will be entitled to terminate the contract. HITES shall also be entitled to recover from the contractor liquidated damages equivalent to 10% of the contract value or the amount equivalent to security deposit/ performance guarantee, whichever is higher.
- iv. To immediately cancel the contract, if already signed, without giving any compensation to the BIDDER.
- v. To recover all sums already paid by HITES, and in case of an Indian BIDDER with interest thereon at 2% higher than the prevailing Prime Lending Rate of State Bank of India, while in case of a BIDDER from a country other than India with interest thereon at 2% higher than the LIBOR. If any outstanding payment is due to the BIDDER from HITES in connection with any other contract for any other stores, such outstanding payment could also be utilized to recover the aforesaid amount.
- vi. To encash the advance bank guarantee and performance guarantee /warranty bond, if furnished by the BIDDER, in order to recover the payments already made by HITES, along with interest.
- vii. To cancel all or any other contract with the BIDDER. The BIDDER shall be liable to pay compensation for any loss or damage to HITES resulting from such cancellation/recession and HITES shall be entitled to deduct the amount so payable from the money(s) due to the BIDDER.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

- viii. To debar the BIDDER from participating in future bidding processes of HITES for a minimum period of five (5) years, which may be further extended at the discretion of HITES or until Independent External Monitors is satisfied that the Counterparty will not commit any future violation.
- ix. To recover all sums paid in violation of this Pact by BIDDER(s) to any middleman or agent or broker with a view to securing the contract.
- x. In cases where irrevocable Letters of credit have been received in respect of any contract signed by HITES with the BIDDER, the same shall not be opened.
- xi. Forfeiture of performance guarantee in case of a decision by HITES to forfeit the same without assigning any reason for imposing sanction for violation of the pact.

5.2 HITES will be entitled to all or any of the actions mentioned in para 5.1(i) to (x) of this pact also on the commission by the BIDDER or any one employed by it or acting on its behalf (whether with or without the knowledge of the BIDDER), of an offence as defined in Chapter IX of the Indian Penal Code, 1860 or Prevention of Corruption Act, 1988 or any other statute enacted for prevention of corruption.

5.3 The decision of HITES to the effect that a breach of the provisions of this Pact has been committed by the BIDDER shall be final and conclusive on the BIDDER. However, the BIDDER can approach the Independent External Monitor(s) appointed for the purposes of this Pact.

**Clause.6. Fall Clause**

The BIDDER undertakes that it has not supplied/is not supplying similar product/systems or subsystems OR providing similar services at a price / charge lower than that offered in the present bid in respect of any other Ministry/Department of the Government of India or PSU and if it is found any stage that similar product/systems or sub systems was supplied by the BIDDER to any to the Ministry/Department of the Government of India or a PSU at a lower price, then that very price, with due allowance for elapsed time will be applicable to the present case and the difference in the cost would be refunded by the BIDDER to HITES, if the contract has already been concluded.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

**Clause .7. Independent External Monitor(s)**

- 7.1 HITES has appointed Independent External Monitor(s) (hereinafter referred to as Monitor(s)) for this Pact in consultation with the Central Vigilance Commission (Name and addresses of the Monitor(s) to be given).
- 7.2 The responsibility of the Monitor(s) shall be to review independently and objectively, whether and to what extent the parties comply with the obligations under this Pact.
- 7.3 The Monitor(s) shall not be subject to instructions by the representatives of the parties and perform their functions neutrally and independently.
- 7.4 Both the parties accept that the Monitor(s) have the right to access all the documents relating to the project/ procurement, including minutes of meetings.
- 7.5 As soon as the Monitor(s) notices, or has reason to believe, a violation of this pact, he will so inform the CVO.
- 7.6 The BIDDER(S) accepts that the Monitor(s) have the right to access without restriction to all project documentation of HITES including that provided by the BIDDER. The BIDDER will also grant the Monitor(s), upon his request and demonstration of a valid interest, unrestricted and unconditional access to his project documentation. The same is applicable to subcontractors engaged by the BIDDER. The Monitor(s) shall be under contractual obligation to treat the information and documents of the BIDDER/ Subcontractor(s) with confidentiality.
- 7.7 HITES will provide to the Monitor(s) sufficient information about all meetings among the parties related to the Project provided such meeting could have an impact on the contractual relation between the parties. The parties will offer to the Monitor(s) option to participate in such meetings.
- 7.8 The Monitor(s) will submit a written report to the CVO of HITES within 8 to 10 weeks from the date of reference or intimation to him by HITES/BIDDER and, should consent arise, submit proposals for correcting problematic situations.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

**Clause.8.Criminal charges against violating Bidder(s)/  
\_\_\_\_\_ Contractor(s)/ Subcontractor(s)**

If HITES obtains knowledge of conduct of a Bidder, Contractor or Subcontractor, or of an employee or a representative or an associate of a Bidder, Contractor or Subcontractor which constitutes corruption, or if HITES has substantive suspicion in this regard, HITES will inform the same to the Chief Vigilance Officer.

**Clause.9. Facilitation of Investigation**

In case of any allegation of violation of any provisions of this Pact or payment of commission, HITES or its agencies shall be entitled to examine all the documents, including the Books of Accounts of the BIDDER and the BIDDER shall provide necessary information and documents in English and shall extend all possible help for the purpose of such examination.

**Clause.10. Law and Place of Jurisdiction**

Both the Parties agree that this Pact is subject to Indian Law. The place of performance and hence this Pact shall be subject to Thiruvananthapuram Jurisdiction.

**Clause.11. Other legal Actions**

The actions stipulated in the Integrity Pact are without prejudice to any other legal action that may follow in accordance with the provisions of the extant law in force relating to any civil or criminal proceedings.

**Clause.12. Validity and Duration of the Agreement**

This Pact begins when both parties have legally signed it. It expires for the Contractor/Successful bidder 12 months after the last payment under the contract or the complete execution of the contract to the satisfaction of the both HITES and the BIDDER /Seller, including warranty period, whichever is later, and for all other Bidders/unsuccessful bidders 6 months after the contract has been awarded.

If any claim is made / lodged during this time, the same shall be binding and continue to be valid despite the lapse of this pact as specified above, unless it is discharged / determined by Chairman and Managing Director of HITES.

HLL Infra Tech Services Ltd.

Division : \_\_\_\_\_

Tender No: \_\_\_\_\_

**Clause. 13. Other provisions**

13.1 Changes and supplements as well as termination notices need to be made in writing. Both the Parties declare that no side agreements have been made to this Integrity Pact.

13.1 If the Contractor is a partnership or a consortium, this agreement must be signed by all partners or consortium members.

13.1 Should one or several provisions of this agreement turn out to be invalid, the remainder of this agreement remains valid. In this case, the parties will strive to come to an agreement to their original intentions

INWITNESS THEREOF the parties have signed and executed this pact at the place and date first above mentioned in the presents of following witnesses:

**HLL Infra Tech Services Ltd.**

**Bidder**

\_\_\_\_\_  
Director & CEO

\_\_\_\_\_  
Chief Executive

Witness

Witness

1.....

1.....

2.....

2.....

\* Provisions of these clauses would be amended /deleted in line with the policy of the HITES in regard to involvement of Indian agents of foreign suppliers.