

Amendment No. 6

Date: 11/04/2017

Sub: Amendment No.06 to the Tender Enquiry Document for extension of due date and Clarification to Technical specifications

Ref: i. Tender Enquiry No.: HLL/PCD/PMSSY-II/12/16-17 dated 17.01.2017.

ii. Amendment No. 5 dated 28/03/2017.

The due date for submission of bids is revised as follows:

Sch no	Event Number	Name of Item	Closing date & time for submission of online bids	Closing date & time for submission of tender fee and EMD in physical form	Time and date of opening of online bids
2	3000001614	Neurosurgical Operating microscope	20.04.2017 18:00	21.04.2017 11:30	21.04.2017 14:00

Technical Clarification**Schedule No. 02****Neurosurgical Operating Microscope****1. Existing Specification:**

L. 1. Should have installed ICG, ALA **and** sodium fluorescence (Yellow -560)

Read As:

L. 1. Should have installed ICG, ALA **or** sodium fluorescence (Yellow -560)

Note:

All the three facilities should be available with the machine.

The consignee will choose ALA or sodium fluorescence (Yellow -560) based on their requirement.

2. Existing Specification:

N. Accessories: **(Price should be quoted separately and will be freezed for 5 years)**

1. Spare Xenon lamp - 1 no
2. Disposable drapes for microscopes- 250nos
3. Vials for Sodium Fluorescence - 560: 10 nos
4. Vials for ICG drugs - 50 nos
5. **Vials for ALA drugs – 5 nos**

Read As:

N. Accessories:

1. **Spare Xenon lamp - 2 nos (Price should be quoted separately and will be freezed for 5 years)**
2. Disposable drapes for microscopes- 250nos
3. Vials for Sodium Fluorescence - 560: 10 nos
4. Vials for ICG drugs - 50 nos
5. Deleted

NOTE: If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e., up to 03.10.2017.

All other terms and conditions of the tender enquiry remain unaltered