

Sub: Augmentation of Nuclear Medicine Department with PET- CT unit & Dual Head Gamma Camera at JIPMER, Puducherry

Tender no.: HLL/ID/13/102 Dated 04.12.2013

S. NO	Query received from Neo Rad, India for PET CT tender through email	REPLY BY HLL
1	As we mentioned earlier, page 97 clause 5 in pet ct tender, trajectory Planning software comes default with any pet ct and no need for Additional trajectory planning or treatment planning software in needle Positioning system. Hence we kindly request you to issue a clarification regarding above mention Points.	<u>The clarification given by the concerned department of JIPMER is as follows:</u> The trajectory planning software is not the part of PET CT and it is required for accurate localization of site and minimizing the chance of false negatives due to human error. So automated software with dedicated monitor is required.