

05-05-2018

**Amendment No. 3**

**Sub: Amendment to the referred tender enquiry for Sch. 10 (Event No. 300002735) Physiograph Single Channel with time and event & Sch. 11 ((Event No. 300002736) Physiograph three channel with time and event**

**Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/AIIMS/01/18-19 dated: 10.04.2018**

The following changes are being incorporated in the above referred Tender Enquiry Document

**Section VII****Technical Specifications**

<b>Schedule No. 10 - Physiograph Single Channel with time and event (Rfx no. 300002735)</b>		
<b>Tender Page &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
Page 46	Student Physiograph should be single channel console with 9 speed (.5, 1, 2, 5, 10, 20, 25,30 & 50 mm/sec) chart drive, time & event markers and appropriate transducers and stimulator	Student Physiograph should be single channel console with 9 speed (.5, 1, 2, 5, 10, 20, 25,30 & 50 mm/sec) <b>TFT display</b> , time & event markers and appropriate transducers and stimulator
Page 46	Transducers: Pressure, volume, muscle activity/ force, Isotonic fine movement	Transducers: Pressure, Volume, Muscle activity / force, Isotonic fine movement, <b>EEG / EOG / EMG disc electrodes, GSR transducer, Pulse Transducer, ECG Electrodes, Respiration Belt, Temperature Electrode for Human Physiology</b>
Page 46	Ink bottle EP to EP lead Perpex pen Steel wire Motor Belt Chart paper Z- fold	<b>Deleted</b>

<b>Schedule No. 11 - Physiograph three channel with time and event (Rfx no. 300002736)</b>		
<b>Tender Page &amp; Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
Page 46	Console with time & Event channel and stimulator for human experiments	<b>TFT Console with time &amp; Event channel and <b>Inbuilt</b> stimulator for human experiments</b>
Page 47	Chart paper Z folds 250 folds 10 no.s	<b>Deleted</b>
Page 47	Extra pen with Cradles 01 no.s Ink ½ Ltr	<b>Deleted</b>
Page 47	<b>Added para:</b>	GSR Electrodes- 02 nos

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note:**

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**