

Amendment No. 4**07.07.2018****Sub: Amendment to the Bidding Document**

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/18/18-19 dated 25.05.2018 read with its Amendment No. 01, 02 & 03 dated 08.06.18, 19.06.18 and 26.06.18.

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION – I**NOTICE INVITING BIDS (NIB)**

Description	Existing*	Amended as.*
Last date and time of online submission of tender	09.07.2018 at 12:00 noon	17.07.2018 at 12:00 noon
Last date and time of physical submission of EMD, Tender processing Fee, any other document specified in the Bidding Document	09.07.2018 at 2:00 pm	17.07.2018 at 2:00 pm
Date of tender Opening	09.07.2018 at 2:30 pm	17.07.2018 at 2:30 pm

***Note: The above mentioned due date extensions are applicable only for the tendered item against Rfx no. 3000002858 i.e. 'CSSD System'**

SECTION - VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:**

**Item No. 1 (Rfx/ Event number 3000002858)
CSSD System**

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Amendment
1	Amendment No 2, Page 04, Point 17	Closed Sterilization Containers 300mm x 290mm x 110mm Should have thermo lock drainage, steam penetration valve and stainless steel/	Closed Sterilization Containers 300mm x 290mm x 110mm (+/- 10%) Should have thermo lock drainage, steam penetration valve and stainless steel/ Aluminium top. Having

		Aluminium top. Having biobarrier/ Permanent filter with life 4000- 5000 cycles	biobarrier/ Permanent filter with life 4000- 5000 cycles
2	Amendment No 2, Page 04, Point 18	Closed Sterilization Containers 300mm x 290mm x 140mm Should have thermo lock drainage, steam penetration valve and stainless steel/ Aluminium top. Having biobarrier/ Permanent filter with life 4000- 5000 cycles	Closed Sterilization Containers 300mm x 290mm x 140mm (+/-10%) Should have thermo lock drainage, steam penetration valve and stainless steel/ Aluminium top. Having biobarrier/ Permanent filter with life 4000- 5000 cycles
3	Amendment No 2, Page 04, Point 19	Closed Sterilization Containers 590mm x 280mm x 260mm Should have thermo lock drainage, steam penetration valve and stainless steel/ Aluminium top. Having biobarrier/ Permanent filter with life 4000- 5000 cycles	Closed Sterilization Containers 590mm x 280mm x 260mm (+/-10%) Should have thermo lock drainage, steam penetration valve and stainless steel/ Aluminium top. Having biobarrier/ Permanent filter with life 4000- 5000 cycles
4	Page 67,	Point 48- Free Standing basket rack (15 Baskets) for Sterile store -03 Nos.	Point 48.a- Free Standing basket rack (15 Baskets) for Sterile store Size (LxWxH) : 1850x480x2150 mm(Single) -02 Nos.
5	Page 67,		Added Point 48.b- Free Standing basket rack (15 Baskets) for Sterile store Size (LxWxH): 1850x800x2150 mm(Double) -02 Nos.

SECTION – X

Please refer the new version of Price bid format (i.e “3000002858_Ver1.2”) on HLL e-bid portal for RFX/ Event number 3000002858. The same need to be downloaded for submission of price offer.

All other contents of the Bidding Document including terms & conditions remain unaltered.