

24-01-2020

Amendment No. 02**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-IV/10/URLG/19-20 dated 08-01-2020 and Amendment No 1 dated 13-01-2020**

The following items are added to the tender enquiry document referred above:

**Section VII
Technical Specification**

Para	TENDER SPECIFICATION	READ AS
Item No 02 Pneumatic lithotripter		
10	In Let Pressure (Minimum) (To Control Unit) 4 Kg /cm ² & (Maximum) (To Control Unit) 6 Kg /cm ² 12. Out Let Pressure: 0 - 3 Kg /cm ²	In Let Pressure (Minimum) (To Control Unit) 4 Kg /cm ² & (Maximum) (To Control Unit) 6 Kg /cm ² , Out Let Pressure: 0 - 3 Kg /cm ²
Item No 03 Combined single generator for ultrasonic cutting & coagulation device with advanced Radio frequency energy		Item No 03 Urology workstation for Ultrasonic, Advanced Radiofrequency and Monopolar Energy
A.1	System should be a single/dual generator with single footswitch that provides ultrasonic energy and advanced RF energy technology for soft tissue dissection and vessel sealing	System should be a single/dual generator with single/dual foot switch that provides ultrasonic energy, monopolar cutting and coagulation, bipolar cutting and coagulation (Note: At all places RF to be read as RF/HF)
A.2	Should have the ability to select hand switch or footswitch activation or both for ultrasonic and advanced RF energy instruments and the ability to change selection during use can seal and transect vessel upto 7mm with system feedback mechanism for better patient safety.	Should have the ability to select hand switch or footswitch activation or both for instruments and the ability to change selection during use can seal and transect vessel upto 7mm with system feedback mechanism for better patient safety.
A.6	Should be able to power ultrasonic energy instruments with 35 KHz and above frequency	Should be able to power ultrasonic energy instruments with 35KHz and above frequency , Monopolar cutting 320-400W, Monopolar Coagulation:120W or more, Bipolar cutting 100W or more, Bipolar coagulation 120W or more. [Monopolar and Bipolar should be underwater compatible]
A.8	Should be able to power energy instruments with microprocessor controlled bipolar electrosurgical radiofrequency technology with a quasi-sinusoidal forced impedance output	Should be able to power energy instruments with microprocessor controlled bipolar electrosurgical radiofrequency technology with a quasi- sinusoidal/sinusoidal forced impedance output.

Para	TENDER SPECIFICATION	READ AS
A.9	Should be equipped with smart advanced RF energy technology to measure the tissue impedance and control the power delivery and that can simultaneously seal and transect vessels up to 7mm, large tissue pedicles and vascular bundles with system feedback.	Deleted
A.11	Should provide temperature controlled energy delivery which should maintain tissue temperature approximately at 100 deg C and hand instruments that provide tissue/vessel seal strength to withstand bursting pressure of 7 times the systolic pressure.	Should provide temperature controlled energy delivery which should maintain tissue temperature approximately at 100 deg C and hand instruments that provide tissue/vessel seal strength to withstand bursting pressure of 3 times the systolic pressure.
A. 12	Should be compatible with ultrasonic energy instruments which can simultaneously seal and transect vessels upto 7mm, large pedicles and vascular bundles with ultrasonic energy only with system feedback	Deleted
A. 16	Should provide class I protection against electric shock	Deleted
C.2	Adapters for ultrasonic and radiofrequency instruments	Adapters for ultrasonic and radiofrequency instruments if required
C.4	RF energy 5mm lap hand activated probe capable of sealing blood vessels upto 7mm in diameter, 35cm with 3mm curved tip	a. RF/HF energy 5mm lap hand activated probe capable of sealing blood vessels upto 7mm in diameter, 35cm with 3mm curved tip - 1 no Added para: b. RF/HF energy vessel sealing instrument for open procedures: 1 no
C.5	Ultrasonic 5mm lap hand activated curved coagulating shear with next generation adaptive tissue technology which gives regulated power delivery for improved thermal management and audible feedback tone for generator surgical efficiency, capable of sealing blood vessels upto 7mm in diameter using two hand activation buttons, proprietary non stick coated active blade to reduce tissue sticking, 36cm shaft length with ergonomic handle.	Shears for ultrasonic transducer: Laparoscopic: 10 nos Open: 10 nos

Para	TENDER SPECIFICATION	READ AS
		Added para : C.6: Monopolar accessories: Reusable forcep/pencil with reusable cable : 1no Electrode for monopolar pencil/forcep: 5 no Reusable patient plate with reusable cable: 1 no Reusable bipolar forceps with reusable cable: 1
Item No 04 Urodynamic System		
2	Radiolucent electrohydraulic chair / table capable of supine and sitting position having perineal cut with minimum 300 mm longitudinal slide for C-ARM access. Should have swivel arm mechanism for Uroflow transducer.	2. Video Urodynamic Table/Chair a. Radiolucent motorised chair providing eight positioning functions capable of supine, sitting and standing position b. It should allow patient to remain in one place for preparation, filling and voiding phases of the urodynamic test c. It should eliminate the risk of catheters falling out or getting misplaced during transfer or ambulation d. Chair should be protected against water incursion e. There should be double wheel rolling casters offering unencumbered maneuverability f. There should be side mounted cantilevered design offering antero/posterior, lateral and oblique views without obstruction and to provide easier access to, and visualization of, the perineal area. g. There should be radiolucent carbon fiber seat and back for head-to-toe imaging. h. Weight capacity in standing position should be 250kg.
4	Should display atleast 8 display channel	Should display atleast 6 channels at a time
12	All in one PC computer Windows 10 or more, i7 processor, at least 4 GB RAM, 1 TB HDD with colour laser printer, should have facility	All in one PC computer Windows 10 or more, i5 processor, at least 4 GB RAM, 1 TB HDD with colour laser printer, should have facility
14	Should be supplied 19 inch flat screen monitor for display. Monitor should have facility of RGB output so that it can be projected on screen in conferences	Should be supplied 19" or more flat screen monitor for display. Monitor should have facility of RGB output so that it can be projected on screen in conferences
16	Should have Bluetooth data transfer or equivalent facility for CMG, Pressure flow and other Urodynamic studies	Should have wireless data transfer facility for CMG, Pressure flow and other Urodynamic studies.

Para	TENDER SPECIFICATION	READ AS
25	Reusable pressure transducers - 3 Nos (to be covered under both warranty & CMC)	Reusable pressure transducers - 3 pcs (to be covered under both warranty & CMC) ; Domes: 50 pieces
		Added Para: Price of accessories and consumables should be quoted separately
BOQ.7	Reusable pressure transducers - 3 Nos	Reusable pressure transducers - 3 Nos Domes: 50 pieces
Item No 05 Uroflowmetry system		
2	The Uroflowmetry sensor should be wired/Wireless to the main unit. The sensor should also be operable by wireless Bluetooth mode	The Uroflowmetry sensor should be wireless to the main unit.
3	Should have Database Software and Uroflow Software which includes extensive report printed by PC printer, possibility to add investigation comments and real time on line view of the investigation. Should have the facility of wireless transfer of data using Bluetooth Technology with automatic start. Automatic Investigation and Analysis.	Should have Database Software and Uroflow Software which includes extensive report printed by PC printer, possibility to add investigation comments and real time on line view of the investigation. Should have the facility of wireless transfer of data with automatic start. Automatic Investigation and Analysis.
8	Beakers for Uroflow -5 Nos, Rechargeable Batteries 2set for each (Uroflo & Blader scan), Charger for Rechargeable Batteries 1set for each (Uroflo & Blader scan). Should be supplied as standard	Beakers for Uroflow -5 Nos, 1 set Rechargeable Batteries for Bladder scanner, 1 set batteries for Uroflow. Charger for Rechargeable Batteries 1set (if applicable) Should be supplied as standard
	BOQ	BOQ
3	Rechargeable Batteries (Uroflo & Blader scan) 2 sets each	Rechargeable Batteries (Bladder scan); Batteries for Uroflow 1 set each
4	Charger for Rechargeable Batteries (Uroflo & Blader scan) 1 sets each	Charger for Rechargeable Batteries (Blader scan) - 1 set (If applicable)
Item No 06 - Holmium Laser 100W		
7	It should be supplied with twin/double foot pedal	It should be supplied with single/double foot pedal
8	It should have repetition rate of 5-50Hz or more	It should have repetition rate of 5-30Hz or more
16.ix	Ceramic Scissors 1	Ceramic Scissors / Ceramic Cutter - 1 no
BOQ-10	Ceramic Scissors 1 No	Ceramic Scissors / Ceramic Cutter - 1 no

Para	TENDER SPECIFICATION	READ AS
Item No 07 Tissue Morcellator		
B.1	One control box	Suction & Shaver system incorporated through Control Box.
B.3.a	Outer Blade –Outer dia 0.5cm or less, length 39cm or more	Outer Blade –Outer dia 0.5cm or less, length 35cm or more
B.3.b	Inner Blade –Outer dia 0.39cm or less, length 40cm or more	Inner blade should be compatible to outer blade
16	Single Chamber Imported Tissue Evacuator – 1	Single Chamber Tissue Evacuator – 1
12	LASER bridge – 2	LASER bridge/ Laser guide rod – 2
		Added para: 20. All accessories required for smooth functioning of system to be supplied as standard
		Added para: 21. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 22. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: 23. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 24. Price of consumables/accessories should be quoted separately in the price bid
Item No 08 Flexible Cysto Nephroscope		
4.2	All consumables required for installation and standardization of system to be given free of cost	All accessories required for installation and standardization of system to be given free of cost
		Added para: 8. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided

Para	TENDER SPECIFICATION	READ AS
		Added para: 9.If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 10.Price of consumables/accessories should be quoted separately in the price bid
Item No 09 Turp, Cystoscope & Optical Urethrotome		
4	Flexible Grasping forceps Compatible Grasping forceps 33cm or more to be provided to fit the purpose - 4 nos.	Flexible Grasping forceps Compatible Grasping forceps 31cm or more to be provided to fit the purpose - 4 nos.
5	Flexible cup biopsy forcep Compatible cup biopsy forceps 33cm or more to be provided to fit the purpose - 2 no.	Flexible cup biopsy forcep Compatible cup biopsy forceps 31cm or more to be provided to fit the purpose - 2 no.
8	Urethrotome sheath with 1 working element	Urethrotome sheath with 1 working element
	Straight cold knife 5 nos.	Straight cold knife 30 nos.
10	Working Element set	
	Working element set passive type compatible with offered resectoscope sheath with standard accessories like, Collins knife, HF cord, Protection tube, cutting loop to be provided - 1 set comprising of following	One Working element passive type (compatible with offered resectoscope sheath) with standard accessories like, Collins knife, HF cord, Protection tube, cutting loop to be provided comprising of following
	Collins knife - 5 nos.	Collins knife - 30 nos.
		Added para: 17. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 18. Price of consumables/accessories should be quoted separately in the price bid
Item No 10 Laparoscopic Set		
4.a	Atraumatic bowel Grasping Forceps, both jaw opening 5 x 330 mm, ratchet, non locking 2 no.	Atraumatic bowel Grasping Forceps, both jaw opening 5 mm, ratchet, non locking 2 no.
4.b	Insert only of atraumatic bowel Grasping Forceps, both jaw opening 5 x 330 mm, ratchet: 2 no.	Insert only of atraumatic bowel Grasping Forceps, both jaw opening 5 mm, ratchet: 2 no.

Para	TENDER SPECIFICATION	READ AS
4.c	Maryland Forceps, 5 x 330 mm, Monopolar 2 no.	Maryland Forceps, 5 mm, Monopolar 2 no.
4.d	Metzenbaum Scissors, 5 x 330 mm, Monopolar 2 no	Metzenbaum Scissors, 5 mm, Monopolar 2 no
4.e	Insert only of Metzenbaum Scissors, 5 x 330 mm, Monopolar 2 no	Insert only of Metzenbaum Scissors, 5 mm, Monopolar 2 no
4.f	Hook Scissors, 5 x 330 mm, Monopolar 2 no	Hook Scissors, 5 mm, Monopolar 2 no
4.h	Straight Needle Holder, 5 x 330 mm 2 no	Straight Needle Holder, 5 mm 2 no
5.a	Trocar & Cannula(metallic) 5-6mm: 5no 11mm: 5no	Trocar & Cannula(metallic) 5-6mm: 5no, 10-11mm: 5no
6.f	The system should have digital zoom and integrated optical zoom with auto focus to enhance the quality of image size & cross speciality standardization of the camera system, regardless of the telescope used	The system should have digital zoom and optical zoom to enhance the quality of image size & cross speciality standardization of the camera system, regardless of the telescope used.
6.k	Lens F14-30mm \pm 10 %	Lens F16 -30mm \pm 10 %
6.m	The Camera should be capable of directly recording full HD quality photos (1920x1080p) and full HD quality videos(1920x1080p) into an internal memory of 300 GB or more / into an external hard drive of 300GB or more. If not, a medical grade HD recorder capable of directly recording full HD quality photos (1920x1080p) and full HD quality videos(1920x1080p) with internal memory of 300GB or more /external hard drive of 300 GB or more must be supplied	Deleted
7.b	26" full HD medical grade monitor in 16:9/10 HDTV format, LED/LCD display	26" or more full HD medical grade monitor in 16:9/10 HDTV format, LED /LED backlit display
7.d	SDI/HD-SDI	DVI /HD-SDI
7.f	TFT/LCD/LED screen stand/Fixtures for connecting to Pendant System/Ceiling Light Arm	Deleted
8.b	Adjustable flow rate of 30 litres per minute or more and pressure range adjustable between 0 to 30 mm Hg	Adjustable flow rate of 45 litres per minute or more and pressure range adjustable between 0 to 30 mm Hg

Para	TENDER SPECIFICATION	READ AS
8.h	Should have smoke evacuation available in the system or separate smoke evacuation system should be offered	Deleted
8.j	Should include pin-index connection to small/big gas cylinder with regulator, high pressure hose, mains cord, silicone autoclavable tubing set, universal wrench and gas filter	Should include pin-index connection to small/big gas cylinder with regulator, high pressure hose, accessory to connect with central pipeline , mains cord, silicone autoclavable tubing set, universal wrench and gas filter
9	LIGHT SOURCE (Xenon 300) with TWO Spare Bulbs	LED LIGHT SOURCE
9.a	Xenon cold light fountain with 300 watts xenon lamp	LED light source equivalent to 300 W Xenon having lamp life time of atleast 5000 hours
9.b	Colour temperature of at least 5800 °K	Deleted
9.c	Manual and automatic adjustment of light intensity	Deleted
9.d	Brightness control to be regulated manually or automatically via the output signal of a video camera	Deleted
9.e	Lamp life 500 hrs or more	Deleted
9.f	Display of lamp life/Bulb usage meter warning light	Deleted
10.f	Accessories should include silicone suction tubing set with reusable pressure domes, bacterial filter and suction bottles with cap (minimum 5 ltrs.)	Accessories should include silicone suction tubing set with reusable pressure domes, bacterial filter and suction bottles with cap (minimum 3 ltrs.)
11.d	Preferable adjustable arm for fixation to either side for fixing the TFT monitor	Adjustable arm for fixation to either side for fixing the monitor
	12. The complete system must be European CE (with 4 digit notified body no.) or USFDA approved/BIS approved	13. The complete system must be European CE (with 4 digit notified body no.) or USFDA approved/BIS approved
13	Suitable online UPS 2kVA with 30min battery backup for the complete system should be supplied	Deleted
		Added para: 14. A separate medical grade recording system with resolution 1920 x 1080p with 2TB inbuilt memory should be provided

Para	TENDER SPECIFICATION	READ AS
		Added para: 15. The Laparoscopic system should have smoke evacuation available in the system or separate smoke evacuation system should be offered
		Added para: 16. The telescopes, Camera, Monitor, Insufflator, Light Source and all their subparts must be from the same OEM
		Added para: 17. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 18. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch, a temporary standby machine of same specification must be provided
		Added para: 19. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		20. Price of consumables/accessories should be quoted separately in the price bid
		Added Para BOQ
1		Telescope 10mm 30 degree: 2 No.
2		Telescope 5 -5.5mm 30 degree: 1 No
3		Telescope 10mm 0 degree: 1 No
4		Laparoscopic Hand Instrument set as per specification (para 4 & 5)- 1 set
5		Camera control unit, camera & light source - 1 set
6		Light guide cable - 2 nos
7		Monitor - 1 no
8		CO2 Electronic INSUFFLATOR - 2 nos
9		Suction irrigation unit - 1 no
10		Video cart (same oem) - 1 no
11		5-6 kg cylinder - 2 nos

Para	TENDER SPECIFICATION	READ AS
12		10-12 kg cylinder - 2 nos
13		Medical grade recording system - 1 no
Item No 11 PCNL set		
		Added para: F. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: G. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: H. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		I. Price of consumables/accessories should be quoted separately in the price bid
Item No 12 Uretero renolescope - Pediatric		
4.5	All adapters required for smooth functioning of system to be supplied as standard	All accessories required for smooth functioning of system to be supplied as standard
		Added para: 6. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 7. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: 8. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 9. Price of consumables/accessories should be quoted separately in the price bid

Para	TENDER SPECIFICATION	READ AS
Item No 13 - Uretero renoscope Adult		
3.2	Long Arm (Working Length – 430 mm or more) Autoclave with offset Eyepiece, Distal sheath tip 7.5-9Fr. Atraumatic Viewing angle 5 -10 degree with laterally placed eyepiece. - 1 no. Channel size 5 fr or more	Long Arm (Working Length – 430 mm or more) Autoclave with offset Eyepiece, Distal sheath tip 7.5-9Fr. Atraumatic Viewing angle 5 -12 degree with laterally placed eyepiece. - 1 no. Channel size 5 fr or more
4.6	All adapters required for smooth functioning of system to be supplied as standard	All accessories required for smooth functioning of system to be supplied as standard
		Added para: 6.Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 7.There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: 8. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 9. Price of consumables/accessories should be quoted separately in the price bid
Item No. 14 - Cystoscope & Resectoscope – Peadiatric		
D.a	2. Cold knife set of 2.7mm set of 6-1 set	Deleted
		Added para: 5.Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 6.There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: 7.If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same

Para	TENDER SPECIFICATION	READ AS
		8.Price of consumables/accessories should be quoted separately in the price bid
		Added para: 9.All accessories required for smooth functioning of system to be supplied as standard
Item No 15 - Camera system for Endourology		
1.f	The system should have digital zoom and integrated optical zoom with auto focus to enhance the quality of image size & cross speciality standardization of the camera system, regardless of the telescope used.	The system should have digital zoom and optical zoom to enhance the quality of image size & cross speciality standardization of the camera system, regardless of the telescope used.
1.k	Lens F14-30mm ± 10 %	Lens F16 -30mm ± 10 %
1.m	The Camera should be capable of directly recording full HD quality photos(1920x1080p) and full HD quality videos(1920x1080p) into an internal memory of 300 GB or more / into an external hard drive of 300GB or more. If not, a medical grade HD recorder capable of directly recording full HD quality photos(1920x1080p)and full HD quality videos(1920x1080p) with internal memory of 300GB or more /external hard drive of 300 GB or more must be supplied.	Deleted
2.b	26" full HD medical grade monitor in 16:9/10 HDTV format, LED/LCD display	26" or more full HD medical grade monitor in 16:9/10 HDTV format, LED /LED backlit display
2.d	SDI/HD-SDI input	DVI /HD-SDI
2.f	TFT/LCD/LED screen stand/Fixtures for connecting to Pendant System/Ceiling Light Arm	Deleted
3	LIGHT SOURCE (Xenon 300) with TWO Spare Bulbs	LED LIGHT SOURCE
3.a	Xenon cold light fountain with 300 watts xenon lamp	LED light source equivalent to 300 W Xenon having lamp life time of atleast 5000 hours
3.b	Colour temperature of at least 5800 °K	Deleted
3.c	Manual and automatic adjustment of light intensity	Deleted

Para	TENDER SPECIFICATION	READ AS
3.d	Brightness control to be regulated manually and /or automatically via the output signal of a video camera	Deleted
3.e	Lamp life 500 hrs or more	Deleted
3.f	Display of lamp life/Bulb usage meter warning light	Deleted
4.d	Preferable adjustable arm for fixation to either side for fixing the TFT monitor	Adjustable arm for fixation to either side for fixing the monitor
		Added para: 6. A separate medical grade recording system with resolution 1920 x 1080p with 2TB inbuilt memory should be provided
		Added para: 7. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 8. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided
		Added para: 9. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 10. Price of consumables/accessories should be quoted separately in the price bid
Item No 16 - Flexible Uretroscope		
		Added para: 15. Declaration from principal manufacturer that quoted items are brand new and not refurbished
		Added para: 16. There will be 95% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. If non functioning period exceed 2 weeks at a stretch , a temporary standby machine of same specification must be provided

Para	TENDER SPECIFICATION	READ AS
		Added para: 17. If accessories are not supplied from the OEM, declaration should be submitted from OEM that they do not manufacture the same
		Added para: 18. Price of consumables/accessories should be quoted separately in the price bid
		Added para: 19. All accessories required for smooth functioning of system to be supplied as standard
Item No 17 OT Table - Urology		
A.3	Removable and Interchangeable head/ leg sections to suit different applications	Removable and Interchangeable head and leg sections to suit different applications with buttons on corded handset to detect orientation
A.15	Table should have a mechanism/sensor to prevent/detect collision	Should have anti collision system to detect/prevent accident or breakage if any object comes under the OT table top
C.13	Maximum patient weight : 375kg or more at centered column and 220 kg or more in all articulations	Maximum patient weight: 250 kgs or more in all articulations
D.9	Clamp for locking X Ray cassette -1	Deleted
D.10	Lithotomy leg holders "Geopel type" (adult and paediatric)-1set each	Lithotomy leg holders "Geopel type" -1set
		Added para: D.13 Extender for Urological surgeries with perineal cut with compatible drain tray - 1 no
BOQ.11	Lithotomy leg holders "Geopel type" (adult and paediatric) 1 Set each	Lithotomy leg holders "Geopel type" -1set
BOQ.14		Added para: Extender for Urological surgeries with perineal cut with compatible drain tray - 1 no
Item No 18 Ultrasound for Urology Applications Ultrasound for Urology Applications		
25	Should be Upgradable to MR fusion , T/R and T/P.	Should be Upgradable to MR fusion (T/R and T/P) by same OEM

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. **Techno Commercial Sheet v1.0 for this tender is being uploaded on CPPP. The same need to be downloaded for submission.**

- ii. Price Bid format v1.0 for this tender is being uploaded on CPPP. The same need to be downloaded for submission.**
- iii. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**