

**Amendment No. 15**

**Date: 18.01.2016**

**Sub: Amendment to the Tender Enquiry Document**

**Ref: Tender Ref No.: HLL/PCD/PMSSY-II/NAGPUR/10/15-16 dated 07/10/2015 and subsequent amendments**

**TECHNICAL AMENDMENT**

**Schedule No. 1**

**DEXA Scan**

- 1. Existing Para 5:** Para 5: Scan Time: A/P Spine  $\leq$  30 sec; Femer  $\leq$  30 sec  
**Read as:** Scan Time: A/P Spine  $\leq$  **60 sec**; Femer  $\leq$  **60 sec**
- 2. Added Para Under Para 13:** 3 D software for volumetric bone geometry , cortical bone thickness and volumetric bone mineral density for anticipating fracture risk
- 3. Added Para:** Suitable UPS with 30min backup for the entire system
- 4. Added Para under Turnkey:** Earthing of the Equipment will be the responsibility of the vendor

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**