

Clarification Regarding Technical Specifications & Turnkey

Date: 14.10.2014

Subject: Clarification regarding technical specifications and turnkey

- Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/11/13-14 dated 19.12.2013.
(ii) Amendment No.1 dtd.21.01.2014.
(iii) Amendment No.2 dtd. 10.02.2014.
(iv) Amendment No.3 dtd. 27.02.2014.
(v) Amendment No.4 dtd. 14.03.2014.
(vi) Amendment No.5 dtd. 27.03.2014.
(vii) Amendment No.6 dtd. 16.04.2014.
(viii) Amendment No.7 dtd. 09.05.2014.
(ix) Amendment No.8 dtd. 17.05.2014.
(x) Amendment No.9 dtd. 19.06.2014.
(xi) Amendment No.10 dtd. 21.07.2014
(xii) Amendment No.11 dtd. 16.08.2014
(xiii) Amendment No.12 dtd. 02.09.2014
(xiv) Amendment No.13 dtd. 15.09.2014
(xv) Amendment No. 14 dtd. 19.09.2014
(xvi) Amendment No. 15 dtd. 13.10.2014

The following clarifications regarding technical specifications are made further to the TED and Amendments as referred.

Section VII Technical Specifications

Item No: 13 CT Simulator System

1) For:

The Turnkey Scope of Work –MULTISLICE CT SIMULATOR

1.The bidders should inspect the proposed site offered by the Consignee Institute in which the CT system has to be installed and they are required to submit the plan for the complete CT Scan Centre on a turnkey basis. The scope of work includes complete Civil work, Electrical, Plumbing, Furnishing, Air-conditioning and Fire fighting for the construction of CT Scan Centre.

Read as:

- 1. Principal Vendor (Supplier of System A) will be responsible for quoting for turnkey for both the systems even though the System B(CT Simulator System is not quoted by them) is not supplied by them.** The Principal Vendor (Supplier of System A) should inspect the proposed site offered by the Consignee Institute in which the CT system has to be installed and they are required to submit the plan for the complete CT Scan Centre on a turnkey basis. The scope of work includes complete Civil work, Electrical, Plumbing, Furnishing, Air-conditioning and Fire fighting for the construction of CT Scan Centre.

2) For:

3. The cost of Turnkey for the area of 140 sq.ft and Air-conditioning of Tonnage 18 TR will be considered for Ranking / Evaluation purpose.

Read as:

3. The cost of Turnkey for the area of **140 sq.m** and Air-conditioning of Tonnage 18 TR will be considered for Ranking / Evaluation purpose.

All other terms and conditions of the tender enquiry remain unaltered.