

Amendment No.1 to Tender Enquiry Document

Sub: Amendment to Tender Enquiry Document.

Ref: NIT No. HLL/PCD/NIB-19/15-16 Dated 21.12.2015

The following amendments are issued with respect to the above tender Enquiry Document

Existing:

SECTION - VII Technical Specifications

Schedule No. 1 Wireless Temperature Monitoring Data Logger

The Detailed Technical Specifications are as given below:

Page No.	General Characteristics	
44	Signal strength and coverage	Wireless coverage range should be minimum 250 meters in the open area. Data logging system should be capable of connecting 200+ data loggers to single receiver/computer
44	Parameters required	Temperature range from (-) 40°C to (+) 80 °C

Read As:

SECTION - VII Technical Specifications

Schedule No. 1 Wireless Temperature Monitoring Data Logger

The Detailed Technical Specifications are as given below:

Page No.	General Characteristics	
44	Signal strength and coverage	Wireless coverage range should be minimum 100 meters in the open area. Data logging system should be capable of connecting 100+ data loggers to single receiver/computer
44	Parameters required	Temperature range from (-) 40°C to (+) 80 °C Humidity range 0 to 100%RH

SECTION - IX
Qualification Criteria

Existing:

02. (a) The Manufacturer should have supplied in last Five years from the date of Tender Opening, atleast 50% of the quoted quantity of the similar item performing similar function which has been meeting major specification parameters which is functioning satisfactorily any where. (*“Similar equipment”, means the “Chassis” or fabricators of the truck/bus body*”).

Read As:

02. (a) The Manufacturer should have supplied in last Five years from the date of Tender Opening, atleast 50% of the quoted quantity of the similar item performing similar function which has been meeting major specification parameters which is functioning satisfactorily any where.

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.