

Amendment No. 3

Date: 24/02/2014

Subject: Amendment to the tender Enquiry Document

Ref: Tender Enquiry No.: HLL/PCD/PMSSY-II/Nagpur/01/13-14 dated 29/12/2013 read with Amendment no. 1 dated 28/01/2014 and Amendment no. 2 dated 15/02/2014

The following changes have been incorporated in the referred tender .

Section – VII Technical Specifications

Item No: 9

ICCU Beds Motorized

Existing

3.6 - Should have step less pneumatic adjustment for Trendlenburg (15° approx), anti-trendlenburg (15° approx)

Read As

3.6 - Should have step less pneumatic adjustment for Trendlenburg (10° approx), anti-trendlenburg (13° approx)

Existing

6.2 - Battery backup for one hour of operations.

Read as

6.2 - Deleted

All other contents of the tender enquiry read with its earlier amendment including terms & conditions remain unaltered.