

TENDER ENQUIRY NOTICE

HLL/ID/C & M/ KKD-Med Equip-Phase II/2014-15 /523

05.06.2015

Name of Work : Supply of Medical Equipment for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II.

Limited tenders are invited on behalf of The Principal, Govt: Medical College, Kozhikode for the Supply of Medical Equipment for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II. .The specification for the items is enclosed herewith.

Address to which the offers are to be submitted:

Deputy Vice President (Technical)

HLL Lifecare Limited

"ADARSH", TC - 6/1718 (1)

Vettamukku, Thirumala P O

Thiruvananthapuram - 695 006

Ph : 0471 - 2365872/73

Fax: 0471 - 2368144

Submission of Offer

The offer shall be submitted in single bid system. *The offer should be submitted in a sealed envelope superscribing "Supply of Medical Equipment for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II.*

Last date and time of Receipt of offers :15.06.2015 11.00 am.

Date and time of Opening of Offers :15.06.2015 11.30 am

Payment terms

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

75 % payment of the contract price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Consignee Receipt Certificate as per in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by HLL.
- (v) Certificate of origin.

b) On Acceptance:

Balance 25 % payment would be made after satisfactory completion against 'Final Acceptance Certificate' of goods/equipments to be issued by the consignees and endorsed by HLL subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

The address to which the items have to be supplied is as follows:

The Principal, Govt: Medical College ,Kozhikode
Kozhikode -673008

Delivery period

The Equipments shall be supplied within 1 month from the date of supply order.

**The supplier should give CMC for the quoted item for 5 years after 2 years warranty period.
Cost of CMC will not be considered for price comparison purpose.**

1	2	3	4	5					6
Sch edu le No.	BRIEF DESCRIPTI ON OF GOODS	QUANT ITY. (Nos.)	Basic price of the equipment excluding tax	CMC Cost for Each Unit year wise*.					Total CMC Contract Cost for 5 Years
				1 st	2 nd	3 ^r d	4 th	5 th	
				a	b	c	d	e	
1									
2									
3									

The cost of CMC quoted shall be exclusive of service tax . The service tax shall be extra as applicable.

Sl No:	Equipment Name	Qty	EMD in Rs
1.1	WPS Electronic Tapping test	1	400
1.2	Modified WISCONSIN card Sorting test	1	500

1.3	STROOP colour and word test (Adults)	1	300
2	Micromotorfor ENT (Indigenous)	2	4000
3	Electrocautery for microscope	1	5000

Other Conditions

1. Warranty should be for minimum 2 years from the date of final acceptance of the equipments and the Warranty has to be given in the name of client.

Performance Security

2. Within Fifteen (15) days from date of the issue of purchase order the successful bidder, shall furnish performance security for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.
3. Demo of equipment and sample of instruments should be arranged when asked for
4. Earnest Money as shown in the above tabular column has to be submitted along with the tender in the form of a Demand Draft/ Fixed Deposit Receipt (FDR) of a scheduled bank issued in favour of HLL Lifecare Limited, Thiruvananthapuram which should be placed in a separate sealed cover marked "Earnest Money" shall be submitted along with the tender.
5. The rate quoted should be inclusive of all taxes, duties, transport, loading and unloading up to consignee site.
6. The tenderer shall submit the compliance for technical specifications. The make and Model should also be attached along with the bid.
7. The original Literature or the relevant part of the user/service manual should be attached as proof.
8. The offer shall be valid for 3 months.
9. The invoice/ bills shall be raised in the name of "The Principal, Govt: Medical College, Kozhikode a/c HLL Lifecare Ltd" and the same shall be submitted to the HLL Engineer in charge.
10. TIN no. of HLL should not be mentioned in any of the documents as HLL is acting on behalf of the client in the capacity of consultant alone.
11. The completed works shall be handed over to the client after due verification by HLL Engineer in Charge.
- 12. The bidders should seal and sign the Tender Enquiry Notice and should be submitted along with the offer.**

Deputy Vice President (Tech)

HLL Lifecare Limited

1	MATERIAL FOR PSYCHOMETRY
1.1	WPS ELECTRONIC TAPPING TEST
	Should be an accurate, easy-to-use finger or foot tapping measure for neuropsychological evaluation
	Compact electronic tapper should automatically start a 10-second timer as soon as the first tap is made. The digital display should show a dash (--) until 10 seconds are up--at which point it shows the number of taps made during that interval. Additional taps, made after the 10-second period should be elapsed and should not be recorded.
	Should operate with 9 V battery
	Should be supplied with plastic casing
1.2	MODIFIED WISCONSIN CARD SORTING TEST
	The M-WCST should be a modification of the original Wisconsin Card Sorting Test that eliminates all cards from the original 128-card deck that share more than one attribute with a stimulus card (e.g., each response card whose stimuli match a stimulus card in both color and form was eliminated). This results in a 48-card deck, which is used along with four stimulus cards.
	This modifications should enable impaired and elderly adults to more easily understand the directions, reducing frustration.
	Four scores are calculated for the M-WCST: Number of Categories Correct, Number of Perseverative Errors, Number of Total Errors, and Percent of Perseverative Errors. In addition, an Executive Function Composite score can be calculated.
1.3	STROOP COLOUR AND WORD TEST (ADULTS)
	The Stroop is based on the observation that individuals can read words much faster than they can identify and name colors. The cognitive dimension tapped by the Stroop is associated with cognitive flexibility, resistance to interference from outside stimuli, creativity, and psychopathology--all of which influence the individual's ability to cope with cognitive stress and process complex input.
	The test should feature a three-page test booklet. On the first page, the words "RED," "GREEN," and "BLUE," are printed in black ink and repeated randomly in columns. On the second page, the item "XXXX" appears repeatedly in columns, printed in red, green, or blue ink. On the third page (referred to as the interference page), the words "RED," "GREEN," and "BLUE" are printed in red, green, or blue ink--but in no case do the words and the colors in which they are printed match.

	For example, the word "BLUE" appears in either red or green ink. The subject's task is to look at each page and move down the columns, reading words or naming the ink colors as quickly as possible, within a given time limit
	The test yields three scores, based on the number of items completed on each of the three stimulus sheets. In addition, you can calculate an interference score, which is useful in determining the individual's cognitive flexibility, creativity, and reaction to cognitive stress
2	MICROMOTOR FOR ENT
2.1.1	Technical Specification
	Should have a brushless micro motor with variable speed from 10,000 to 35,000rpm.
	Should have at least four speed control modes.
	Should have an automatic cooling fan for continuous high speed performance.
	Should have an on/off foot control.
	Should have provision for irrigation.
	Should be supplied with straight hand piece for micro motor and contra angle hand piece for micro motor.
	Should be supplied with the following Burs set.
	• Tungsten Carbide burs - 3 nos. (Assorted sizes)
	• Diamond burs - 3 nos. (Assorted sizes)
	• Steel cutting burs - 5 nos. (Assorted sizes)
2.1.2	Power Supply
	Power input to be 220-240VAC, 50Hz fitted with Indian plug
2.1.3	Standards, Safety and Training
	Manufacturer should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
2.1.4	Documentation
	User/Technical/Maintenance manual to be supplied in English
	Certificate of calibration and inspection from factory
3	ELECTROCAUTERY FOR DERMATOLOGY
3.1	Technical Specification
	Removal of Warts, Desiccation of sebaceous gland disorders, Epilation, Destruction of skin cancers such as basal cell carcinoma
	Should be microprocessor controlled device with LCD screen
	Dual frequency: 1.7 MHz , 4 MHz
	Power Output: 120 watts

	Finger Switch as well as foot switch control
	Should have atleast two output wave forms
	Should have atleast five mode of operation including dessication, and fulguration
	Should have bipolar and monopolar modes
	The electrodes and patient plate should be sterilizable
	Power supply: 230V,50 Hz
	Protection class:Class 1,Type BF
3.2	Acessories,Spares and Consumables
	System as per specs-01
	Patient plate -02 nos
	Bipolar electrodes with cable-04 nos
	Monopolar electrodes with cable-04 nos
	Power cord -01
3.3	Power Supply
	Power input to be 220-240VAC, 50Hz fitted with Indian plug
3.4	Standards, Safety and Training
	Manufacturer should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
	The quoted model should have US FDA/ European CE/ BIS certificate and copy of the same should be enclosed along with the technical bid.
3.5	Documentation
	User/Technical/Maintenance manual to be supplied in English
	Certificate of calibration and inspection from factory