

AMENDMENT No.4

No: HLL/PCD/UIP/GMSD-1/2012-13/

Date: 12.06.2013

Tender Enquiry No.: HLL/PCD/UIP/GMSD-1/2012-13 dated 19.02.2013

NOTICE INVITING TENDERS (NIT)

Technical specifications: Equipment specifications for Wireless data loggers Schedule-1

8.1 Clause:Operational details: Existing:

Internal memory of minimum 10000 data points for backup in event of system failure.

Read As:

Internal memory of 3,000 minimum data points for backup in event of system failure.

8.2 Clause: Operational details: Existing:

System should be capable of capturing data from 1 minute to 90 minutes

Read as:

System should be capable of capturing data from 1 minute to 24 hours

8.3 Clause: Alert System for deviation from High and Low Threshold Temp: Existing:

Alerts are to be generated through phone, email, etc. Possibility of connecting to suitable external hooter/buzzer/light.

Read as:

Alerts are to be generated through phone, fax, email, print out etc. Possibility of connecting to suitable external hooter/buzzer/light.

Other technical specifications will remain same.

-END-