

## **Amendment No. 1**

**Date: 08/07/2014**

**Subject: Amendment to the Tender Enquiry Document**

**Ref: NIT No: HLL/PCD/GNCTD/07/MAMC/14-15 dated 02.06.2014**

The following changes have been incorporated in the referred tender .

### **Section – VII Technical Specifications**

#### **Item No. 7**

**Existing Item Name: Desktop Discreet Random Access Analyzer**

**Amended as: Discreet Random Access Analyzer**

- 1. Existing Para:** Probe: Ceramic coated with capacitive liquid level sensor for sample and reagent both.

**Read as:** Probe: **Ceramic/ Teflon** coated with capacitive liquid level sensor for sample and reagent both.

#### **Item No. 8**

#### **Liquid Based Cytology Machine**

- 1. Existing Para 9:** System should have provision for future up gradation for including HPV hybrid capture assay, automated screening using image analysis software or integrated image analysis system & computer assisted imaging system.

**Read as:** System should be **compatible** for including HPV hybrid capture assay, automated screening using image analysis software or integrated image analysis system & computer assisted imaging system.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**