

AMENDMENT NO.2 Dated 02.03.2024

Reference NIT No.: HLL/CHO/SD/MKTG/2023-24/06 DT. 19-02-2024

The following amendment has been incorporated in the above referred NIT No. **HLL/CHO/SD/MKTG/2023-24/06 DT. 19-02-2024** the Supply of Diagnostic KITS under rate contract for 24 months

1. Notice Inviting tender (NIT)

Bid submission and Opening Date and Time

FOR

Last date and time for online submission of online bids	02.03.2024 (15: 00 Hrs.)
Date and time of opening of e-tender	04.03.2024 (15: 00 Hrs.)

MAY BE READ AS

Last date and time for online submission of online bids	11.03.2024 (15: 00 Hrs.)
Date and time of opening of e-tender	12.03.2024 (15: 00 Hrs.)

2. MAY BE READ AS

S. No.	Reference / Tender page number	Particulars as per tender document / Amendment	Amendment and may read as under
1	Certificates (Page no: 11)	03 years Market Standing Certificate	CA certified 03 years market standing certificate along with valid product license issued by FDA.
2	Products details – Specifications Page no. 11	Products details – Specifications	Approved Specifications and product packaging details provided at Annexure -1

Annexure – I: Approved Specifications For Pregnancy Rapid cards (in single pack size in wallet) Immunochromatography Lateral Flow For qualitative detection of pregnancy in urine hCG (Card / Cassette format)

Test Specimen	Women Urine
Testing Principle	Single step self-performing sandwiched immunoassay using colloidal gold and anti hCG antibodies in lateral flow Immunochromatography format
Time to Result	Less than 5 minutes
Sensitivity	Not less than 25 Milli IU per ml of urine hCG
Specificity	100% (No reactivity with other gonadotropin hormones LH and FSH)
Built in control	Should have built in control for correctness of test procedure
Nitrocellulose Paper	Nitrocellulose paper coated with anti hCG antibodies for test band and appropriate reagents for control band
Pouch	Triple layered laminated pouch having aluminium foil in the middle layer
Silica gel	Each test pack should have moisture indicating silica gel
Dropper	Each test pouch should have disposable dropper for urine specimen addition

Shelf Life	24 months from the date of manufacturing
Storage	The kit should be stable at 2 – 30°C (Room Temperature)
Pack Size and product packaging	<ul style="list-style-type: none">- Each sealed pouch contains a test card, silica gel desiccant, dropper for addition of sample 02 drops for testing as per product insert- Each Wallet / Mono carton contains a sealed pouch and product insert- Each Inner Carton contains 32 wallets- Each Shipper Box contains 24 Inner Cartons

Vice President (Sourcing)