

Amendment No. 2**19-01-2018****Sub: Amendment to the Bidding Document****Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/09/17-18 dated 01.01.2018 read with its Amendment No. 01 dated 11.01.2018**

The following changes have been authorised and incorporated in the above referred Bidding Document.

SECTION – VII**A. TECHNICAL SPECIFICATION:****Item No. 02 (RFx/Event number - 3000002576)****CSSD System**

Sl. No	Ref. Tender Page & Para	Existing Tender Specification	Amended Tender Specification
1	Para 9, Page 56 & 57	Bidder has to quote OPEX (Operational Cost) rate per hospital bed per day on annual basis as mentioned in Annexure 3. This OPEX rate must be inclusive of the cost of manpower, consumables any other recurring cost incurred for the receiving, sorting, washing, disinfecting, drying, packing, storage, issuing, quality control, routine cleaning etc. of the complete CSSD area as per SOP prescribed in tender. This OPEX (operational cost) rate shall be the basis for calculating OPEX (operational cost) of the CSSD for bid ranking purpose by multiplying the quoted annual OPEX rate with 710 beds per year. However, the OPEX shall	Bidder has to quote OPEX (Operational Cost) rate per hospital bed per day on annual basis as mentioned in Annexure 3. This OPEX rate must be inclusive of the cost of manpower, consumables any other recurring cost incurred for the receiving, sorting, washing, disinfecting, drying, packing, storage, issuing, quality control, routine cleaning etc. of the complete CSSD area as per SOP prescribed in tender. This OPEX (operational cost) rate shall be the basis for calculating OPEX (operational cost) of the CSSD for bid ranking purpose by multiplying the quoted annual OPEX rate with 710 beds per year. However, the OPEX shall be

Sl. No	Ref. Tender Page & Para	Existing Tender Specification	Amended Tender Specification
		be paid on monthly basis by multiplying the quoted annual OPEX rate per hospital bed per day with the actual number of operational hospital beds during that month as certified by the competent authority at NCI-AIIMS.	paid on monthly basis by multiplying the quoted annual OPEX rate per hospital bed per day with the actual number of operational hospital beds during that month as certified by the competent authority at NCI-AIIMS. The minimum number of Operational Beds for the purpose of OPEX shall not be less than 100 from the date of commissioning of the CSSD system.
2	Para 21, Page 63	Door: The sterilizer supplied should be pneumatically (Compressed Air) /electrical operated double door with fully automatic vertical sliding movement along with door safety features.	Door: The sterilizer supplied should be pneumatically (Compressed Air) /electrical operated double door with fully automatic Vertical / Horizontal sliding movement along with door safety features.
3	Para 21, Page 63	Door Safety System: The Sterilizer should be supplied with Automatic (Manual opening in case of automatic mechanism failure) vertical sliding door.	Door Safety System: The Sterilizer should be supplied with Automatic (Manual opening in case of automatic mechanism failure) Vertical/Horizontal sliding door.
4	Para 21, Page 64	Steam Generator: The sterilizer should have inbuilt steam generator of adequate capacity. In inbuilt steam generator should be made of 316L quality stainless steel.	Steam Generator: The sterilizer should have Inbuilt/External steam generator of adequate capacity. Steam generator should be made of 316L quality stainless steel. The steam generator/boilers must be licensed as per Indian Boiler act 1920 and regulations as amended from time to time (as applicable).

Item No. 03 (RFx/Event number- 3000002577)**Laundry System**

Sl. No	Ref. Tender Page & Para	Existing Tender Specification	Amended Tender Specification
1	Para 18 Page 98	Flatwork Ironer -with Feeder, Folder & Stacker (Roller Size- 450-480 Ø mm x 3000-3300 mm length)	Flatwork Ironer -with Feeder, Folder & Stacker (Roller Size- minimum 450 Ø mm x minimum 3000 mm length)
2	Para 19 Page 98	Flatwork Ironer -with Feeder, Folder & Stacker (Roller Size- 450-480 Ø mm x 1900-2000 mm length)	Flatwork Ironer -with Feeder, Folder & Stacker (Roller Size- minimum 450 Ø mm x minimum 1900 mm length)

Clarification: -

The bidder must submit necessary product catalogue/brochure/datasheet etc. in the technical bid and clearly highlight the technical points to enable purchaser to evaluate the various parameters asked in the technical specifications and in SOP, i.e. washing machine/ sluicing machine, cycle time, total load etc. must be clearly specified in the compliance sheet & wherever applicable to enable bid evaluation.

SECTION - VIII**QUALIFICATION CRITERIA**

For Item at Sl. No. 3 (Rfx/Event number 3000002577) (i.e Laundry System)

Existing Qualification Criteria for this item is amended and superseded by the following:

- Status:** The bidder must be a reputed manufacturer of the Laundry equipment or its authorized **agent** (Sub-authorization is not acceptable for the supply of equipment). The firm should have National/International recognition in execution of Laundry System of international standards and sustained history of credible performance. The firm manufacturing the Laundry equipment should have ISO 9001:2000 or above Certification.

2. Minimum Work of Similar Nature:

Eligible bidder(s) should have in the past 7 (seven) years prior to closing of bid submission, successfully executed Laundry orders at Govt. hospitals/institutions/Hotels etc. globally as stated below:

- a. One order of Laundry on SITC basis for a minimum value of 80% of the estimated CAPEX cost of Rs. 4 cr.

Or,

- b. Two orders of Laundry on SITC basis, each for a minimum value of 50% of the estimated CAPEX cost of Rs. 4 cr.

Or,

- c. Three orders of Laundry on SITC basis, each for a minimum value of 40% of the estimated CAPEX cost of Rs. 4 cr.

Note:

(i) The copies of order(s) along with the completion certificate(s) indicating that the specified order(s) have been completed, are to be submitted with technical bid. The value of such executed order(s) shall be brought to the current costing level by enhancing the actual value of order(s) at simple rate of 7% per annum, computed from the date of completion of such execution to the last date of receipt of tenders. Executed order value shall be limited to the upfront charges paid (DDP price) for items in the ordered BOQ, on Supply, Installation, Testing & Commissioning (SITC) basis, inclusive of warranty. CAMC & any other recurring costs will not be taken into consideration.

(ii) The Global experience of the OEM of Laundry equipment shall also be considered for 'Minimum Work of Similar Nature' in case:

- a. the bidder is a 100% owned Indian Subsidiary of an International firm. In this regard, relevant document to be submitted for proof of being an 'Indian Subsidiary'

Or,

- b. the authorized agent (bidder) is also a Laundry Operator with an average annual turnover of at least Rs. 1.0 cr. per year in the business of Laundry Operations during the last 3 financial years prior to closing of bid submission. In this regard, a copy of the turnover certificate from Chartered Accountant to be submitted along with the technical bid.

3. Financial Status: Eligible Bidders should not have incurred any loss in more than 2 years during the last five years ending 31st March 2017 or 30th June 2017 or 30th September 2017 or 31st December 2017. Copy of audited Profit & Loss account and Balance Sheet (duly self certified) for the immediate last five consecutive financial years should be submitted along with the bid.

All other contents of the tender enquiry including terms & conditions remain unaltered.