

**Amendment No. 3**

03-11-2017

**Sub: Amendment to the Bidding Document**

**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/02/17-18 dated 14.09.2017 read with its Amendment No. 1 and Amendment No. 2 dated 25.10.2017 and 27.10.2017 respectively.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - III  
SPECIAL INSTRUCTIONS TO BIDDERS (SIB)**

Following Para added in this section:

**33. Comparison of Bids**

33.2 Unit Prices for all optional items/accessories/services (if any) asked in the tender specifications must be quoted separately by all the bidders in their price bid. Such unit prices after multiplying by the required quantity shall be added and taken into consideration for comparison and ranking of bids.

**SECTION - VII  
TECHNICAL SPECIFICATION AND GENERAL POINTS**

**A. TECHNICAL SPECIFICATION:****Item sl. no. 04****DIGITAL SUBTRACTION ANGIOGRAPHY**

<b>Sl. No</b>	<b>Tender Page &amp; Para Ref</b>	<b>EXISTING</b>	<b>AMENDED AS</b>
1	Page 74 4. X- Ray Tube Para 4.5	4.5 X-ray tube should have secondary grid switching to reduce the soft X-rays.	4.5 X-ray tube should have <b>Primary / Secondary</b> grid switching to reduce the soft X-rays.
2	Page 74 5 Digital Imaging System : Para 5.1	5.1 A flat panel detector of dimensions: 40 x 30 cm or larger. Pixel size 160 micron or less.	5.1 A flat panel detector of dimensions: 40 x 30 cm or larger. Pixel size <b>200</b> micron or less.
3	Page 74-75 5. Digital Imaging System Para 5.3	5.3 It should be capable of acquisition and processing in 1920 x 2048 matrix with at least 16 bit digitization.	5.3 It should be capable of acquisition and processing in <b>1500</b> x 2048 matrix with at least <b>14 bit</b> digitization.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**