

## AMENDMENT-04

TENDER REF: HLL/PCD/PMSSY-II/NAGPUR/09/15-16

### Plasma Sterilizer

#### 1. Existing specifications

**Para 4.2:** Should be supplied with accessories to specification like instruments tray of all size and consumables to run at least 500 cycles of sterilization.

**Read as:**

**Para 4.2:** Instrument trays (small, medium, large) - 2 Nos, Chemical indicators -2000, Biological indicators – 100, Cartridges required to operate 500 cycles with maximum load Tyvek roll of at least 5 different size - 10 rolls each size (These are to be supplied along with equipment)

#### 2. Existing specifications

**Para 4.3:** Price of all consumables (including gas cartridge, labels, indicators, packing material, tray, adapter etc) required for five years (approx. 1000 cycles of full load) should be offered by the bidder. This would be considered for price comparison.

**Read as:**

**Para 4.3:** Instrument trays (small, medium, large) - 2 Nos, Chemical indicators -4000, Biological indicators – 200, Cartridges required to operate 1000 cycles with maximum load Tyvek roll of at least 5 different size - 20 rolls each size. This would be considered for price comparison.

Unit price of these items also to be quoted and fixed for Warranty and CMC period

#### 2. Existing specifications

**Para 7.1:** Certified to be in compliance with **ISO/EN 14937**. -Standards for sterilization equipment.

**Read as:**

**Para 7.1:** Certified to be in compliance with **EN ISO 14937**. -Standards for sterilization equipment Unit price of these items also to be quoted and fixed for at least 5years.

### **3. Existing specifications**

**Para 8.7:** The supplier must furnish satisfactory service report for about 3 years or more from at least 3 users preferably Govt. institutions.

**Read as:**

**Para 8.7:** DELETED

**4. Added Para:** Plasma sterilizer should be square or rectangular in shape