

Amendment No. 1

Date: 30/09/2016

Sub: Amendment to the Tender Enquiry Document.

Ref: Tender Enquiry No.: HLL/PCD/NIB-LTE-01/16-17 dated 15.09.2016

The following changes are being incorporated in the referred tender enquiry.

TECHNICAL SPECIFICATIONS

Schedule No. 05 pH Meter

1. Existing specification:

Para 7. Potential range 0 to +/- 1999.9 mV

Read as:

Potential range **0 to ±1200 mV**

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids