

Amendment No. 7

Date: 14/12/2015

Subject: Amendment to the Tender Enquiry Document

Ref: Tender Enquiry No.: HLL/HITES/PCD/RC-ME/01/2015 dated 06/10/2015 read with amendment no. 1, 2, 3, 4, 5 & 6 dated 21.11.2015, 23.11.2015, 30.11.2015, 05.12.2015, 07.12.2015 & 10.12.2015 respectively.

The following changes have been incorporated in the referred Tender Enquiry Document (TED).

<p style="text-align: center;">SECTION VII Technical Specification</p>
--

Group: Radiology

Item No. 23

Colour doppler flow ultrasound

1. Existing Para:- 2.1 - 13) The system should have high dynamic range of 170dB with scanning depth of 30cm or more.

Read as:- Para:- 2.1 - 13) The system should have high dynamic range of **180dB (+/- 20)** with scanning depth of 30cm or more.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

Prospective bidders are advised to check the website regularly prior to the closing date and time of online submission of tenders.

-- End --