

Amendment No. 10

Date: 14/01/2016

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/07/15-16 dated 28/10/2015 and subsequent amendments

The following changes have been incorporated in the referred NIT.

Section VII Technical Specification

Item Sl. No. 1 **CT SCAN – 64 ROWS / 128 Slice**

Existing Para:

Scope of work for turnkey CT (Page 50) Point 3: The cost of Turnkey for the area of 2500sq.ft and Air-conditioning of Tonnage 21 TR will be considered for Ranking / Evaluation purpose.

Read as:

The cost of Turnkey for the area of 1500 sq.ft and Air-conditioning of Tonnage 13 TR will be considered for Ranking / Evaluation purpose. However the bidder should quote unit rate for all the items. The payment shall be made as per actuals.

Item Sl. No. 17 **128 SLICE MDCT WITH INDEPENDENT 64 OR MORE ROWS OF DETECTOR**

Added Para under Turnkey Specifications (Page 81):

The cost of Turnkey for the area of 1500 sq.ft and Air-conditioning of Tonnage 13 TR will be considered for Ranking / Evaluation purpose. However the bidder should quote unit rate for all the items. The payment shall be made as per actuals.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids